



Benzoic acid




$\text{Response} = 1.868\text{e-}001 * \text{Amt} - 6.949\text{e-}002$
 Coef of Det (r^2) = 0.997332 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080421.M
 Calibration Table Last Updated: Wed Aug 04 17:59:00 2021

Hexachlorocyclopentadiene




Response = $2.342 \times 10^{-1} \times \text{Amt} - 6.312 \times 10^{-2}$
Coef of Det (r^2) = 0.997307 Curve Fit: Linear
Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080421.M
Calibration Table Last Updated: Wed Aug 04 17:59:00 2021

2,4-Dinitrophenol



Response = 1.642e-001 * Amt - 4.454e-002
 Coef of Det (r^2) = 0.997910 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080421.M
 Calibration Table Last Updated: Wed Aug 04 17:59:00 2021

Pentachlorophenol



Response = 1.179e-001 * Amt - 2.118e-002
 Coef of Det (r^2) = 0.998943 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080421.M
 Calibration Table Last Updated: Wed Aug 04 17:59:00 2021