

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080422\
 Data File : BF129587.D
 Acq On : 05 Aug 2022 00:52
 Operator : CG\JU
 Sample : N3931-18
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RVE-129

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 08/05/2022
 Supervised By : Jagrut Upadhyay 08/05/2022

Quant Time: Aug 05 02:10:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 05 02:05:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	159560	20.000	ng	0.00
21) Naphthalene-d8	8.139	136	572051	20.000	ng	0.00
39) Acenaphthene-d10	9.898	164	244884	20.000	ng	0.00
64) Phenanthrene-d10	11.386	188	351743	20.000	ng	0.00
76) Chrysene-d12	14.033	240	296680	20.000	ng	0.00
86) Perylene-d12	15.515	264	237352	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	1212846	123.823	ng	0.01
7) Phenol-d6	6.516	99	1485477	120.656	ng	0.00
23) Nitrobenzene-d5	7.422	82	939681	89.670	ng	-0.01
42) 2,4,6-Tribromophenol	10.692	330	281600	119.607	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	1504775	95.058	ng	0.00
79) Terphenyl-d14	12.974	244	1270765	80.808	ng	0.00
Target Compounds						Qvalue
75) Fluoranthene	12.598	202	50681	2.605	ng	99
78) Pyrene	12.833	202	67791	2.732	ng	96
88) Benzo(b)fluoranthene	15.074	252	32371m	2.056	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080422\
Data File : BF129587.D
Acq On : 05 Aug 2022 00:52
Operator : CG\JU
Sample : N3931-18
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RVE-129

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 08/05/2022
Supervised By :Jagrut Upadhyay 08/05/2022

Quant Time: Aug 05 02:10:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 02:05:37 2022
Response via : Initial Calibration

