


Data Path : Z:\HPCHEM1\BNA F\Data\BF080517\
 Data File : BF097416.D
 Acq On : 6 Aug 2017 00:58
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 07 16:02:51 2017



AutoFind: Scans 1601, 1602, 1603; Background Corrected with Scan 1596

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.0	15680	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.8	11561	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.1	14551	PASS
197	198	0.00	2	0.8	245	PASS
198	198	100	100	100.0	29031	PASS
199	198	5	9	6.6	1928	PASS
275	198	10	60	22.6	6566	PASS
365	198	1	100	2.5	715	PASS
441	442	0.01	24	16.3	4547	PASS
442	198	50	100	95.8	27826	PASS
443	442	15	24	20.4	5688	PASS