

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080624\
Data File : BF138806.D
Acq On : 06 Aug 2024 13:19
Operator : RC/JU
Sample : PB162267TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB162267TB

Quant Time: Aug 06 13:48:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	51259	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	214070	20.000	ng	0.00
39) Acenaphthene-d10	9.875	164	124665	20.000	ng	0.00
64) Phenanthrene-d10	11.363	188	235815	20.000	ng	0.00
76) Chrysene-d12	13.998	240	158113	20.000	ng	0.00
86) Perylene-d12	15.462	264	106451	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	450960	135.805	ng	0.02
7) Phenol-d6	6.487	99	603175	135.293	ng	0.00
23) Nitrobenzene-d5	7.410	82	389048	88.854	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	161135	157.794	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	748330	90.191	ng	0.00
79) Terphenyl-d14	12.951	244	940607	99.601	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080624\
Data File : BF138806.D
Acq On : 06 Aug 2024 13:19
Operator : RC/JU
Sample : PB162267TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB162267TB

Quant Time: Aug 06 13:48:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration

