

Data Path : Z:\HPCHEM1\BNA F\DATA\BF080715\
 Data File : BF080929.D
 Acq On : 7 Aug 2015 16:21
 Operator : TP/UM
 Sample : G3083-12DL 5X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :


Quant Time: Aug 08 01:59:00 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF080715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 08 01:33:05 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	75385	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	275982	20.00	ng	-0.01
38) Acenaphthene-d10	9.85	164	147238	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	309941	20.00	ng	0.00
75) Chrysene-d12	13.95	240	273196	20.00	ng	-0.01
86) Perylene-d12	15.36	264	221551	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	57128	12.38	ng	0.00
7) Phenol-d6	6.43	99	47468	7.51	ng	-0.01
23) Nitrobenzene-d5	7.37	82	78814	13.51	ng	-0.01
41) 2,4,6-Tribromophenol	10.62	330	31406	19.42	ng	-0.01
44) 2-Fluorobiphenyl	9.16	172	133971	12.81	ng	-0.01
78) Terphenyl-d14	12.90	244	163106	12.33	ng	-0.01

Target Compounds				Qvalue		
8) 2-Chlorophenol	6.58	128	33971	6.36	ng	83
10) Phenol	6.44	94	58652	7.82	ng	91
17) 2-Methylphenol	7.06	107	20178	4.43	ng	97
19) n-Nitroso-di-n-propylamine	7.37	70	10502	2.27	ng	# 73
20) 3+4-Methylphenols	7.21	107	40500	7.07	ng	# 49
24) Nitrobenzene	7.21	77	12540	2.09	ng	# 52
25) Isophorone	7.37	82	67558	6.02	ng	# 65
26) 2-Nitrophenol	7.71	139	57098	22.59	ng	# 86
27) 2,4-Dimethylphenol	7.74	122	33483	7.08	ng	96
29) 2,4-Dichlorophenol	7.95	162	14438	3.71	ng	86
32) Benzoic acid	7.74	122	33483	11.82	ng	# 15
35) Caprolactam	8.64	113	2830	2.02	ng	# 14
36) 4-Chloro-3-methylphenol	8.64	107	112462	23.70	ng	98
37) 2-Methylnaphthalene	8.64	142	79739	8.23	ng	# 23
42) 2,4,6-Trichlorophenol	9.08	196	57756	17.36	ng	99
43) 2,4,5-Trichlorophenol	9.12	196	31869	9.51	ng	# 93
50) 2,6-Dinitrotoluene	9.85	165	18892	7.48	ng	# 5
53) 2,4-Dinitrophenol	9.89	184	13612	14.32	ng	# 71
55) 4-Nitrophenol	9.94	139	17117	7.34	ng	# 46
56) 2,4-Dinitrotoluene	9.85	165	18892	5.59	ng	# 7
69) Pentachlorophenol	11.13	266	21184	9.98	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.81 min Scan# 413
Delta R.T. 0.00 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 75385
Ion Ratio Lower Upper
152 100
150 155.7 127.4 191.2
115 59.2 52.6 78.8




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 12.38 ng
RT: 5.38 min Scan# 288
Delta R.T. 0.00 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Tgt Ion:112 Resp: 57128
Ion Ratio Lower Upper
112 100
64 73.6 66.1 99.1
63 36.7 34.6 52.0



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 7.51 ng

RT: 6.43 min Scan# 380

Delta R.T. -0.01 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21

Instrument :

BNA_F

ClientSampled :


Tot Ion: 99 Resp: 47468

Ion Ratio Lower Upper




99 100

42 15.0 14.4 21.6

71 27.5 25.8 38.8



Abundance Ion 99.00 (98.70 to 99.70): BF0
Ion 42.00 (41.70 to 42.70): BF0
Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 6.36 ng

RT: 6.58 min Scan# 393

Delta R.T. -0.01 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21


Tgt Ion: 128 Resp: 33971

Ion Ratio Lower Upper



128 100


130 29.6 12.4 52.4

64 69.2 32.1 72.1



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF0





#10

Phenol

Concen: 7.82 ng

RT: 6.44 min Scan# 381

Delta R.T. -0.01 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 94 Resp: 58652

Ion Ratio Lower Upper




94 100

65 27.6 11.8 51.8

66 36.9 22.9 62.9



Abundance Ion 94.00 (93.70 to 94.70): BF0
Ion 65.00 (64.70 to 65.70): BF0
Ion 66.00 (65.70 to 66.70): BF0



#17

2-Methylphenol

Concen: 4.43 ng

RT: 7.06 min Scan# 435

Delta R.T. -0.01 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21

Tgt Ion: 107 Resp: 20178


Ion Ratio Lower Upper

107 100



108 114.9 90.3 135.5


77 45.8 34.6 51.8

79 44.8 33.6 50.4



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0









#19
n-Nitroso-di-n-propylamine
Concen: 2.27 ng
RT: 7.37 min Scan# 462
Delta R.T. 0.14 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21




Instrument :
BNA_F
ClientSampleId :


Tot Ion: 70 Resp: 10502
Ion Ratio Lower Upper
70 100
42 49.6 55.7 83.5#
101 0.0 7.0 10.6#
130 0.0 13.8 20.8#



#20
3+4-Methylphenols
Concen: 7.07 ng
RT: 7.21 min Scan# 448
Delta R.T. -0.01 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Tgt Ion:107 Resp: 40500
Ion Ratio Lower Upper
107 100
108 78.4 69.2 109.2
77 32.3 113.0 153.0#
79 23.0 10.3 50.3









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.09 min Scan# 525
Delta R.T. -0.01 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Instrument :
BNA_F
ClientSampleId :

Tot Ion:136	Resp:	275982
Ion Ratio	Lower	Upper
136	100	
137	11.1	8.5 12.7
54	15.4	7.1 10.7#
68	7.5	3.8 5.8#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#23
Nitrobenzene-d5
Concen: 13.51 ng
RT: 7.37 min Scan# 462
Delta R.T. -0.01 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Tgt Ion: 82	Resp:	78814
Ion Ratio	Lower	Upper
82	100	
128	42.0	34.1 51.1
54	59.4	46.5 69.7



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF0









#24
 Nitrobenzene
 Concen: 2.09 ng
 RT: 7.21 min Scan# 448
 Delta R.T. -0.18 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
77	100		
123	0.0	28.7	43.1#
65	8.6	12.5	18.7#



#25
 Isophorone
 Concen: 6.02 ng
 RT: 7.37 min Scan# 462
 Delta R.T. -0.27 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.0	5.8	8.6#
138	0.0	14.1	21.1#









#26
 2-Nitrophenol
 Concen: 22.59 ng
 RT: 7.71 min Scan# 492
 Delta R.T. 0.00 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21




Instrument :
 BNA_F
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
139	100		
109	28.1	24.1	36.1
65	52.8	54.0	81.0



#27
 2,4-Dimethylphenol
 Concen: 7.08 ng
 RT: 7.74 min Scan# 495
 Delta R.T. -0.01 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Tgt Ion	Ratio	Lower	Upper
122	100		
107	118.2	90.2	135.4
121	57.0	46.0	69.0









#29
2,4-Dichlorophenol
Concen: 3.71 ng
RT: 7.95 min Scan# 513
Delta R.T. 0.00 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
162	100		
164	69.3	42.3	82.3
98	34.5	27.7	67.7




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BF0






#32
Benzoic acid
Concen: 11.82 ng
RT: 7.74 min Scan# 495
Delta R.T. -0.13 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Tgt Ion	Ratio	Lower	Upper
122	100		
105	3.7	103.6	143.6#
77	37.3	70.4	110.4#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF0









#35
 Caprolactam
 Concen: 2.02 ng
 RT: 8.64 min Scan# 573
 Delta R.T. 0.09 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:113 Resp: 2830
 Ion Ratio Lower Upper
 113 100
 55 110.7 196.4 236.4#
 56 9.9 137.9 177.9#



#36
 4-Chloro-3-methylphenol
 Concen: 23.70 ng
 RT: 8.64 min Scan# 573
 Delta R.T. -0.01 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Tgt Ion:107 Resp: 112462
 Ion Ratio Lower Upper
 107 100
 144 21.8 17.3 25.9
 142 70.1 54.4 81.6









#37
 2-Methylnaphthalene
 Concen: 8.23 ng
 RT: 8.64 min Scan# 573
 Delta R.T. -0.17 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:142 Resp: 79739
 Ion Ratio Lower Upper
 142 100
 141 10.3 71.7 107.5#
 115 1.7 30.9 46.3#




Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E






#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.85 min Scan# 679
 Delta R.T. 0.00 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Tgt Ion:164 Resp: 147238
 Ion Ratio Lower Upper
 164 100
 162 105.8 80.5 120.7
 160 48.2 36.6 54.8



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 19.42 ng
 RT: 10.62 min Scan# 747
 Delta R.T. -0.01 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:330 Resp: 31406
 Ion Ratio Lower Upper
 330 100
 332 94.7 76.2 114.4
 141 40.0 27.8 41.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 17.36 ng
 RT: 9.08 min Scan# 612
 Delta R.T. 0.00 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Tgt Ion:196 Resp: 57756
 Ion Ratio Lower Upper
 196 100
 198 95.9 75.9 113.9
 200 30.5 23.1 34.7



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 9.51 ng
 RT: 9.12 min Scan# 615
 Delta R.T. -0.01 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
196	100		
198	97.2	78.9	118.3
97	47.9	29.1	43.7
132	28.0	19.1	28.7




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 12.81 ng
 RT: 9.16 min Scan# 619
 Delta R.T. -0.01 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.2	29.4	44.0
170	22.5	19.7	29.5



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#50
 2,6-Dinitrotoluene
 Concen: 7.48 ng
 RT: 9.85 min Scan# 679
 Delta R.T. 0.22 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:165 Resp: 18892
 Ion Ratio Lower Upper
 165 100
 63 0.0 77.6 116.4#
 89 0.0 64.5 96.7#




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 89.00 (88.70 to 89.70): BF0






#53
 2,4-Dinitrophenol
 Concen: 14.32 ng
 RT: 9.89 min Scan# 683
 Delta R.T. -0.01 min
 Lab File: BF080929.D
 Acq: 7 Aug 2015 16:21

Tgt Ion:184 Resp: 13612
 Ion Ratio Lower Upper
 184 100
 63 113.6 59.7 89.5#
 154 74.9 53.8 80.8



Abundance Ion 184.00 (183.70 to 184.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 154.00 (153.70 to 154.70): E





#55

4-Nitrophenol

Concen: 7.34 na

RT: 9.94 min Scan# 687

Delta R.T. -0.02 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21

Instrument :

BNA_F

ClientSampleId :


Tot Ion:139 Resp: 17117

Ion Ratio Lower Upper

139 100

109 80.5 44.7 84.7




65 151.9 59.5 99.5#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 5.59 ng

RT: 9.85 min Scan# 679

Delta R.T. -0.18 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21

Tgt Ion:165 Resp: 18892


Ion Ratio Lower Upper

165 100

63 0.0 56.7 85.1#

89 0.0 80.4 120.6#

182 0.0 1.5 2.3#






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.32 min Scan# 808
Delta R.T. 0.00 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Instrument :
BNA_F
ClientSampled :

Tot Ion:188 Resp: 309941
Ion Ratio Lower Upper
188 100
94 10.4 8.8 13.2
80 11.5 9.5 14.3




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0






#69
Pentachlorophenol
Concen: 9.98 ng
RT: 11.13 min Scan# 791
Delta R.T. 0.00 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Tgt Ion:266 Resp: 21184
Ion Ratio Lower Upper
266 100
268 67.7 51.6 77.4
264 64.1 50.1 75.1



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E





#75

Chrvsene-d12

Concen: 20.00 ng

RT: 13.95 min Scan# 1038

Delta R.T. -0.01 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 273196

Ion Ratio Lower Upper

240 100

120 11.1 5.2 7.8#

236 26.6 20.2 30.4




Abundance Ion 240.00 (239.70 to 240.70): E


Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E

13.95



Time-->



#78

Terphenyl-d14

Concen: 12.33 ng

RT: 12.90 min Scan# 946

Delta R.T. -0.01 min

Lab File: BF080929.D

Acq: 7 Aug 2015 16:21


Tgt Ion:244 Resp: 163106

Ion Ratio Lower Upper

244 100

212 7.8 5.8 8.8

122 15.1 9.4 14.2#




Abundance Ion 244.00 (243.70 to 244.70): E




Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E

12.90



Time-->



#86
Pervlene-d12
Concen: 20.00 ng
RT: 15.36 min Scan# 1161
Delta R.T. 0.00 min
Lab File: BF080929.D
Acq: 7 Aug 2015 16:21

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.4	29.0
265	21.7	17.0	25.4

