



Data Path : Z:\HPCHEM1\BNA F\DATA\BF080717\
Data File : BF097425.D
Acq On : 7 Aug 2017 13:08
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 08 08:43:44 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 16:02:51 2017
Response via : Initial Calibration



TIC: BF097425.D

(76) Benzidine



12.245min (+0.000) 0.00ng

response 451550

Ion	Exp%	Act%
184.00	100	100
185.00	19.40	14.60#
183.00	12.20	11.90
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF080717\
Data File : BF097425.D
Acq On : 7 Aug 2017 13:08
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 08 08:43:44 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 16:02:51 2017
Response via : Initial Calibration



TIC: BF097425.D

(69) Pentachlorophenol (C)

10.739min (-0.006) 0.00ng

response 43153

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	58.54
264.00	64.60	58.57
0.00	0.00	0.00