


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF080719\
 Data File : BF116039.D
 Acq On : 7 Aug 2019 9:34
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 07 11:13:18 2019



AutoFind: Scans 1619, 1620, 1621; Background Corrected with Scan 1614

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	64458	PASS
68	69	0.00	2	1.7	1117	PASS
69	198	0.00	100	36.0	65469	PASS
70	69	0.00	2	0.6	394	PASS
127	198	10	80	48.9	88890	PASS
197	198	0.00	2	0.5	931	PASS
198	198	100	100	100.0	181792	PASS
199	198	5	9	6.7	12108	PASS
275	198	10	60	26.6	48408	PASS
365	198	1	100	3.3	6028	PASS
441	198	0.01	100	12.3	22393	PASS
442	198	50	100	79.7	144965	PASS
443	442	15	24	18.9	27377	PASS