

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080721\
 Data File : BF124966.D
 Acq On : 7 Aug 2021 9:57
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 07 11:06:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 06 11:27:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.457	0.480	-5.0	112	0.04
3	Pyridine	1.061	1.079	-1.7	102	0.03
4	n-Nitrosodimethylamine	0.754	0.797	-5.7	105	0.04
5 S	2-Fluorophenol	1.221	1.228	-0.6	107	0.00
6	Aniline	1.626	1.674	-3.0	107	0.01
7 S	Phenol-d6	1.539	1.586	-3.1	106	0.00
8	2-Chlorophenol	1.339	1.336	0.2	106	0.00
9	Benzaldehyde	0.838	0.762	9.1	98	0.00
10 C	Phenol	1.588	1.640	-3.3	110	0.00
11	bis(2-Chloroethyl)ether	1.192	1.234	-3.5	106	0.01
12	1,3-Dichlorobenzene	1.458	1.470	-0.8	106	0.00
13 C	1,4-Dichlorobenzene	1.457	1.488	-2.1	105	0.00
14	1,2-Dichlorobenzene	1.366	1.411	-3.3	108	0.01
15	Benzyl Alcohol	1.110	1.088	2.0	101	0.00
16	2,2'-oxybis(1-Chloropropane	1.909	1.941	-1.7	110	0.00
17	2-Methylphenol	1.002	1.040	-3.8	111	0.00
18	Hexachloroethane	0.571	0.602	-5.4	113	0.00
19 P	n-Nitroso-di-n-propylamine	0.913	0.903	1.1	108	0.00
20	3+4-Methylphenols	1.331	1.360	-2.2	109	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.01
22	Acetophenone	0.490	0.489	0.2	108	0.01
23 S	Nitrobenzene-d5	0.382	0.379	0.8	105	0.00
24	Nitrobenzene	0.370	0.370	0.0	103	0.00
25	Isophorone	0.653	0.656	-0.5	107	0.00
26 C	2-Nitrophenol	0.188	0.192	-2.1	105	0.01
27	2,4-Dimethylphenol	0.266	0.273	-2.6	106	0.00
28	bis(2-Chloroethoxy)methane	0.371	0.358	3.5	103	0.00
29 C	2,4-Dichlorophenol	0.297	0.302	-1.7	106	0.00
30	1,2,4-Trichlorobenzene	0.332	0.327	1.5	103	0.00
31	Naphthalene	1.028	1.041	-1.3	107	0.00
32	Benzoic acid	0.138	0.127	8.0	96	0.00
33	4-Chloroaniline	0.383	0.397	-3.7	107	0.01
34 C	Hexachlorobutadiene	0.199	0.199	0.0	102	0.01
35	Caprolactam	0.079	0.083	-5.1	101	0.01
36 C	4-Chloro-3-methylphenol	0.310	0.305	1.6	104	0.00
37	2-Methylnaphthalene	0.691	0.686	0.7	106	0.01
38	1-Methylnaphthalene	0.646	0.660	-2.2	109	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	0.574	0.564	1.7	104	0.01
41 P	Hexachlorocyclopentadiene	0.139	0.196	-41.0#	100	0.00
42 S	2,4,6-Tribromophenol	0.179	0.171	4.5	98	0.00
43 C	2,4,6-Trichlorophenol	0.382	0.385	-0.8	109	0.00
44	2,4,5-Trichlorophenol	0.387	0.399	-3.1	103	0.00
45 S	2-Fluorobiphenyl	1.367	1.376	-0.7	107	0.01
46	1,1'-Biphenyl	1.490	1.495	-0.3	107	0.01
47	2-Chloronaphthalene	1.186	1.187	-0.1	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080721\
 Data File : BF124966.D
 Acq On : 7 Aug 2021 9:57
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 07 11:06:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 06 11:27:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.348	0.348	0.0	106	0.00
49	Acenaphthylene	1.801	1.842	-2.3	112	0.01
50	Dimethylphthalate	1.355	1.373	-1.3	108	0.00
51	2,6-Dinitrotoluene	0.302	0.301	0.3	104	0.01
52 C	Acenaphthene	1.092	1.094	-0.2	108	0.01
53	3-Nitroaniline	0.318	0.325	-2.2	112	0.00
54 P	2,4-Dinitrophenol	0.132	0.120	9.1	94	0.00
55	Dibenzofuran	1.706	1.744	-2.2	110	0.01
56 P	4-Nitrophenol	0.210	0.212	-1.0	102	0.00
57	2,4-Dinitrotoluene	0.410	0.417	-1.7	107	0.00
58	Fluorene	1.308	1.335	-2.1	109	0.00
59	2,3,4,6-Tetrachlorophenol	0.290	0.306	-5.5	109	0.00
60	Diethylphthalate	1.391	1.381	0.7	108	0.00
61	4-Chlorophenyl-phenylether	0.618	0.647	-4.7	109	0.01
62	4-Nitroaniline	0.309	0.304	1.6	102	0.00
63	Azobenzene	1.360	1.313	3.5	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	0.125	0.116	7.2	101	0.00
66 c	n-Nitrosodiphenylamine	0.648	0.641	1.1	105	0.00
67	4-Bromophenyl-phenylether	0.219	0.218	0.5	105	0.01
68	Hexachlorobenzene	0.238	0.232	2.5	104	0.00
69	Atrazine	0.195	0.182	6.7	99	0.00
70 C	Pentachlorophenol	0.102	0.090	11.8	91	0.00
71	Phenanthrene	1.085	1.072	1.2	105	0.01
72	Anthracene	1.088	1.091	-0.3	108	0.01
73	Carbazole	0.998	0.987	1.1	107	0.00
74	Di-n-butylphthalate	1.268	1.209	4.7	104	0.00
75 C	Fluoranthene	1.122	1.096	2.3	105	0.01
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzidine	0.535	0.611	-14.2	119	0.00
78	Pyrene	1.575	1.663	-5.6	104	0.01
79 S	Terphenyl-d14	1.186	1.195	-0.8	101	0.01
80	Butylbenzylphthalate	0.701	0.723	-3.1	103	0.01
81	Benzo(a)anthracene	1.303	1.307	-0.3	101	0.00
82	3,3'-Dichlorobenzidine	0.344	0.346	-0.6	97	0.01
83	Chrysene	1.260	1.248	1.0	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.973	0.984	-1.1	97	0.00
85 c	Di-n-octyl phthalate	1.534	1.472	4.0	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	1.191	1.185	0.5	100	0.00
88	Benzo(b)fluoranthene	1.404	1.474	-5.0	99	0.00
89	Benzo(k)fluoranthene	1.277	1.259	1.4	95	0.00
90 C	Benzo(a)pyrene	1.196	1.187	0.8	97	0.01
91	Dibenzo(a,h)anthracene	1.047	1.042	0.5	98	0.01
92	Benzo(g,h,i)perylene	1.004	0.995	0.9	101	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080721\
Data File : BF124966.D
Acq On : 7 Aug 2021 9:57
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 07 11:06:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 06 11:27:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0