




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080724\
Data File : BF138833.D
Acq On : 07 Aug 2024 10:30
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 07 11:59:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration



TIC: BF138833.D\data.ms

(70) Pentachlorophenol (C)




11.169min (-0.006) 57532.53 ng

response 68125

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.50	62.36
264.00	62.20	63.20
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080724\
Data File : BF138833.D
Acq On : 07 Aug 2024 10:30
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 07 11:59:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration



TIC: BF138833.D\data.ms

(77) Benzidine

12.698min (-0.006) 0.00 ng

response 323216

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	14.17
183.00	12.00	12.73
0.00	0.00	0.00