

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF080816\
 Data File : BF089661.D
 Acq On : 8 Aug 2016 14:54
 Operator : UM/SJ
 Sample : PB92687BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB92687BL

Quant Time: Aug 09 03:46:10 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF080416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 05 01:39:23 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	44063	20.00	ng	-0.02
21) Naphthalene-d8	9.45	136	188351	20.00	ng	-0.02
38) Acenaphthene-d10	12.26	164	92681	20.00	ng	-0.03
63) Phenanthrene-d10	14.66	188	185263	20.00	ng	-0.03
75) Chrysene-d12	18.37	240	128515	20.00	ng	-0.02
86) Perylene-d12	20.02	264	98339	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.31	112	342100	130.13	ng	0.00
7) Phenol-d6	6.96	99	462399	129.66	ng	-0.02
23) Nitrobenzene-d5	8.33	82	284992	83.69	ng	-0.03
41) 2,4,6-Tribromophenol	13.57	330	96499	126.17	ng	-0.02
44) 2-Fluorobiphenyl	11.23	172	554626	77.77	ng	-0.02
78) Terphenyl-d14	17.03	244	500143	76.83	ng	-0.02

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF080816\
Data File : BF089661.D
Acq On : 8 Aug 2016 14:54
Operator : UM/SJ
Sample : PB92687BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB92687BL

Quant Time: Aug 09 03:46:10 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF080416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 01:39:23 2016
Response via : Initial Calibration

