

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF080816\  
 Data File : BF089685.D  
 Acq On : 9 Aug 2016 6:29  
 Operator : UM/SJ  
 Sample : H4323-01  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 WC-DECONFLUID-13

Quant Time: Aug 09 07:16:32 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF080416.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Aug 05 01:39:23 2016  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min)  |
|--------------------------------|-------|------|----------|--------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4      | 7.46  | 152  | 47593    | 20.00  | ng    | 0.01      |
| 21) Naphthalene-d8             | 9.46  | 136  | 193146   | 20.00  | ng    | -0.01     |
| 38) Acenaphthene-d10           | 12.27 | 164  | 91038    | 20.00  | ng    | -0.02     |
| 63) Phenanthrene-d10           | 14.66 | 188  | 143454   | 20.00  | ng    | -0.03     |
| 75) Chrysene-d12               | 18.38 | 240  | 100304   | 20.00  | ng    | -0.01     |
| 86) Perylene-d12               | 20.06 | 264  | 71676    | 20.00  | ng    | 0.01      |
| System Monitoring Compounds    |       |      |          |        |       |           |
| 5) 2-Fluorophenol              | 0.00  | 112  | 0        | 0.00   | ng    |           |
| 7) Phenol-d6                   | 7.04  | 99   | 385793   | 100.15 | ng    | 0.06      |
| 23) Nitrobenzene-d5            | 8.35  | 82   | 306362   | 87.73  | ng    | -0.01     |
| 41) 2,4,6-Tribromophenol       | 13.58 | 330  | 85170    | 113.37 | ng    | -0.01     |
| 44) 2-Fluorobiphenyl           | 11.23 | 172  | 560959   | 80.08  | ng    | -0.02     |
| 78) Terphenyl-d14              | 17.04 | 244  | 360470   | 70.95  | ng    | -0.01     |
| Target Compounds               |       |      |          |        |       |           |
| 31) Naphthalene                | 9.50  | 128  | 605868   | 59.02  | ng    | Qvalue 99 |
| 37) 2-Methylnaphthalene        | 10.62 | 142  | 21496    | 3.40   | ng    | 99        |
| 49) Dimethylphthalate          | 11.92 | 163  | 28586    | 4.01   | ng    | # 90      |
| 83) Bis(2-ethylhexyl)phthalate | 18.46 | 149  | 42610    | 8.16   | ng    | # 89      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF080816\  
Data File : BF089685.D  
Acq On : 9 Aug 2016 6:29  
Operator : UM/SJ  
Sample : H4323-01  
Misc :  
ALS Vial : 26 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
WC-DECONFLUID-13

Quant Time: Aug 09 07:16:32 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF080416.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Aug 05 01:39:23 2016  
Response via : Initial Calibration

