




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF080916\
Data File : BF089688.D
Acq On : 9 Aug 2016 14:56
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 10 07:02:55 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF080416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 01:39:23 2016
Response via : Initial Calibration



TIC: BF089688.D

(76) Benzidine




16.697min (-0.034) 0.00ng

response 82414

Ion	Exp%	Act%
184.00	100	100
185.00	14.40	14.89
183.00	12.30	11.99
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF080916\
Data File : BF089688.D
Acq On : 9 Aug 2016 14:56
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 10 07:02:55 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF080416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 01:39:23 2016
Response via : Initial Calibration



TIC: BF089688.D

(69) Pentachlorophenol (C)

14.388min (-0.034) 0.00ng

response 10692

Ion	Exp%	Act%
265.70	100	100
268.00	63.20	64.26
264.00	61.90	66.63
0.00	0.00	0.00