

Data Path : Z:\HPCHEM1\BNA F\DATA\BF080917\
 Data File : BF097539.D
 Acq On : 10 Aug 2017 2:11
 Operator : SJ/JU
 Sample : I4677-07
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-F9-WSW

Quant Time: Aug 10 14:13:22 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 07 16:02:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.40	152	113330	20.00	ng	-0.01
21) Naphthalene-d8	7.68	136	468183	20.00	ng	-0.02
38) Acenaphthene-d10	9.42	164	174895	20.00	ng	-0.02
63) Phenanthrene-d10	10.90	188	281300	20.00	ng	-0.02
75) Chrysene-d12	13.52	240	175777	20.00	ng	-0.02
86) Perylene-d12	14.86	264	185824	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.99	112	803623	106.90	ng	-0.01
7) Phenol-d6	6.07	99	978018	113.96	ng	-0.01
23) Nitrobenzene-d5	6.97	82	584754	73.27	ng	-0.02
41) 2,4,6-Tribromophenol	10.21	330	144633	104.67	ng	-0.02
44) 2-Fluorobiphenyl	8.76	172	825957	80.15	ng	-0.02
78) Terphenyl-d14	12.48	244	481512	55.85	ng	-0.02
Target Compounds						
11) bis(2-Chloroethyl)ether	6.15	93	68214	8.472	ng	88
29) 2,4-Dichlorophenol	7.60	162	15702	2.457	ng	93
49) Dimethylphthalate	9.16	163	132423	9.971	ng	98
69) Pentachlorophenol	10.73	266	6401	4.914	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF080917\
Data File : BF097539.D
Acq On : 10 Aug 2017 2:11
Operator : SJ/JU
Sample : I4677-07
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-F9-WSW

Quant Time: Aug 10 14:13:22 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 16:02:51 2017
Response via : Initial Calibration

