

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF080918\
 Data File : BF108075.D
 Acq On : 10 Aug 2018 8:43
 Operator : JU/SJ
 Sample : J4388-05DL 4X
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S-3ADL

Quant Time: Aug 10 10:50:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 12:24:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.01	152	285452	20.00	ng	0.00
21) Naphthalene-d8	8.30	136	1075987	20.00	ng	0.00
38) Acenaphthene-d10	10.07	164	524439	20.00	ng	0.00
63) Phenanthrene-d10	11.56	188	881505	20.00	ng	0.00
75) Chrysene-d12	14.22	240	669391	20.00	ng	0.00
86) Perylene-d12	15.79	264	594464	20.00	ng	0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	485068	30.77	ng	0.00
7) Phenol-d6	6.62	99	673145	32.13	ng	-0.01
23) Nitrobenzene-d5	7.57	82	417684	21.66	ng	0.00
41) 2,4,6-Tribromophenol	10.85	330	135304	25.87	ng	0.00
44) 2-Fluorobiphenyl	9.37	172	755489	23.13	ng	0.00
78) Terphenyl-d14	13.15	244	624618	23.73	ng	0.00


Target Compounds					Qvalue	
25) Isophorone	7.57	82	417672	11.364	ng	# 64
32) Benzoic acid	7.96	122	3288	6.279	ng	92
49) Dimethylphthalate	9.77	163	122359	3.349	ng	# 99
50) 2,6-Dinitrotoluene	10.07	165	70808	9.264	ng	# 25
51) Acenaphthene	10.10	154	96178	3.250	ng	99
57) Fluorene	10.61	166	136413	4.019	ng	98
64) 4,6-Dinitro-2-methylphenol	10.85	198	354	5.454	ng	# 1
70) Phenanthrene	11.59	178	1936497	46.576	ng	97
71) Anthracene	11.64	178	426254	10.003	ng	99
72) Carbazole	11.79	167	281864	7.445	ng	98
74) Fluoranthene	12.79	202	2820915	62.167	ng	94
77) Pyrene	13.02	202	2385766	56.295	ng	95
80) Benzo(a)anthracene	14.21	228	1227873	31.962	ng	98
82) Chrysene	14.25	228	1161003	32.965	ng	97
85) Indeno(1,2,3-cd)pyrene	17.40	276	500299	16.394	ng	# 86
87) Benzo(b)fluoranthene	15.32	252	1848958	52.052	ng	96
88) Benzo(k)fluoranthene	15.43	252	185923	5.694	ng	94
89) Benzo(a)pyrene	15.72	252	905039	28.453	ng	95
90) Dibenzo(a,h)anthracene	17.41	278	122468	4.653	ng	98
91) Benzo(g,h,i)perylene	17.90	276	440396	16.994	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF080918\
Data File : BF108075.D
Acq On : 10 Aug 2018 8:43
Operator : JU/SJ
Sample : J4388-05DL 4X
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S-3ADL

Quant Time: Aug 10 10:50:37 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 12:24:24 2018
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.01 min Scan# 795
 Delta R.T. 0.00 min
 Lab File: BF108075.D
 Acq: 10 Aug 2018 8:43

Instrument :
 BNA_F
 ClientSampleId :
 S-3ADL




Tgt Ion	Ratio	Lower	Upper
152	100		
150	154.7	124.6	186.8
115	61.4	50.1	75.1




#5

2-Fluorophenol
 Concen: 30.765 ng
 RT: 5.63 min Scan# 560
 Delta R.T. 0.00 min
 Lab File: BF108075.D
 Acq: 10 Aug 2018 8:43

Tgt Ion	Ratio	Lower	Upper
112	100		
64	66.8	47.9	71.9
63	34.5	23.7	35.5





#7

Phenol-d6

Concen: 32.128 ng

RT: 6.62 min Scan# 728

Delta R.T. -0.01 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Instrument :

BNA_F

ClientSampled :

S-3ADL


Tgt Ion: 99 Resp: 673145

Ion Ratio Lower Upper




99 100

42 23.7 14.5 21.7#

71 35.4 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.30 min Scan# 1014

Delta R.T. 0.00 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Tgt Ion: 136 Resp: 1075987


Ion Ratio Lower Upper

136 100



137 10.9 8.9 13.3


54 10.0 6.6 9.8#

68 6.0 5.0 7.4



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 21.661 ng
 RT: 7.57 min Scan# 890
 Delta R.T. -0.01 min
 Lab File: BF108075.D
 Acq: 10 Aug 2018 8:43

Instrument :
 BNA_F
 ClientSampleId :
 S-3ADL

Tgt Ion	Ratio	Lower	Upper
82	100		
128	41.6	36.1	54.1
54	57.6	41.4	62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#25
 Isophorone
 Concen: 11.364 ng
 RT: 7.57 min Scan# 890
 Delta R.T. -0.26 min
 Lab File: BF108075.D
 Acq: 10 Aug 2018 8:43

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.0	5.2	7.8#
138	0.0	14.9	22.3#



Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 95.00 (94.70 to 95.70): BF1
 Ion 138.00 (137.70 to 138.70): BF1









#32
Benzoic acid
Concen: 6.279 ng
RT: 7.96 min Scan# 956
Delta R.T. -0.08 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Instrument :
BNA_F
ClientSampleId :
S-3ADL

Tgt Ion:122 Resp: 3288
Ion Ratio Lower Upper
122 100
105 108.8 101.1 141.1
77 95.0 70.7 110.7




Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF1






#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.07 min Scan# 1314
Delta R.T. 0.00 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Tgt Ion:164 Resp: 524439
Ion Ratio Lower Upper
164 100
162 102.2 86.1 129.1
160 46.6 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 25.865 ng
 RT: 10.85 min Scan# 1448
 Delta R.T. -0.01 min
 Lab File: BF108075.D
 Acq: 10 Aug 2018 8:43

Instrument :
 BNA_F
 ClientSampled :
 S-3ADL

Tgt Ion	Ratio	Lower	Upper
330	100		
332	94.8	77.0	115.6
141	34.7	29.9	44.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 23.128 ng
 RT: 9.37 min Scan# 1196
 Delta R.T. -0.01 min
 Lab File: BF108075.D
 Acq: 10 Aug 2018 8:43

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.7	44.5
170	23.6	19.6	29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#49
Dimethylphthalate
Concen: 3.349 ng
RT: 9.77 min Scan# 1263
Delta R.T. -0.01 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Instrument :
BNA_F
ClientSampleId :
S-3ADL

Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.4	2.7	4.1#
164	10.0	8.2	12.2




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 9.264 ng
RT: 10.07 min Scan# 1314
Delta R.T. 0.22 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Tgt Ion	Ratio	Lower	Upper
165	100		
63	1.4	44.7	67.1#
89	1.3	44.0	66.0#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1





#51

Acenaphthene

Concen: 3.250 ng

RT: 10.10 min Scan# 1319

Delta R.T. -0.01 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Instrument :

BNA_F

ClientSampleId :

S-3ADL


Tgt Ion:154 Resp: 96178

Ion Ratio Lower Upper

154 100

153 115.2 91.2 136.8




152 55.3 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#57

Fluorene

Concen: 4.019 ng

RT: 10.61 min Scan# 1407

Delta R.T. -0.01 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43


Tgt Ion:166 Resp: 136413

Ion Ratio Lower Upper

166 100

165 98.3 79.8 119.8



167 12.1 10.6 16.0




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.56 min Scan# 1568
Delta R.T. 0.00 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43




Instrument :
BNA_F
ClientSampleId :
S-3ADL


Tgt Ion:188 Resp: 881505
Ion Ratio Lower Upper
188 100
94 9.5 8.5 12.7
80 10.6 9.2 13.8



#64
4,6-Dinitro-2-methylphenol
Concen: 5.454 ng
RT: 10.85 min Scan# 1448
Delta R.T. 0.20 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Tgt Ion:198 Resp: 354
Ion Ratio Lower Upper
198 100
51 418.7 37.7 77.7#
105 246.3 36.3 76.3#





#70

Phenanthrene

Concen: 46.576 ng

RT: 11.59 min Scan# 1573

Delta R.T. 0.00 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Instrument :

BNA_F

ClientSampled :

S-3ADL


Tgt Ion:178 Resp: 1936497

Ion Ratio Lower Upper

178 100

176 21.8 15.8 23.8




179 17.0 13.0 19.6



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#71

Anthracene

Concen: 10.003 ng

RT: 11.64 min Scan# 1581

Delta R.T. 0.00 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43


Tgt Ion:178 Resp: 426254

Ion Ratio Lower Upper

178 100

176 18.9 15.4 23.2



179 15.1 12.6 19.0




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#72

Carbazole

Concen: 7.445 ng

RT: 11.79 min Scan# 1607

Delta R.T. 0.00 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Instrument :

BNA_F

ClientSampleId :

S-3ADL


Tgt Ion:167 Resp: 281864

Ion Ratio Lower Upper

167 100

166 22.6 17.6 26.4




139 12.8 10.9 16.3



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Fluoranthene

Concen: 62.167 ng

RT: 12.79 min Scan# 1777

Delta R.T. 0.01 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43


Tgt Ion:202 Resp: 2820915

Ion Ratio Lower Upper

202 100

101 12.8 0.0 34.1



203 21.6 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.22 min Scan# 2020

Delta R.T. 0.00 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Instrument :

BNA_F

ClientSampleId :

S-3ADL


Tgt Ion:240 Resp: 669391

Ion Ratio Lower Upper

240 100

120 11.2 9.5 14.3




236 26.8 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 56.295 ng

RT: 13.02 min Scan# 1816

Delta R.T. 0.01 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43


Tgt Ion:202 Resp: 2385766

Ion Ratio Lower Upper

202 100

200 23.8 17.4 26.2



203 20.4 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 23.725 ng

RT: 13.15 min Scan# 1838

Delta R.T. 0.00 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Instrument :

BNA_F

ClientSampleId :

S-3ADL


Tgt Ion:244 Resp: 624618

Ion Ratio Lower Upper

244 100

212 6.8 5.4 8.0




122 9.1 11.0 16.4#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 31.962 ng

RT: 14.21 min Scan# 2018

Delta R.T. 0.00 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43


Tgt Ion:228 Resp: 1227873

Ion Ratio Lower Upper

228 100

226 29.0 22.6 33.8



229 21.0 15.8 23.6




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 32.965 ng

RT: 14.25 min Scan# 2025

Delta R.T. 0.01 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43

Instrument :

BNA_F

ClientSampleId :

S-3ADL


Tgt Ion:228 Resp: 1161003

Ion Ratio Lower Upper

228 100

226 31.4 25.0 37.6




229 23.2 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 16.394 ng

RT: 17.40 min Scan# 2561

Delta R.T. 0.01 min

Lab File: BF108075.D

Acq: 10 Aug 2018 8:43


Tgt Ion:276 Resp: 500299

Ion Ratio Lower Upper

276 100

138 18.9 25.0 37.6#



277 23.5 20.1 30.1




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Perylene-d12
Concen: 20.000 ng
RT: 15.79 min Scan# 2287
Delta R.T. 0.01 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43




Instrument :
BNA_F
ClientSampleId :
S-3ADL


Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.2	28.8
265	21.3	17.5	26.3



#87
Benzo(b)fluoranthene
Concen: 52.052 ng
RT: 15.32 min Scan# 2207
Delta R.T. 0.01 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.4	17.0	25.6
125	10.0	9.0	13.6









#88
Benzo(k)fluoranthene
Concen: 5.694 ng
RT: 15.43 min Scan# 2226
Delta R.T. 0.09 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43




Instrument :
BNA_F
ClientSampleId :
S-3ADL


Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.9	17.9	26.9
125	11.6	10.2	15.2



#89
Benzo(a)pyrene
Concen: 28.453 ng
RT: 15.72 min Scan# 2275
Delta R.T. 0.01 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.7	17.1	25.7
125	10.2	9.8	14.6









#90
Dibenzo(a,h)anthracene
Concen: 4.653 ng
RT: 17.41 min Scan# 2563
Delta R.T. -0.01 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Instrument :
BNA_F
ClientSampleId :
S-3ADL

Tgt Ion	Ratio	Lower	Upper
278	100		
139	17.9	15.9	23.9
279	24.0	19.0	28.4



#91
Benzo(g,h,i)perylene
Concen: 16.994 ng
RT: 17.90 min Scan# 2646
Delta R.T. 0.01 min
Lab File: BF108075.D
Acq: 10 Aug 2018 8:43

Tgt Ion	Ratio	Lower	Upper
276	100		
277	22.2	17.9	26.9
138	19.6	21.8	32.8#

