

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080921\
 Data File : BF125006.D
 Acq On : 9 Aug 2021 17:29
 Operator : JU/CG
 Sample : PB138266BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB138266BL

Quant Time: Aug 09 18:42:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 11:03:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.816	152	7788	20.000	ng	# 0.00
21) Naphthalene-d8	8.098	136	31561	20.000	ng	# 0.00
39) Acenaphthene-d10	9.857	164	17660	20.000	ng	0.00
64) Phenanthrene-d10	11.339	188	33432	20.000	ng	0.00
76) Chrysene-d12	13.980	240	24293	20.000	ng	# 0.00
86) Perylene-d12	15.433	264	19011	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	62197	130.824	ng	0.01
7) Phenol-d6	6.457	99	75563	126.048	ng	0.00
23) Nitrobenzene-d5	7.381	82	49639	82.277	ng	-0.01
42) 2,4,6-Tribromophenol	10.645	330	20114	127.486	ng	0.00
45) 2-Fluorobiphenyl	9.175	172	101612	84.192	ng	0.00
79) Terphenyl-d14	12.933	244	121721	84.519	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080921\
Data File : BF125006.D
Acq On : 9 Aug 2021 17:29
Operator : JU/CG
Sample : PB138266BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB138266BL

Quant Time: Aug 09 18:42:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 09 11:03:44 2021
Response via : Initial Calibration

