



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080921\
 Data File : BF125004.D
 Acq On : 9 Aug 2021 16:23
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 10 11:32:36 2021






AutoFind: Scans 1615, 1616, 1617; Background Corrected with Scan 1611

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.2	4846	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.2	3942	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.0	5125	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	10051	PASS
199	198	5	9	8.8	880	PASS
275	198	10	60	25.3	2547	PASS
365	198	1	100	1.2	116	PASS
441	198	0.01	100	8.6	868	PASS
442	442	50	100	100.0	7020	PASS
443	442	15	24	17.9	1258	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080921\
 Data File : BF125004.D
 Acq On : 9 Aug 2021 16:23
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Aug 09 17:33:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 11:03:44 2021
 Response via : Initial Calibration



TIC: BF125004.D\data.ms

(77) Benzidine

12.680min (+ 0.000) 0.00 ng




response 42360

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	14.06
183.00	10.10	11.66
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080921\
 Data File : BF125004.D
 Acq On : 9 Aug 2021 16:23
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Aug 09 17:33:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 11:03:44 2021
 Response via : Initial Calibration



TIC: BF125004.D\data.ms

(70) Pentachlorophenol (C)

11.145min (-0.006) 0.00 ng

response 6620

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	57.00
264.00	69.90	60.98
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/9/2021	BNA_F	<u>BF125004.D</u>
Compound Name	Response	Retention Time
DDT	23797	13.49
DDD	0	13.21
DDE	0	19.871
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
0	23797	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP