

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081017\
 Data File : BF097577.D
 Acq On : 11 Aug 2017 00:10
 Operator : SJ/JU
 Sample : I4697-13
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-8B

Manual Integrations
 APPROVED

sohil
 8/14/2017 6:35:08 PM

Quant Time: Aug 11 05:30:02 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 07 16:02:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.39	152	112264	20.00	ng	-0.02
21) Naphthalene-d8	7.67	136	473390	20.00	ng	-0.02
38) Acenaphthene-d10	9.42	164	195003	20.00	ng	-0.02
63) Phenanthrene-d10	10.89	188	290540	20.00	ng	-0.02
75) Chrysene-d12	13.53	240	177159	20.00	ng	-0.01
86) Perylene-d12	14.87	264	189859	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.99	112	752344	101.03	ng	-0.02
7) Phenol-d6	6.07	99	855613	100.64	ng	-0.02
23) Nitrobenzene-d5	6.96	82	510926	63.32	ng	-0.03
41) 2,4,6-Tribromophenol	10.21	330	128851	83.63	ng	-0.02
44) 2-Fluorobiphenyl	8.75	172	693060	60.32	ng	-0.02
78) Terphenyl-d14	12.47	244	384926	44.30	ng	-0.02
Target Compounds						
10) Phenol	6.08	94	28671	2.843	ng	Qvalue # 58
31) Naphthalene	7.69	128	188741	8.458	ng	98
37) 2-Methylnaphthalene	8.39	142	91430	6.471	ng	97
45) 1,1'-Biphenyl	8.85	154	49865	2.994	ng	97
48) Acenaphthylene	9.28	152	53133	2.824	ng	# 96
49) Dimethylphthalate	9.16	163	167418	11.306	ng	98
51) Acenaphthene	9.45	154	214792	19.382	ng	99
54) Dibenzofuran	9.62	168	253104	17.147	ng	98
57) Fluorene	9.96	166	341897	30.818	ng	98
70) Phenanthrene	10.93	178	1841892	118.318	ng	100
71) Anthracene	10.97	178	604741	37.929	ng	100
72) Carbazole	11.14	167	157595	10.050	ng	98
74) Fluoranthene	12.11	202	1457342	95.560	ng	99
77) Pyrene	12.33	202	1344071	92.673	ng	98
80) Benzo(a)anthracene	13.52	228	644363	56.213	ng	99
82) Chrysene	13.55	228	539547	48.203	ng	96
83) Bis(2-ethylhexyl)phthalate	13.52	149	20954	2.175	ng	94
85) Indeno(1,2,3-cd)pyrene	16.06	276	247270	25.763	ng	95
87) Benzo(b)fluoranthene	14.52	252	660332m	57.859	ng	
88) Benzo(k)fluoranthene	14.54	252	153686m	14.131	ng	
89) Benzo(a)pyrene	14.82	252	509369	48.765	ng	97
90) Dibenzo(a,h)anthracene	16.05	278	53362m	6.230	ng	
91) Benzo(g,h,i)perylene	16.40	276	224983	25.047	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF081017\
Data File : BF097577.D
Acq On : 11 Aug 2017 00:10
Operator : SJ/JU
Sample : I4697-13
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-8B

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM

Quant Time: Aug 11 05:30:02 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 16:02:51 2017
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.39 min Scan# 800
Delta R.T. -0.02 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-8B




Tot Ion	Ratio	Lower	Upper
152	100		
150	161.6	126.2	189.2
115	66.6	52.2	78.2

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 101.025 ng
RT: 4.99 min Scan# 561
Delta R.T. -0.02 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Lower	Upper
112	100		
64	68.0	46.0	69.0
63	34.9	23.4	35.0



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 100.640 ng

RT: 6.07 min Scan# 745

Delta R.T. -0.02 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Instrument :

BNA_F





ClientSampleId :

NYSEG-PH4-ESMI-8B

Tot Ion:	99	Resp:	855613
Ion Ratio	Lower	Upper	
99	100		
42	24.6	13.5	20.3#
71	33.6	24.5	36.7

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM



#10

Phenol

Concen: 2.843 ng




RT: 6.08 min Scan# 747


Delta R.T. -0.02 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Tgt Ion:	94	Resp:	28671
Ion Ratio	Lower	Upper	
94	100		
65	43.5	9.9	49.9
66	77.5	23.1	63.1#





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 7.67 min Scan# 1018

Delta R.T. -0.02 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Instrument :

BNA_F


ClientSampleId :

NYSEG-PH4-ESMI-8B




Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	10.4	8.5	12.7	
54	10.3	4.9	7.3	
68	5.5	4.1	6.1	

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM



Abundance Ion 136.00 (135.70 to 136.70): E



#23

Nitrobenzene-d5

Concen: 63.316 ng


RT: 6.96 min Scan# 897

Delta R.T. -0.03 min



Lab File: BF097577.D


Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	44.4	34.0	51.0	
54	69.4	42.2	63.2	



Abundance Ion 82.00 (81.70 to 82.70): BF0





#31

Naphthalene

Concen: 8.458 ng

RT: 7.69 min Scan# 1021

Delta R.T. -0.02 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Instrument :

BNA_F





ClientSampleId :

NYSEG-PH4-ESMI-8B

Tot Ion:128	Resp:	188741
Ion Ratio	Lower	Upper
128 100		
129 10.7	9.0	13.6
127 12.6	10.8	16.2

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM



#37

2-Methylnaphthalene

Concen: 6.471 ng




RT: 8.39 min Scan# 1139


Delta R.T. -0.02 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Tgt Ion:142	Resp:	91430
Ion Ratio	Lower	Upper
142 100		
141 91.5	71.3	106.9
115 32.2	27.1	40.7









#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.42 min Scan# 1314
 Delta R.T. -0.02 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-8B

Tot Ion	Ratio	Resp	Lower	Upper
164	100	195003		
162	104.1	82.6	123.8	
160	45.7	36.2	54.2	




Manual Integrations
 APPROVED


sohil
 8/14/2017 6:35:08 PM



#41
 2,4,6-Tribromophenol
 Concen: 83.631 ng
 RT: 10.21 min Scan# 1449
 Delta R.T. -0.02 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	128851		
332	98.8	76.6	115.0	
141	43.0	31.4	47.2	









#44
2-Fluorobiphenyl
Concen: 60.316 ng
RT: 8.75 min Scan# 1201
Delta R.T. -0.02 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-8B

Tot Ion	Ratio	Resp	Lower	Upper
172	100	693060		
171	37.1		29.6	44.4
170	24.9		19.8	29.8




Manual Integrations
APPROVED


sohil
8/14/2017 6:35:08 PM



#45
1,1'-Biphenyl
Concen: 2.994 ng
RT: 8.85 min Scan# 1218
Delta R.T. -0.02 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	49865		
153	39.6		21.2	61.2
76	12.2		0.0	29.9








#48
Acenaphthylene
Concen: 2.824 ng
RT: 9.28 min Scan# 1291
Delta R.T. -0.02 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-8B

Tot Ion	Ratio	Resp	Lower	Upper
152	100	53133		
151	21.1	16.8	25.2	
153	17.6	10.8	16.2	




Manual Integrations
APPROVED


sohil
8/14/2017 6:35:08 PM



#49
Dimethylphthalate
Concen: 11.306 ng
RT: 9.16 min Scan# 1270
Delta R.T. -0.03 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	167418		
194	3.4	2.6	4.0	
164	9.5	8.4	12.6	





#51

Acenaphthene

Concen: 19.382 ng

RT: 9.45 min Scan# 1320

Delta R.T. -0.02 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-8B

Tot Ion:154 Resp: 214792

Ion Ratio Lower Upper

154 100

153 112.0 90.5 135.7


152 55.1 43.6 65.4

Manual Integrations

APPROVED

sohil




8/14/2017 6:35:08 PM



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 17.147 ng

RT: 9.62 min Scan# 1349

Delta R.T. -0.02 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10


Tgt Ion:168 Resp: 253104

Ion Ratio Lower Upper

168 100

139 39.3 30.6 45.8



169 13.9 10.8 16.2




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E






#57
Fluorene
 Concen: 30.818 ng
 RT: 9.96 min Scan# 1407
 Delta R.T. -0.03 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-8B




Tot Ion	Ratio	Resp	Lower	Upper
166	100	341897		
165	99.8	78.8	118.2	
167	14.6	10.6	16.0	

Manual Integrations
 APPROVED

sohil
 8/14/2017 6:35:08 PM




Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E






#63
Phenanthrene-d10
 Concen: 20.000 ng
 RT: 10.89 min Scan# 1565
 Delta R.T. -0.02 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	290540		
94	13.2	9.4	14.2	
80	12.9	10.2	15.2	



Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BF0
 Ion 80.00 (79.70 to 80.70): BF0









#70
Phenanthrene
Concen: 118.318 ng
RT: 10.93 min Scan# 1571
Delta R.T. -0.01 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-8B

Tot Ion	Ratio	Lower	Upper
178	100		
176	20.0	16.2	24.4
179	15.9	12.6	19.0




Manual Integrations
APPROVED


sohil
8/14/2017 6:35:08 PM



#71
Anthracene
Concen: 37.929 ng
RT: 10.97 min Scan# 1579
Delta R.T. -0.02 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.8	23.8
179	15.8	12.7	19.1





#72

Carbazole

Concen: 10.050 ng

RT: 11.14 min Scan# 1607

Delta R.T. -0.01 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-8B

Tot Ion:167 Resp: 157595

Ion Ratio Lower Upper

167 100


166 22.1 18.1 27.1

139 13.4 11.7 17.5

Manual Integrations
APPROVED

sohil




8/14/2017 6:35:08 PM



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Fluoranthene

Concen: 95.560 ng

RT: 12.11 min Scan# 1772

Delta R.T. -0.01 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10


Tgt Ion:202 Resp: 1457342

Ion Ratio Lower Upper

202 100

101 14.0 0.0 33.2



203 18.3 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 13.53 min Scan# 2013

Delta R.T. -0.01 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-8B

Tot Ion:240 Resp: 177159

Ion Ratio Lower Upper

240 100





120 14.5 5.8 8.8#

236 26.9 20.5 30.7

Manual Integrations
APPROVED

sohil

8/14/2017 6:35:08 PM



#77

Pyrene

Concen: 92.673 ng

RT: 12.33 min Scan# 1810

Delta R.T. -0.01 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10




Tgt Ion:202 Resp: 1344071


Ion Ratio Lower Upper

202 100

200 22.9 17.6 26.4

203 18.4 14.4 21.6









#78
 Terphenyl-d14
 Concen: 44.300 ng
 RT: 12.47 min Scan# 1834
 Delta R.T. -0.02 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-8B

Tot Ion	Ratio	Lower	Upper
244	100		
212	6.8	6.0	9.0
122	16.0	10.3	15.5




Manual Integrations
 APPROVED


sohil
 8/14/2017 6:35:08 PM



#80
 Benzo(a)anthracene
 Concen: 56.213 ng
 RT: 13.52 min Scan# 2011
 Delta R.T. -0.01 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.0	22.6	33.8
229	20.5	16.3	24.5





#82

Chrysene

Concen: 48.203 ng

RT: 13.55 min Scan# 2017

Delta R.T. -0.01 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Instrument :

BNA_F


ClientSampleId :

NYSEG-PH4-ESMI-8B




Tot Ion:	228	Resp:	539547
Ion Ratio	Lower	Upper	
228	100		
226	32.2	24.9	37.3
229	22.9	15.8	23.6

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#83

Bis(2-ethylhexyl)phthalate

Concen: 2.175 ng

RT: 13.52 min Scan# 2012




Delta R.T. -0.02 min


Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Tgt Ion:	149	Resp:	20954
Ion Ratio	Lower	Upper	
149	100		
167	22.6	21.0	31.4
279	2.6	1.9	2.9

Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E






#85
 Indeno(1,2,3-cd)pyrene
 Concen: 25.763 ng
 RT: 16.06 min Scan# 2443
 Delta R.T. -0.01 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-8B




Tot Ion	Ratio	Lower	Upper
276	100		
138	31.9	27.8	41.6
277	22.3	20.2	30.4

Manual Integrations
 APPROVED

sohil
 8/14/2017 6:35:08 PM




Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E






#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 14.87 min Scan# 2242
 Delta R.T. 0.01 min
 Lab File: BF097577.D
 Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Lower	Upper
264	100		
260	22.8	19.5	29.3
265	24.2	17.2	25.8



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E









#87
Benzo(b)fluoranthene
Concen: 57.859 ng m
RT: 14.52 min Scan# 2182
Delta R.T. -0.01 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-8B

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.7	18.1	27.1
125	15.6	9.8	14.8




Manual Integrations
APPROVED


sohil
8/14/2017 6:35:08 PM



#88
Benzo(k)fluoranthene
Concen: 14.131 ng m
RT: 14.54 min Scan# 2185
Delta R.T. -0.01 min
Lab File: BF097577.D
Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.5	18.4	27.6
125	15.8	13.1	19.7





#89

Benzo(a)pyrene

Concen: 48.765 ng

RT: 14.82 min Scan# 2233

Delta R.T. 0.00 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10





Instrument :
BNA_F

ClientSampleId :
NYSEG-PH4-ESMI-8B

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.5	26.3
125	16.3	11.0	16.4

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM



#90

Dibenzo(a,h)anthracene

Concen: 6.230 ng m




RT: 16.05 min Scan# 2442




Delta R.T. -0.01 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Tgt Ion	Ratio	Lower	Upper
278	100		
139	0.0	16.6	24.8#
279	26.2	19.3	28.9





#91

Benzo(a,h,i)perylene

Concen: 25.047 ng

RT: 16.40 min Scan# 2502

Delta R.T. -0.01 min

Lab File: BF097577.D

Acq: 11 Aug 2017 00:10

Tot Ion:	276	Resp:	224983
Ion Ratio	Lower	Upper	
276	100		
277	24.5	18.6	28.0
138	35.3	25.4	38.0

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-8B

Manual Integrations
APPROVED

sohil
8/14/2017 6:35:08 PM

