

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081021\
 Data File : BF125030.D
 Acq On : 10 Aug 2021 14:36
 Operator : JU/CG
 Sample : SP5637
 Misc : 625 SURROGATE
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP5637

Quant Time: Aug 10 15:13:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 10 11:32:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.810	152	6069	20.000	ng	0.00
21) Naphthalene-d8	8.092	136	24428	20.000	ng	0.00
39) Acenaphthene-d10	9.845	164	9726	20.000	ng	0.00
64) Phenanthrene-d10	11.333	188	26399	20.000	ng	0.00
76) Chrysene-d12	13.968	240	19875	20.000	ng	# 0.00
86) Perylene-d12	15.415	264	17038	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.445	112	37387	100.913	ng	0.01
7) Phenol-d6	6.445	99	52522	112.428	ng	0.00
23) Nitrobenzene-d5	7.375	82	52355	112.119	ng	0.00
42) 2,4,6-Tribromophenol	10.639	330	13528	155.688	ng	0.00
45) 2-Fluorobiphenyl	9.169	172	71543	107.634	ng	0.00
79) Terphenyl-d14	12.921	244	127173	107.934	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081021\
Data File : BF125030.D
Acq On : 10 Aug 2021 14:36
Operator : JU/CG
Sample : SP5637
Misc : 625 SURROGATE
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP5637

Quant Time: Aug 10 15:13:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 10 11:32:36 2021
Response via : Initial Calibration

