

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081024\
 Data File : BF138914.D
 Acq On : 10 Aug 2024 17:21
 Operator : RC/JU
 Sample : P3476-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MUL07

Quant Time: Aug 12 01:16:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	34659	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	144676	20.000	ng	0.00
39) Acenaphthene-d10	9.869	164	74060	20.000	ng	-0.01
64) Phenanthrene-d10	11.357	188	112514	20.000	ng	-0.01
76) Chrysene-d12	13.998	240	66059	20.000	ng	# 0.00
86) Perylene-d12	15.462	264	79830	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	200794	89.430	ng	0.00
7) Phenol-d6	6.481	99	173151	57.439	ng	0.00
23) Nitrobenzene-d5	7.404	82	276827	93.550	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	86232	142.144	ng	0.00
45) 2-Fluorobiphenyl	9.192	172	485374	98.471	ng	-0.01
79) Terphenyl-d14	12.939	244	351881	89.184	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081024\
Data File : BF138914.D
Acq On : 10 Aug 2024 17:21
Operator : RC/JU
Sample : P3476-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MUL07

Quant Time: Aug 12 01:16:56 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration

