

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081117\
 Data File : BF097621.D
 Acq On : 12 Aug 2017 5:31
 Operator : SJ/JU
 Sample : I4646-04 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-L8-BASE

Manual Integrations
 APPROVED

sohil
 8/14/2017 6:55:21 PM

Quant Time: Aug 14 01:32:41 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 11 15:55:57 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.51	152	108236	20.00	ng	0.00
21) Naphthalene-d8	9.54	136	441649	20.00	ng	0.00
38) Acenaphthene-d10	12.36	164	186818	20.00	ng	0.00
63) Phenanthrene-d10	14.76	188	277670	20.00	ng	0.00
75) Chrysene-d12	18.46	240	158503	20.00	ng	0.00
86) Perylene-d12	20.13	264	154882	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	89772	13.18	ng	0.05
7) Phenol-d6	7.10	99	132586	15.70	ng	0.04
23) Nitrobenzene-d5	8.43	82	81595	11.58	ng	0.01
41) 2,4,6-Tribromophenol	13.67	330	19000	13.38	ng	0.00
44) 2-Fluorobiphenyl	11.32	172	128388	10.37	ng	0.00
78) Terphenyl-d14	17.12	244	57565	7.52	ng	0.00
Target Compounds						
11) bis(2-Chloroethyl)ether	7.16	93	22836	3.078	ng	92
31) Naphthalene	9.57	128	57291	2.590	ng	99
49) Dimethylphthalate	12.02	163	33962	2.359	ng	99
67) Hexachlorobenzene	14.15	284	30316	9.661	ng	# 84
70) Phenanthrene	14.80	178	69117	4.322	ng	98
77) Pyrene	16.87	202	28217	2.153	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081117\
Data File : BF097621.D
Acq On : 12 Aug 2017 5:31
Operator : SJ/JU
Sample : I4646-04 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-L8-BASE

Manual Integrations
APPROVED

sohil
8/14/2017 6:55:21 PM

Quant Time: Aug 14 01:32:41 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 11 15:55:57 2017
Response via : Initial Calibration

