



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081121\
 Data File : BF125051.D
 Acq On : 11 Aug 2021 8:42
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 10 11:32:36 2021






AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1610

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.0	3308	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	41.8	3212	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.4	3568	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	7685	PASS
199	198	5	9	6.8	520	PASS
275	198	10	60	28.4	2184	PASS
365	198	1	100	1.5	118	PASS
441	198	0.01	100	10.4	802	PASS
442	442	50	100	100.0	5848	PASS
443	442	15	24	19.2	1125	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081121\
 Data File : BF125051.D
 Acq On : 11 Aug 2021 8:42
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Aug 11 10:57:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 10 11:32:36 2021
 Response via : Initial Calibration



TIC: BF125051.D\data.ms

(77) Benzidine

12.674min (+ 0.006) 0.00 ng




response 30603

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	14.14
183.00	10.10	10.46
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081121\
Data File : BF125051.D
Acq On : 11 Aug 2021 8:42
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 11 10:57:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 10 11:32:36 2021
Response via : Initial Calibration



TIC: BF125051.D\data.ms

(70) Pentachlorophenol (C)

11.139min (-0.000) 0.00 ng

response 4738

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	59.18
264.00	69.90	57.08
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/11/2021	BNA_F	<u>BF125051.D</u>
Compound Name	Response	Retention Time
DDT	16925	13.48
DDD	0	13.23
DDE	0	12.74
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
0	16925	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP