

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081217\
 Data File : BF097660.D
 Acq On : 13 Aug 2017 7:20
 Operator : SJ/JU
 Sample : I4697-02DL 10X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-1BDL

Quant Time: Aug 14 05:27:31 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 11 15:55:57 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	121216	20.00	ng	0.00
21) Naphthalene-d8	9.53	136	502282	20.00	ng	-0.01
38) Acenaphthene-d10	12.35	164	199403	20.00	ng	-0.02
63) Phenanthrene-d10	14.75	188	274633	20.00	ng	-0.01
75) Chrysene-d12	18.45	240	190294	20.00	ng	-0.01
86) Perylene-d12	20.12	264	185786	20.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.49	112	41077	5.39	ng	0.09
7) Phenol-d6	7.12	99	77688	8.21	ng	0.06
23) Nitrobenzene-d5	8.45	82	45239	5.65	ng	0.02
41) 2,4,6-Tribromophenol	13.66	330	7941	5.24	ng	0.00
44) 2-Fluorobiphenyl	11.30	172	69325	5.25	ng	-0.02
78) Terphenyl-d14	17.10	244	31754	3.45	ng	-0.02


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Naphthalene	9.56	128	180005	7.15	ng	98
33) 4-Chloroaniline	9.56	127	22764	2.18	ng	# 37
37) 2-Methylnaphthalene	10.70	142	34470	2.18	ng	94
48) Acenaphthylene	12.13	152	129034	6.18	ng	97
50) 2,6-Dinitrotoluene	12.35	165	24625	7.76	ng	# 33
51) Acenaphthene	12.40	154	74170	5.93	ng	98
54) Dibenzofuran	12.70	168	122767	7.01	ng	97
55) 4-Nitrophenol	12.70	139	46341	29.28	ng	# 28
57) Fluorene	13.24	166	135804	9.43	ng	97
70) Phenanthrene	14.79	178	787868	49.81	ng	99
71) Anthracene	14.87	178	322865	19.95	ng	100
72) Carbazole	15.18	167	48127	3.12	ng	98
74) Fluoranthene	16.56	202	839052	58.70	ng	98
77) Pyrene	16.86	202	693364	44.06	ng	96
80) Benzo(a)anthracene	18.48	228	295731	26.61	ng	96
82) Chrysene	18.48	228	295731	26.75	ng	98
85) Indeno(1,2,3-cd)pyrene	21.33	276	136344	15.86	ng	95
87) Benzo(b)fluoranthene	19.72	252	473792	42.47	ng	# 97
88) Benzo(k)fluoranthene	19.72	252	473792	42.17	ng	98
89) Benzo(a)pyrene	20.06	252	280534	27.40	ng	96
90) Dibenzo(a,h)anthracene	21.33	278	33677	4.59	ng	# 76
91) Benzo(g,h,i)perylene	21.69	276	118805	14.77	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081217\
Data File : BF097660.D
Acq On : 13 Aug 2017 7:20
Operator : SJ/JU
Sample : I4697-02DL 10X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL

Quant Time: Aug 14 05:27:31 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 11 15:55:57 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.50 min Scan# 989
Delta R.T. -0.01 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL

Tot Ion:152 Resp: 121216
Ion Ratio Lower Upper
152 100
150 153.2 126.2 189.2
115 65.9 52.2 78.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 5.39 ng
RT: 5.49 min Scan# 647
Delta R.T. 0.09 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Tgt Ion:112 Resp: 41077
Ion Ratio Lower Upper
112 100
64 62.3 46.0 69.0
63 35.3 23.4 35.0#



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 8.21 ng

RT: 7.12 min Scan# 924

Delta R.T. 0.06 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-1BDL


Tot Ion: 99 Resp: 77688

Ion Ratio Lower Upper

99 100

42 21.8 13.5 20.3#




71 30.3 24.5 36.7



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 9.53 min Scan# 1333

Delta R.T. -0.01 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20

Tgt Ion: 136 Resp: 502282


Ion Ratio Lower Upper

136 100

137 10.9 8.5 12.7

54 11.6 4.9 7.3#

68 6.5 4.1 6.1#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 5.65 ng
 RT: 8.45 min Scan# 1149
 Delta R.T. 0.02 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20




Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-1BDL


Tot Ion	82	Resp	45239
Ion Ratio	Lower	Upper	
82	100		
128	41.9	34.0	51.0
54	66.2	42.2	63.2#



#31
 Naphthalene
 Concen: 7.15 ng
 RT: 9.56 min Scan# 1339
 Delta R.T. -0.01 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20

Tgt Ion	128	Resp	180005
Ion Ratio	Lower	Upper	
128	100		
129	9.9	9.0	13.6
127	12.9	10.8	16.2









#33
4-Chloroaniline
Concen: 2.18 ng
RT: 9.56 min Scan# 1339
Delta R.T. -0.15 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20




Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL


Tot Ion:	127	Resp:	22764
Ion	Ratio	Lower	Upper
127	100		
129	76.6	26.2	39.2#
65	0.0	27.8	41.8#
92	0.0	16.8	25.2#



#37
2-Methylnaphthalene
Concen: 2.18 ng
RT: 10.70 min Scan# 1532
Delta R.T. -0.00 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Tgt Ion:	142	Resp:	34470
Ion	Ratio	Lower	Upper
142	100		
141	96.7	71.3	106.9
115	34.0	27.1	40.7









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 12.35 min Scan# 1813
Delta R.T. -0.02 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20




Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL


Tot Ion:164 Resp: 199403
Ion Ratio Lower Upper
164 100
162 100.4 82.6 123.8
160 45.1 36.2 54.2



#41
2,4,6-Tribromophenol
Concen: 5.24 ng
RT: 13.66 min Scan# 2036
Delta R.T. -0.00 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Tgt Ion:330 Resp: 7941
Ion Ratio Lower Upper
330 100
332 93.4 76.6 115.0
141 50.2 31.4 47.2#









#44
2-Fluorobiphenyl
Concen: 5.25 ng
RT: 11.30 min Scan# 1635
Delta R.T. -0.02 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20




Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL


Tot Ion	Ratio	Lower	Upper
172	100		
171	33.4	29.6	44.4
170	22.9	19.8	29.8



#48
Acenaphthylene
Concen: 6.18 ng
RT: 12.13 min Scan# 1775
Delta R.T. -0.01 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.6	16.8	25.2
153	11.8	10.8	16.2






#50
 2,6-Dinitrotoluene
 Concen: 7.76 ng
 RT: 12.35 min Scan# 1813
 Delta R.T. 0.24 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20

Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-1BDL




Tot Ion	Ratio	Lower	Upper
165	100		
63	0.0	34.3	51.5#
89	2.0	37.1	55.7#



Abundance Ion 165.00 (164.70 to 165.70): E


Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0



#51
 Acenaphthene
 Concen: 5.93 ng
 RT: 12.40 min Scan# 1822
 Delta R.T. -0.02 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20



Tgt Ion	Ratio	Lower	Upper
154	100		
153	109.7	90.5	135.7
152	54.0	43.6	65.4




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#54
Dibenzofuran
Concen: 7.01 ng
RT: 12.70 min Scan# 1872
Delta R.T. -0.01 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL

Tot Ion	Ratio	Lower	Upper
168	100		
139	36.1	30.6	45.8
169	13.3	10.8	16.2




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#55
4-Nitrophenol
Concen: 29.28 ng
RT: 12.70 min Scan# 1872
Delta R.T. -0.07 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Tgt Ion	Ratio	Lower	Upper
139	100		
109	1.1	34.1	74.1#
65	3.0	31.4	71.4#



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0









#57
 Fluorene
 Concen: 9.43 ng
 RT: 13.24 min Scan# 1965
 Delta R.T. -0.02 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20




Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-1BDL


Tot Ion:166 Resp: 135804
 Ion Ratio Lower Upper
 166 100
 165 101.3 78.8 118.2
 167 13.4 10.6 16.0



#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 14.75 min Scan# 2221
 Delta R.T. -0.01 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20

Tgt Ion:188 Resp: 274633
 Ion Ratio Lower Upper
 188 100
 94 11.9 9.4 14.2
 80 12.6 10.2 15.2









#70
 Phenanthrene
 Concen: 49.81 ng
 RT: 14.79 min Scan# 2228
 Delta R.T. -0.01 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20




Instrument :
 BNA_F
 ClientSampleId :
 NYSEG-PH4-ESMI-1BDL


Tot Ion	178	Resp	787868
Ion Ratio	100	Lower	Upper
176	20.1	16.2	24.4
179	15.3	12.6	19.0



#71
 Anthracene
 Concen: 19.95 ng
 RT: 14.87 min Scan# 2242
 Delta R.T. -0.01 min
 Lab File: BF097660.D
 Acq: 13 Aug 2017 7:20

Tgt Ion	178	Resp	322865
Ion Ratio	100	Lower	Upper
176	19.6	15.8	23.8
179	15.8	12.7	19.1





#72

Carbazole

Concen: 3.12 ng

RT: 15.18 min Scan# 2294

Delta R.T. 0.01 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-1BDL


Tot Ion:167 Resp: 48127

Ion Ratio Lower Upper

167 100

166 23.4 18.1 27.1




139 15.5 11.7 17.5



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Fluoranthene

Concen: 58.70 ng

RT: 16.56 min Scan# 2528

Delta R.T. -0.01 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20


Tgt Ion:202 Resp: 839052

Ion Ratio Lower Upper

202 100

101 14.9 0.0 33.2



203 17.6 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 18.45 min Scan# 2850

Delta R.T. -0.01 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-1BDL


Tot Ion:240 Resp: 190294

Ion Ratio Lower Upper

240 100

120 15.2 5.8 8.8#




236 25.5 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 44.06 ng

RT: 16.86 min Scan# 2579

Delta R.T. -0.01 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20


Tgt Ion:202 Resp: 693364

Ion Ratio Lower Upper

202 100

200 22.3 17.6 26.4



203 21.1 14.4 21.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 3.45 ng

RT: 17.10 min Scan# 2621

Delta R.T. -0.02 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-1BDL


Tot Ion:244 Resp: 31754

Ion Ratio Lower Upper

244 100

212 7.1 6.0 9.0




122 17.2 10.3 15.5#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 26.61 ng

RT: 18.48 min Scan# 2855

Delta R.T. 0.03 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20


Tgt Ion:228 Resp: 295731

Ion Ratio Lower Upper

228 100

226 31.3 22.6 33.8



229 21.4 16.3 24.5




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 26.75 ng

RT: 18.48 min Scan# 2855

Delta R.T. -0.02 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-1BDL


Tot Ion:228 Resp: 295731

Ion Ratio Lower Upper

228 100

226 31.3 24.9 37.3




229 21.4 15.8 23.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 15.86 ng

RT: 21.33 min Scan# 3340

Delta R.T. -0.01 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20


Tgt Ion:276 Resp: 136344

Ion Ratio Lower Upper

276 100

138 31.5 27.8 41.6



277 23.6 20.2 30.4




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Pervlene-d12
Concen: 20.00 ng
RT: 20.12 min Scan# 3133
Delta R.T. -0.01 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20




Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL


Tot Ion	Ratio	Lower	Upper
264	100		
260	23.0	19.5	29.3
265	21.5	17.2	25.8



#87
Benzo(b)fluoranthene
Concen: 42.47 ng
RT: 19.72 min Scan# 3065
Delta R.T. -0.01 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.4	18.1	27.1
125	14.9	9.8	14.8#









#88
Benzo(k)fluoranthene
Concen: 42.17 ng
RT: 19.72 min Scan# 3065
Delta R.T. -0.04 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20




Instrument :
BNA_F
ClientSampleId :
NYSEG-PH4-ESMI-1BDL


Tot Ion	Ratio	Lower	Upper
252	100		
253	22.4	18.4	27.6
125	14.9	13.1	19.7



#89
Benzo(a)pyrene
Concen: 27.40 ng
RT: 20.06 min Scan# 3124
Delta R.T. -0.01 min
Lab File: BF097660.D
Acq: 13 Aug 2017 7:20

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.1	17.5	26.3
125	16.4	11.0	16.4





#90

Dibenzo(a,h)anthracene

Concen: 4.59 ng

RT: 21.33 min Scan# 3340

Delta R.T. -0.01 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20

Instrument :

BNA_F

ClientSampleId :

NYSEG-PH4-ESMI-1BDL


Tot Ion:278 Resp: 33677

Ion Ratio Lower Upper

278 100

139 0.0 16.6 24.8#




279 27.2 19.3 28.9



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 14.77 ng

RT: 21.69 min Scan# 3400

Delta R.T. 0.00 min

Lab File: BF097660.D

Acq: 13 Aug 2017 7:20


Tgt Ion:276 Resp: 118805

Ion Ratio Lower Upper

276 100

277 25.1 18.6 28.0

138 33.7 25.4 38.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

