

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081217\
 Data File : BF097671.D
 Acq On : 13 Aug 2017 13:19
 Operator : SJ/JU
 Sample : I4682-03
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-G10-WSW

Manual Integrations
 APPROVED

sohil
 8/14/2017 7:08:37 PM

Quant Time: Aug 14 08:09:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 11 15:55:57 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	128500	20.00	ng	0.00
21) Naphthalene-d8	9.53	136	518092	20.00	ng	-0.01
38) Acenaphthene-d10	12.35	164	192025	20.00	ng	-0.02
63) Phenanthrene-d10	14.75	188	247359	20.00	ng	-0.01
75) Chrysene-d12	18.45	240	216724	20.00	ng	-0.01
86) Perylene-d12	20.11	264	192862	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	543925	67.28	ng	0.02
7) Phenol-d6	7.07	99	701049	69.90	ng	0.01
23) Nitrobenzene-d5	8.41	82	399585	48.36	ng	-0.01
41) 2,4,6-Tribromophenol	13.65	330	76486	52.40	ng	-0.01
44) 2-Fluorobiphenyl	11.30	172	496760	39.04	ng	-0.02
78) Terphenyl-d14	17.10	244	239201	22.84	ng	-0.02
Target Compounds						
49) Dimethylphthalate	12.00	163	134849	9.112	ng	97
69) Pentachlorophenol	14.51	266	14314	29.030	ng	97
77) Pyrene	16.86	202	51663	2.883	ng	99
87) Benzo(b)fluoranthene	19.71	252	33806m	2.919	ng	
91) Benzo(g,h,i)perylene	21.70	276	17127	2.051	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081217\
Data File : BF097671.D
Acq On : 13 Aug 2017 13:19
Operator : SJ/JU
Sample : I4682-03
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-G10-WSW

Manual Integrations
APPROVED

sohil
8/14/2017 7:08:37 PM

Quant Time: Aug 14 08:09:57 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 11 15:55:57 2017
Response via : Initial Calibration

