

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081224\
 Data File : BF138941.D
 Acq On : 12 Aug 2024 18:57
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 13 00:59:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	80	0.00
2	1,4-Dioxane	40.000	34.681	13.3	69	0.01
3	Pyridine	40.000	36.472	8.8	73	0.02
4	n-Nitrosodimethylamine	40.000	44.195	-10.5	89	0.03
5 S	2-Fluorophenol	80.000	78.307	2.1	80	0.00
6	Aniline	40.000	40.413	-1.0	82	0.00
7 S	Phenol-d6	80.000	80.323	-0.4	83	0.00
8	2-Chlorophenol	40.000	40.656	-1.6	84	0.00
9	Benzaldehyde	40.000	36.866	7.8	85	0.00
10 C	Phenol	40.000	40.005	-0.0	83	0.00
11	bis(2-Chloroethyl)ether	40.000	37.731	5.7	78	0.00
12	1,3-Dichlorobenzene	40.000	39.486	1.3	82	0.00
13 C	1,4-Dichlorobenzene	40.000	39.082	2.3	81	0.00
14	1,2-Dichlorobenzene	40.000	40.863	-2.2	84	0.00
15	Benzyl Alcohol	40.000	44.187	-10.5	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.273	6.8	77	0.00
17	2-Methylphenol	40.000	40.744	-1.9	83	0.00
18	Hexachloroethane	40.000	41.448	-3.6	85	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	43.084	-7.7	91	0.00
20	3+4-Methylphenols	40.000	42.634	-6.6	90	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	85	0.00
22	Acetophenone	40.000	42.958	-7.4	93	0.00
23 S	Nitrobenzene-d5	80.000	83.120	-3.9	89	0.00
24	Nitrobenzene	40.000	40.708	-1.8	87	0.00
25	Isophorone	40.000	42.082	-5.2	92	0.00
26 C	2-Nitrophenol	40.000	42.229	-5.6	88	0.00
27	2,4-Dimethylphenol	40.000	39.991	0.0	85	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.828	0.4	86	0.00
29 C	2,4-Dichlorophenol	40.000	42.113	-5.3	90	0.00
30	1,2,4-Trichlorobenzene	40.000	40.203	-0.5	86	0.00
31	Naphthalene	40.000	40.346	-0.9	87	0.00
32	Benzoic acid	40.000	36.247	9.4	78	0.04
33	4-Chloroaniline	40.000	39.572	1.1	86	0.00
34 C	Hexachlorobutadiene	40.000	42.833	-7.1	92	0.00
35	Caprolactam	40.000	44.350	-10.9	99	0.02
36 C	4-Chloro-3-methylphenol	40.000	43.701	-9.3	95	0.01
37	2-Methylnaphthalene	40.000	41.582	-4.0	91	0.00
38	1-Methylnaphthalene	40.000	41.368	-3.4	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.595	1.0	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.557	6.1	92	0.00
42 S	2,4,6-Tribromophenol	80.000	85.707	-7.1	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.544	1.1	94	0.00
44	2,4,5-Trichlorophenol	40.000	40.010	-0.0	95	0.00
45 S	2-Fluorobiphenyl	80.000	81.489	-1.9	98	0.00
46	1,1'-Biphenyl	40.000	39.678	0.8	95	0.00
47	2-Chloronaphthalene	40.000	39.260	1.9	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081224\
 Data File : BF138941.D
 Acq On : 12 Aug 2024 18:57
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 13 00:59:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.002	-5.0	101	0.00
49	Acenaphthylene	40.000	39.301	1.7	94	0.00
50	Dimethylphthalate	40.000	42.482	-6.2	104	0.00
51	2,6-Dinitrotoluene	40.000	42.853	-7.1	102	0.00
52 C	Acenaphthene	40.000	39.607	1.0	96	0.00
53	3-Nitroaniline	40.000	39.656	0.9	97	0.00
54 P	2,4-Dinitrophenol	40.000	49.611	-24.0	123	0.00
55	Dibenzofuran	40.000	40.392	-1.0	99	0.00
56 P	4-Nitrophenol	40.000	48.291	-20.7	117	0.02
57	2,4-Dinitrotoluene	40.000	44.626	-11.6	107	0.01
58	Fluorene	40.000	41.711	-4.3	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.335	-5.8	104	0.00
60	Diethylphthalate	40.000	45.826	-14.6	114	0.00
61	4-Chlorophenyl-phenylether	40.000	42.400	-6.0	105	0.00
62	4-Nitroaniline	40.000	41.287	-3.2	102	0.01
63	Azobenzene	40.000	41.055	-2.6	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.723	-1.8	106	0.01
66 c	n-Nitrosodiphenylamine	40.000	39.654	0.9	104	0.00
67	4-Bromophenyl-phenylether	40.000	40.512	-1.3	107	0.00
68	Hexachlorobenzene	40.000	39.922	0.2	107	0.00
69	Atrazine	40.000	39.795	0.5	105	0.00
70 C	Pentachlorophenol	40.000	42.996	-7.5	112	0.00
71	Phenanthrene	40.000	40.592	-1.5	108	0.00
72	Anthracene	40.000	40.321	-0.8	108	0.00
73	Carbazole	40.000	39.994	0.0	107	0.00
74	Di-n-butylphthalate	40.000	47.197	-18.0	124	0.00
75 C	Fluoranthene	40.000	40.919	-2.3	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	27.552	31.1#	59	0.00
78	Pyrene	40.000	41.245	-3.1	106	0.00
79 S	Terphenyl-d14	80.000	84.266	-5.3	108	0.00
80	Butylbenzylphthalate	40.000	44.235	-10.6	101	-0.01
81	Benzo(a)anthracene	40.000	38.119	4.7	89	0.00
82	3,3'-Dichlorobenzidine	40.000	40.462	-1.2	94	0.00
83	Chrysene	40.000	39.203	2.0	95	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.141	2.1	87	-0.01
85 c	Di-n-octyl phthalate	40.000	35.264	11.8	79	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	77	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.967	7.6	73	0.01
88	Benzo(b)fluoranthene	40.000	43.198	-8.0	84	0.00
89	Benzo(k)fluoranthene	40.000	37.471	6.3	77	0.00
90 C	Benzo(a)pyrene	40.000	40.008	-0.0	79	0.00
91	Dibenzo(a,h)anthracene	40.000	37.139	7.2	74	0.00
92	Benzo(g,h,i)perylene	40.000	35.919	10.2	70	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081224\
Data File : BF138941.D
Acq On : 12 Aug 2024 18:57
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 13 00:59:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0