

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081318\
 Data File : BF108115.D
 Acq On : 13 Aug 2018 14:49
 Operator : JU/SJ
 Sample : J4272-15 2X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-080918-H

Manual Integrations
 APPROVED

Sohil
 8/14/2018 2:36:29 PM

Quant Time: Aug 13 17:19:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 12:24:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.02	152	283882	20.00	ng	0.01
21) Naphthalene-d8	8.31	136	1027643	20.00	ng	0.00
38) Acenaphthene-d10	10.07	164	461204	20.00	ng	0.00
63) Phenanthrene-d10	11.57	188	755216	20.00	ng	0.01
75) Chrysene-d12	14.23	240	699946	20.00	ng	0.01
86) Perylene-d12	15.81	264	562406	20.00	ng	0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.64	112	763004	48.66	ng	0.01
7) Phenol-d6	6.64	99	1013452	48.64	ng	0.00
23) Nitrobenzene-d5	7.58	82	656164	35.63	ng	0.00
41) 2,4,6-Tribromophenol	10.87	330	200217	43.52	ng	0.00
44) 2-Fluorobiphenyl	9.38	172	1177322	40.98	ng	0.00
78) Terphenyl-d14	13.16	244	1034521	37.58	ng	0.01
Target Compounds						
49) Dimethylphthalate	9.77	163	112038	3.487	ng	Qvalue # 99
70) Phenanthrene	11.59	178	207736	5.832	ng	99
71) Anthracene	11.65	178	73911	2.024	ng	98
74) Fluoranthene	12.79	202	477214	12.275	ng	95
77) Pyrene	13.02	202	446755	10.082	ng	99
80) Benzo(a)anthracene	14.22	228	254678	6.340	ng	96
82) Chrysene	14.26	228	207823	5.643	ng	98
85) Indeno(1,2,3-cd)pyrene	17.45	276	72966	2.287	ng	# 90
87) Benzo(b)fluoranthene	15.34	252	252214m	7.505	ng	
88) Benzo(k)fluoranthene	15.36	252	84599m	2.739	ng	
89) Benzo(a)pyrene	15.74	252	167327	5.560	ng	94
91) Benzo(g,h,i)perylene	17.95	276	67956	2.772	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081318\
Data File : BF108115.D
Acq On : 13 Aug 2018 14:49
Operator : JU/SJ
Sample : J4272-15 2X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-03-080918-H

Manual Integrations
APPROVED

Sohil
8/14/2018 2:36:29 PM

Quant Time: Aug 13 17:19:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 12:24:24 2018
Response via : Initial Calibration

