

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129768.D
 Acq On : 14 Aug 2022 17:47
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF081422

Quant Time: Aug 15 02:36:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.498	0.456	8.4	106	0.00
3	Pyridine	1.304	1.206	7.5	107	0.00
4	n-Nitrosodimethylamine	0.601	0.568	5.5	106	0.00
5 S	2-Fluorophenol	1.208	1.105	8.5	108	0.00
6	Aniline	1.724	1.581	8.3	108	0.00
7 S	Phenol-d6	1.513	1.368	9.6	107	0.00
8	2-Chlorophenol	1.345	1.233	8.3	107	0.00
9	Benzaldehyde	0.893	0.799	10.5	107	0.00
10 C	Phenol	1.659	1.515	8.7	108	0.00
11	bis(2-Chloroethyl)ether	1.260	1.159	8.0	107	0.00
12	1,3-Dichlorobenzene	1.452	1.339	7.8	108	0.00
13 C	1,4-Dichlorobenzene	1.483	1.363	8.1	109	0.00
14	1,2-Dichlorobenzene	1.378	1.255	8.9	107	0.00
15	Benzyl Alcohol	1.142	1.059	7.3	109	0.00
16	2,2'-oxybis(1-Chloropropane	1.688	1.541	8.7	107	0.00
17	2-Methylphenol	1.080	0.991	8.2	109	0.00
18	Hexachloroethane	0.557	0.517	7.2	109	0.00
19 P	n-Nitroso-di-n-propylamine	0.956	0.891	6.8	108	0.00
20	3+4-Methylphenols	1.449	1.332	8.1	108	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.487	0.448	8.0	108	0.00
23 S	Nitrobenzene-d5	0.374	0.345	7.8	107	0.00
24	Nitrobenzene	0.377	0.349	7.4	108	0.00
25	Isophorone	0.643	0.600	6.7	108	0.00
26 C	2-Nitrophenol	0.185	0.176	4.9	109	0.00
27	2,4-Dimethylphenol	0.266	0.247	7.1	107	0.00
28	bis(2-Chloroethoxy)methane	0.375	0.352	6.1	107	0.00
29 C	2,4-Dichlorophenol	0.291	0.270	7.2	106	0.00
30	1,2,4-Trichlorobenzene	0.307	0.282	8.1	107	0.00
31	Naphthalene	1.046	0.960	8.2	108	0.00
32	Benzoic acid	0.113	0.118	-4.4	116	0.00
33	4-Chloroaniline	0.411	0.387	5.8	109	0.00
34 C	Hexachlorobutadiene	0.187	0.173	7.5	106	0.00
35	Caprolactam	0.090	0.084	6.7	107	0.00
36 C	4-Chloro-3-methylphenol	0.313	0.291	7.0	108	0.00
37	2-Methylnaphthalene	0.686	0.640	6.7	108	0.00
38	1-Methylnaphthalene	0.667	0.621	6.9	109	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	0.556	0.509	8.5	107	0.00
41 P	Hexachlorocyclopentadiene	0.117	0.125	-6.8	114	0.00
42 S	2,4,6-Tribromophenol	0.201	0.187	7.0	107	0.00
43 C	2,4,6-Trichlorophenol	0.361	0.341	5.5	109	0.00
44	2,4,5-Trichlorophenol	0.389	0.360	7.5	105	0.00
45 S	2-Fluorobiphenyl	1.317	1.187	9.9	107	0.00
46	1,1'-Biphenyl	1.519	1.390	8.5	107	0.00
47	2-Chloronaphthalene	1.202	1.096	8.8	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129768.D
 Acq On : 14 Aug 2022 17:47
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF081422

Quant Time: Aug 15 02:36:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.380	0.357	6.1	108	0.00
49	Acenaphthylene	1.823	1.655	9.2	107	0.00
50	Dimethylphthalate	1.434	1.329	7.3	109	0.00
51	2,6-Dinitrotoluene	0.311	0.291	6.4	107	0.00
52 C	Acenaphthene	1.106	1.016	8.1	108	0.00
53	3-Nitroaniline	0.344	0.317	7.8	107	0.00
54 P	2,4-Dinitrophenol	0.118	0.128	-8.5	113	0.00
55	Dibenzofuran	1.662	1.505	9.4	108	0.00
56 P	4-Nitrophenol	0.153	0.154	-0.7	110	0.00
57	2,4-Dinitrotoluene	0.404	0.371	8.2	110	0.00
58	Fluorene	1.387	1.256	9.4	108	0.00
59	2,3,4,6-Tetrachlorophenol	0.285	0.270	5.3	106	0.00
60	Diethylphthalate	1.430	1.294	9.5	107	0.00
61	4-Chlorophenyl-phenylether	0.623	0.565	9.3	107	0.00
62	4-Nitroaniline	0.347	0.317	8.6	107	0.00
63	Azobenzene	1.366	1.267	7.2	109	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	0.117	0.119	-1.7	108	0.00
66 c	n-Nitrosodiphenylamine	0.675	0.625	7.4	109	0.00
67	4-Bromophenyl-phenylether	0.234	0.218	6.8	108	0.00
68	Hexachlorobenzene	0.254	0.236	7.1	108	0.00
69	Atrazine	0.207	0.190	8.2	109	0.00
70 C	Pentachlorophenol	0.065	0.059	9.2	96	0.00
71	Phenanthrene	1.117	1.020	8.7	108	0.00
72	Anthracene	1.111	1.017	8.5	108	0.00
73	Carbazole	1.008	0.906	10.1	107	0.00
74	Di-n-butylphthalate	1.298	1.198	7.7	110	0.00
75 C	Fluoranthene	1.143	1.016	11.1	107	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzidine	0.529	0.435	17.8	97	0.00
78	Pyrene	1.738	1.703	2.0	107	0.00
79 S	Terphenyl-d14	1.202	1.157	3.7	107	0.00
80	Butylbenzylphthalate	0.735	0.714	2.9	110	0.00
81	Benzo(a)anthracene	1.376	1.271	7.6	108	0.00
82	3,3'-Dichlorobenzidine	0.426	0.366	14.1	103	0.00
83	Chrysene	1.288	1.177	8.6	107	0.00
84	Bis(2-ethylhexyl)phthalate	0.873	0.833	4.6	112	0.00
85 c	Di-n-octyl phthalate	1.204	1.140	5.3	113	0.00
86 I	Perylene-d12	1.000	1.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	1.373	1.354	1.4	109	0.00
88	Benzo(b)fluoranthene	1.370	1.261	8.0	109	0.00
89	Benzo(k)fluoranthene	1.313	1.134	13.6	111	0.00
90 C	Benzo(a)pyrene	1.078	0.980	9.1	110	0.00
91	Dibenzo(a,h)anthracene	1.133	1.128	0.4	108	0.00
92	Benzo(g,h,i)perylene	1.130	1.134	-0.4	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
Data File : BF129768.D
Acq On : 14 Aug 2022 17:47
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF081422

Quant Time: Aug 15 02:36:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 15 02:28:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0