

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129768.D
 Acq On : 14 Aug 2022 17:47
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF081422

Quant Time: Aug 15 02:36:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	109	0.00
2	1,4-Dioxane	40.000	36.593	8.5	106	0.00
3	Pyridine	40.000	36.984	7.5	107	0.00
4	n-Nitrosodimethylamine	40.000	37.827	5.4	106	0.00
5 S	2-Fluorophenol	80.000	73.153	8.6	108	0.00
6	Aniline	40.000	36.691	8.3	108	0.00
7 S	Phenol-d6	80.000	72.341	9.6	107	0.00
8	2-Chlorophenol	40.000	36.666	8.3	107	0.00
9	Benzaldehyde	40.000	39.367	1.6	107	0.00
10 C	Phenol	40.000	36.512	8.7	108	0.00
11	bis(2-Chloroethyl)ether	40.000	36.797	8.0	107	0.00
12	1,3-Dichlorobenzene	40.000	36.897	7.8	108	0.00
13 C	1,4-Dichlorobenzene	40.000	36.764	8.1	109	0.00
14	1,2-Dichlorobenzene	40.000	36.435	8.9	107	0.00
15	Benzyl Alcohol	40.000	37.086	7.3	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.513	8.7	107	0.00
17	2-Methylphenol	40.000	36.717	8.2	109	0.00
18	Hexachloroethane	40.000	37.091	7.3	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.293	6.8	108	0.00
20	3+4-Methylphenols	40.000	36.766	8.1	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	36.731	8.2	108	0.00
23 S	Nitrobenzene-d5	80.000	73.862	7.7	107	0.00
24	Nitrobenzene	40.000	37.100	7.2	108	0.00
25	Isophorone	40.000	37.325	6.7	108	0.00
26 C	2-Nitrophenol	40.000	38.080	4.8	109	0.00
27	2,4-Dimethylphenol	40.000	37.120	7.2	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.455	6.4	107	0.00
29 C	2,4-Dichlorophenol	40.000	37.176	7.1	106	0.00
30	1,2,4-Trichlorobenzene	40.000	36.847	7.9	107	0.00
31	Naphthalene	40.000	36.724	8.2	108	0.00
32	Benzoic acid	40.000	41.589	-4.0	116	0.00
33	4-Chloroaniline	40.000	37.623	5.9	109	0.00
34 C	Hexachlorobutadiene	40.000	37.022	7.4	106	0.00
35	Caprolactam	40.000	37.146	7.1	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.124	7.2	108	0.00
37	2-Methylnaphthalene	40.000	37.322	6.7	108	0.00
38	1-Methylnaphthalene	40.000	37.226	6.9	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.593	8.5	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.176	-2.9	114	0.00
42 S	2,4,6-Tribromophenol	80.000	74.463	6.9	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.770	5.6	109	0.00
44	2,4,5-Trichlorophenol	40.000	37.047	7.4	105	0.00
45 S	2-Fluorobiphenyl	80.000	72.098	9.9	107	0.00
46	1,1'-Biphenyl	40.000	36.589	8.5	107	0.00
47	2-Chloronaphthalene	40.000	36.469	8.8	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129768.D
 Acq On : 14 Aug 2022 17:47
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF081422

Quant Time: Aug 15 02:36:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.589	6.0	108	0.00
49	Acenaphthylene	40.000	36.318	9.2	107	0.00
50	Dimethylphthalate	40.000	37.090	7.3	109	0.00
51	2,6-Dinitrotoluene	40.000	37.392	6.5	107	0.00
52 C	Acenaphthene	40.000	36.750	8.1	108	0.00
53	3-Nitroaniline	40.000	36.846	7.9	107	0.00
54 P	2,4-Dinitrophenol	40.000	39.336	1.7	113	0.00
55	Dibenzofuran	40.000	36.228	9.4	108	0.00
56 P	4-Nitrophenol	40.000	40.363	-0.9	110	0.00
57	2,4-Dinitrotoluene	40.000	36.716	8.2	110	0.00
58	Fluorene	40.000	36.214	9.5	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.836	5.4	106	0.00
60	Diethylphthalate	40.000	36.195	9.5	107	0.00
61	4-Chlorophenyl-phenylether	40.000	36.285	9.3	107	0.00
62	4-Nitroaniline	40.000	36.580	8.6	107	0.00
63	Azobenzene	40.000	37.094	7.3	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.451	-1.1	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.038	7.4	109	0.00
67	4-Bromophenyl-phenylether	40.000	37.152	7.1	108	0.00
68	Hexachlorobenzene	40.000	37.144	7.1	108	0.00
69	Atrazine	40.000	36.590	8.5	109	0.00
70 C	Pentachlorophenol	40.000	34.358	14.1	96	0.00
71	Phenanthrene	40.000	36.515	8.7	108	0.00
72	Anthracene	40.000	36.617	8.5	108	0.00
73	Carbazole	40.000	35.951	10.1	107	0.00
74	Di-n-butylphthalate	40.000	36.935	7.7	110	0.00
75 C	Fluoranthene	40.000	35.542	11.1	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzidine	40.000	35.078	12.3	97	0.00
78	Pyrene	40.000	39.193	2.0	107	0.00
79 S	Terphenyl-d14	80.000	77.040	3.7	107	0.00
80	Butylbenzylphthalate	40.000	38.825	2.9	110	0.00
81	Benzo(a)anthracene	40.000	36.965	7.6	108	0.00
82	3,3'-Dichlorobenzidine	40.000	34.384	14.0	103	0.00
83	Chrysene	40.000	36.564	8.6	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.174	4.6	112	0.00
85 c	Di-n-octyl phthalate	40.000	37.863	5.3	113	0.00
86 I	Perylene-d12	20.000	20.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.460	1.3	109	0.00
88	Benzo(b)fluoranthene	40.000	36.826	7.9	109	0.00
89	Benzo(k)fluoranthene	40.000	34.542	13.6	111	0.00
90 C	Benzo(a)pyrene	40.000	36.333	9.2	110	0.00
91	Dibenzo(a,h)anthracene	40.000	39.831	0.4	108	0.00
92	Benzo(g,h,i)perylene	40.000	40.124	-0.3	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
Data File : BF129768.D
Acq On : 14 Aug 2022 17:47
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF081422

Quant Time: Aug 15 02:36:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 15 02:28:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0