

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129780.D
 Acq On : 15 Aug 2022 00:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 15 03:04:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	37.550	6.1	108	-0.01
3	Pyridine	40.000	38.563	3.6	111	-0.02
4	n-Nitrosodimethylamine	40.000	38.433	3.9	107	-0.01
5 S	2-Fluorophenol	80.000	74.106	7.4	109	0.00
6	Aniline	40.000	36.247	9.4	107	0.00
7 S	Phenol-d6	80.000	73.211	8.5	108	0.00
8	2-Chlorophenol	40.000	36.860	7.9	107	0.00
9	Benzaldehyde	40.000	40.462	-1.2	109	0.00
10 C	Phenol	40.000	36.771	8.1	108	0.00
11	bis(2-Chloroethyl)ether	40.000	37.490	6.3	109	0.00
12	1,3-Dichlorobenzene	40.000	37.467	6.3	109	0.00
13 C	1,4-Dichlorobenzene	40.000	36.902	7.7	109	0.00
14	1,2-Dichlorobenzene	40.000	36.773	8.1	108	0.00
15	Benzyl Alcohol	40.000	35.115	12.2	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.031	7.4	108	0.00
17	2-Methylphenol	40.000	37.352	6.6	110	0.00
18	Hexachloroethane	40.000	37.652	5.9	110	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.319	6.7	108	0.00
20	3+4-Methylphenols	40.000	36.857	7.9	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	36.666	8.3	108	0.00
23 S	Nitrobenzene-d5	80.000	73.512	8.1	107	0.00
24	Nitrobenzene	40.000	36.200	9.5	105	0.00
25	Isophorone	40.000	36.903	7.7	107	0.00
26 C	2-Nitrophenol	40.000	38.082	4.8	108	0.00
27	2,4-Dimethylphenol	40.000	36.729	8.2	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.872	7.8	105	0.00
29 C	2,4-Dichlorophenol	40.000	37.283	6.8	106	0.00
30	1,2,4-Trichlorobenzene	40.000	36.811	8.0	107	0.00
31	Naphthalene	40.000	36.642	8.4	107	0.00
32	Benzoic acid	40.000	40.388	-1.0	112	0.00
33	4-Chloroaniline	40.000	36.267	9.3	105	0.00
34 C	Hexachlorobutadiene	40.000	37.099	7.3	107	0.00
35	Caprolactam	40.000	37.428	6.4	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.508	8.7	106	0.00
37	2-Methylnaphthalene	40.000	36.556	8.6	106	0.00
38	1-Methylnaphthalene	40.000	36.434	8.9	107	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.383	9.0	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.905	-2.3	111	0.00
42 S	2,4,6-Tribromophenol	80.000	73.494	8.1	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.517	8.7	104	0.00
44	2,4,5-Trichlorophenol	40.000	37.460	6.3	105	0.00
45 S	2-Fluorobiphenyl	80.000	72.175	9.8	106	0.00
46	1,1'-Biphenyl	40.000	36.369	9.1	106	0.00
47	2-Chloronaphthalene	40.000	36.441	8.9	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129780.D
 Acq On : 15 Aug 2022 00:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 15 03:04:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.730	5.7	107	0.00
49	Acenaphthylene	40.000	36.254	9.4	105	0.00
50	Dimethylphthalate	40.000	36.497	8.8	106	0.00
51	2,6-Dinitrotoluene	40.000	37.336	6.7	106	0.00
52 C	Acenaphthene	40.000	36.675	8.3	106	0.00
53	3-Nitroaniline	40.000	35.561	11.1	102	0.00
54 P	2,4-Dinitrophenol	40.000	37.975	5.1	108	0.00
55	Dibenzofuran	40.000	36.073	9.8	107	0.00
56 P	4-Nitrophenol	40.000	34.382	14.0	93	0.00
57	2,4-Dinitrotoluene	40.000	36.555	8.6	108	0.00
58	Fluorene	40.000	35.982	10.0	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.773	0.6	110	0.00
60	Diethylphthalate	40.000	36.342	9.1	107	0.00
61	4-Chlorophenyl-phenylether	40.000	36.867	7.8	108	0.00
62	4-Nitroaniline	40.000	35.978	10.1	105	0.00
63	Azobenzene	40.000	36.796	8.0	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.038	-2.6	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.871	7.8	108	0.00
67	4-Bromophenyl-phenylether	40.000	37.195	7.0	108	0.00
68	Hexachlorobenzene	40.000	37.838	5.4	110	0.00
69	Atrazine	40.000	37.317	6.7	110	0.00
70 C	Pentachlorophenol	40.000	32.939	17.7	91	0.00
71	Phenanthrene	40.000	36.520	8.7	107	0.00
72	Anthracene	40.000	36.244	9.4	106	0.00
73	Carbazole	40.000	36.035	9.9	107	0.00
74	Di-n-butylphthalate	40.000	37.095	7.3	109	0.00
75 C	Fluoranthene	40.000	36.284	9.3	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	112	0.00
77	Benzydine	40.000	39.755	0.6	112	0.00
78	Pyrene	40.000	38.182	4.5	109	0.00
79 S	Terphenyl-d14	80.000	76.095	4.9	110	0.00
80	Butylbenzylphthalate	40.000	38.736	3.2	114	0.00
81	Benzo(a)anthracene	40.000	36.483	8.8	111	0.00
82	3,3'-Dichlorobenzidine	40.000	35.701	10.7	111	0.00
83	Chrysene	40.000	36.373	9.1	111	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.284	4.3	117	0.00
85 c	Di-n-octyl phthalate	40.000	37.083	7.3	115	0.00
86 I	Perylene-d12	20.000	20.000	0.0	113	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.813	0.5	112	0.00
88	Benzo(b)fluoranthene	40.000	36.636	8.4	111	0.00
89	Benzo(k)fluoranthene	40.000	34.912	12.7	114	0.00
90 C	Benzo(a)pyrene	40.000	36.360	9.1	112	0.00
91	Dibenzo(a,h)anthracene	40.000	40.077	-0.2	111	0.00
92	Benzo(g,h,i)perylene	40.000	39.725	0.7	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
Data File : BF129780.D
Acq On : 15 Aug 2022 00:34
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 15 03:04:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 15 02:28:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0