

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129780.D
 Acq On : 15 Aug 2022 00:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022
 Supervised By :Sohil Jodhani 08/18/2022

Quant Time: Aug 15 03:04:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.822	152	177428	20.000	ng	0.00
21) Naphthalene-d8	8.110	136	687224	20.000	ng	0.00
39) Acenaphthene-d10	9.863	164	368701	20.000	ng	0.00
64) Phenanthrene-d10	11.357	188	626819	20.000	ng	0.00
76) Chrysene-d12	14.004	240	394151	20.000	ng	0.00
86) Perylene-d12	15.468	264	307519	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.457	112	794131	74.106	ng	0.00
7) Phenol-d6	6.481	99	982463	73.211	ng	0.00
23) Nitrobenzene-d5	7.398	82	943721	73.512	ng	0.00
42) 2,4,6-Tribromophenol	10.663	330	271963	73.494	ng	0.00
45) 2-Fluorobiphenyl	9.186	172	1752098	72.175	ng	0.00
79) Terphenyl-d14	12.939	244	1802525	76.095	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.551	88	166014	37.550	ng	99
3) Pyridine	3.322	79	446169	38.563	ng	99
4) n-Nitrosodimethylamine	3.287	42	204927	38.433	ng	98
6) Aniline	6.492	93	554399	36.247	ng	98
8) 2-Chlorophenol	6.622	128	439733	36.860	ng	99
9) Benzaldehyde	6.381	77	290175	40.462	ng	99
10) Phenol	6.492	94	541273	36.771	ng	100
11) bis(2-Chloroethyl)ether	6.563	93	419072	37.490	ng	99
12) 1,3-Dichlorobenzene	6.763	146	482537	37.467	ng	98
13) 1,4-Dichlorobenzene	6.839	146	485605	36.902	ng	99
14) 1,2-Dichlorobenzene	6.992	146	449440	36.773	ng	99
15) Benzyl Alcohol	6.981	79	355747	35.115	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.098	45	554661	37.031	ng	99
17) 2-Methylphenol	7.092	107	357787	37.352	ng	99
18) Hexachloroethane	7.328	117	186064	37.652	ng	99
19) n-Nitroso-di-n-propyla...	7.245	70	316557	37.319	ng	95
20) 3+4-Methylphenols	7.251	107	473908	36.857	ng	97
22) Acetophenone	7.239	105	614053	36.666	ng	98
24) Nitrobenzene	7.416	77	468613	36.200	ng	99
25) Isophorone	7.651	82	815397	36.903	ng	99
26) 2-Nitrophenol	7.728	139	242483	38.082	ng	99
27) 2,4-Dimethylphenol	7.769	122	335394	36.729	ng	97
28) bis(2-Chloroethoxy)met...	7.857	93	475696	36.872	ng	98
29) 2,4-Dichlorophenol	7.981	162	372269	37.283	ng	99
30) 1,2,4-Trichlorobenzene	8.051	180	387757	36.811	ng	99
31) Naphthalene	8.128	128	1316634	36.642	ng	99
32) Benzoic acid	7.922	122	156524	40.388	ng	98
33) 4-Chloroaniline	8.186	127	512466	36.267	ng	99
34) Hexachlorobutadiene	8.239	225	237958	37.099	ng	98
35) Caprolactam	8.575	113	116217m	37.428	ng	
36) 4-Chloro-3-methylphenol	8.681	107	393269	36.508	ng	99
37) 2-Methylnaphthalene	8.822	142	862111	36.556	ng	100
38) 1-Methylnaphthalene	8.922	142	835065	36.434	ng	99
40) 1,2,4,5-Tetrachloroben...	8.986	216	373103	36.383	ng	99
41) Hexachlorocyclopentadiene	8.963	237	91541	40.905	ng	97
43) 2,4,6-Trichlorophenol	9.110	196	243303	36.517	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
 Data File : BF129780.D
 Acq On : 15 Aug 2022 00:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022
 Supervised By :Sohil Jodhani 08/18/2022

Quant Time: Aug 15 03:04:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.157	196	268518	37.460	ng	100
46) 1,1'-Biphenyl	9.286	154	1018698	36.369	ng	99
47) 2-Chloronaphthalene	9.310	162	807624	36.441	ng	98
48) 2-Nitroaniline	9.416	65	264252	37.730	ng	99
49) Acenaphthylene	9.727	152	1218381	36.254	ng	100
50) Dimethylphthalate	9.586	163	964644	36.497	ng	99
51) 2,6-Dinitrotoluene	9.657	165	214065	37.336	ng	99
52) Acenaphthene	9.898	154	747780	36.675	ng	99
53) 3-Nitroaniline	9.833	138	225480	35.561	ng	95
54) 2,4-Dinitrophenol	9.945	184	90441	37.975	ng	98
55) Dibenzofuran	10.069	168	1105001	36.073	ng	95
56) 4-Nitrophenol	10.022	139	96986	34.382	ng	97
57) 2,4-Dinitrotoluene	10.069	165	272075	36.555	ng	98
58) Fluorene	10.416	166	920305	35.982	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.198	232	209229	39.773	ng	# 82
60) Diethylphthalate	10.286	149	957771	36.342	ng	100
61) 4-Chlorophenyl-phenyle...	10.404	204	423157	36.867	ng	99
62) 4-Nitroaniline	10.451	138	230139m	35.978	ng	
63) Azobenzene	10.563	77	926923	36.796	ng	99
65) 4,6-Dinitro-2-methylph...	10.480	198	151011	41.038	ng	98
66) n-Nitrosodiphenylamine	10.527	169	780055	36.871	ng	99
67) 4-Bromophenyl-phenylether	10.892	248	273076	37.195	ng	99
68) Hexachlorobenzene	10.963	284	300913	37.838	ng	97
69) Atrazine	11.057	200	242314	37.317	ng	98
70) Pentachlorophenol	11.169	266	70751m	32.939	ng	
71) Phenanthrene	11.380	178	1278465	36.520	ng	100
72) Anthracene	11.433	178	1261787	36.244	ng	100
73) Carbazole	11.592	167	1138792	36.035	ng	99
74) Di-n-butylphthalate	11.904	149	1508730	37.095	ng	100
75) Fluoranthene	12.569	202	1299790	36.284	ng	97
77) Benzidine	12.692	184	381999	39.755	ng	99
78) Pyrene	12.804	202	1307803	38.182	ng	99
80) Butylbenzylphthalate	13.410	149	561408	38.736	ng	100
81) Benzo(a)anthracene	13.992	228	989193	36.483	ng	99
82) 3,3'-Dichlorobenzidine	13.951	252	299796	35.701	ng	98
83) Chrysene	14.033	228	923167	36.373	ng	99
84) Bis(2-ethylhexyl)phtha...	13.963	149	658284	38.284	ng	99
85) Di-n-octyl phthalate	14.574	149	880220	37.083	ng	100
87) Indeno(1,2,3-cd)pyrene	16.956	276	840515	39.813	ng	99
88) Benzo(b)fluoranthene	15.039	252	771455	36.636	ng	99
89) Benzo(k)fluoranthene	15.068	252	705028	34.912	ng	98
90) Benzo(a)pyrene	15.409	252	602946	36.360	ng	100
91) Dibenzo(a,h)anthracene	16.962	278	698116	40.077	ng	99
92) Benzo(g,h,i)perylene	17.403	276	690367	39.725	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081422\
Data File : BF129780.D
Acq On : 15 Aug 2022 00:34
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022


Quant Time: Aug 15 03:04:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 15 02:28:31 2022
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.822 min Scan# 71
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC



Tgt Ion:152 Resp: 177428
Ion Ratio Lower Upper
152 100
150 156.4 126.9 190.3
115 66.1 53.4 80.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#2
1,4-Dioxane
Concen: 37.550 ng
RT: 2.551 min Scan# 45
Delta R.T. -0.012 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion: 88 Resp: 166014
Ion Ratio Lower Upper
88 100
58 76.8 60.9 91.3
43 29.8 23.6 35.4





#3
Pyridine
Concen: 38.563 ng
RT: 3.322 min Scan# 170
Delta R.T. -0.018 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion: 79 Resp: 446169
Ion Ratio Lower Upper
79 100
52 73.5 59.4 89.2
51 36.2 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#4
n-Nitrosodimethylamine
Concen: 38.433 ng
RT: 3.287 min Scan# 170
Delta R.T. -0.012 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion: 42 Resp: 204927
Ion Ratio Lower Upper
42 100
74 126.9 103.1 154.7
44 6.2 5.4 8.2






#5
2-Fluorophenol
Concen: 74.106 ng
RT: 5.457 min Scan# 511
Delta R.T. -0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion:112 Resp: 79413
Ion Ratio Lower Upper
112 100
64 62.6 47.6 71.4
63 32.4 24.6 37.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#6
Aniline
Concen: 36.247 ng
RT: 6.492 min Scan# 715
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion: 93 Resp: 554399
Ion Ratio Lower Upper
93 100
66 87.5 68.7 103.1
65 52.8 41.6 62.4






#7
Phenol-d6
Concen: 73.211 ng
RT: 6.481 min Scan# 713
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion: 99 Resp: 98246
Ion Ratio Lower Upper
99 100
42 18.8 15.0 22.4
71 35.2 28.8 43.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#8
2-Chlorophenol
Concen: 36.860 ng
RT: 6.622 min Scan# 737
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion:128 Resp: 439733
Ion Ratio Lower Upper
128 100
130 33.0 12.5 52.5
64 45.2 24.9 64.9





#9
Benzaldehyde
Concen: 40.462 ng
RT: 6.381 min Scan# 696
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC



Tgt Ion: 77 Resp: 290175
Ion Ratio Lower Upper
77 100
105 100.4 80.3 120.3
106 96.2 74.3 114.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#10
Phenol
Concen: 36.771 ng
RT: 6.492 min Scan# 715
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion: 94 Resp: 541273
Ion Ratio Lower Upper
94 100
65 44.6 24.9 64.9
66 73.9 54.2 94.2









#11
bis(2-Chloroethyl)ether
Concen: 37.490 ng
RT: 6.563 min Scan# 727
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 93 Resp: 419072
Ion Ratio Lower Upper
93 100
63 77.3 55.9 95.9
95 31.7 12.3 52.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#12
1,3-Dichlorobenzene
Concen: 37.467 ng
RT: 6.763 min Scan# 761
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion:146 Resp: 482537
Ion Ratio Lower Upper
146 100
148 66.0 51.2 76.8
75 32.7 26.5 39.7









#13
1,4-Dichlorobenzene
Concen: 36.902 ng
RT: 6.839 min Scan# 774
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:146 Resp: 485609
Ion Ratio Lower Upper
146 100
148 64.4 50.6 75.8
111 44.4 36.2 54.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#14
1,2-Dichlorobenzene
Concen: 36.773 ng
RT: 6.992 min Scan# 800
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion:146 Resp: 449440
Ion Ratio Lower Upper
146 100
148 63.9 51.8 77.6
111 47.2 37.2 55.8









#15
Benzyl Alcohol
Concen: 35.115 ng
RT: 6.981 min Scan# 79
Delta R.T. 0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 79 Resp: 35574
Ion Ratio Lower Upper
79 100
108 74.1 59.2 88.8
77 63.6 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.031 ng
RT: 7.098 min Scan# 818
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion: 45 Resp: 554661
Ion Ratio Lower Upper
45 100
77 46.5 27.4 67.4
79 43.4 24.4 64.4





#17
2-Methylphenol
Concen: 37.352 ng
RT: 7.092 min Scan# 817
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC



Tgt Ion:107 Resp: 35778
Ion Ratio Lower Upper
107 100
108 111.1 90.2 135.2
77 64.9 51.8 77.8
79 60.7 47.4 71.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#18
Hexachloroethane
Concen: 37.652 ng
RT: 7.328 min Scan# 857
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion:117 Resp: 186064
Ion Ratio Lower Upper
117 100
119 98.7 79.1 118.7
201 84.6 66.6 100.0






#19
n-Nitroso-di-n-propylamine
Concen: 37.319 ng
RT: 7.245 min Scan# 842
Delta R.T. 0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion: 70 Resp: 31655
Ion Ratio Lower Upper
70 100
42 55.2 47.6 71.4
101 10.3 8.1 12.1
130 23.5 17.2 25.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#20
3+4-Methylphenols
Concen: 36.857 ng
RT: 7.251 min Scan# 844
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion:107 Resp: 473908
Ion Ratio Lower Upper
107 100
108 91.6 71.3 111.3
77 46.2 20.4 60.4
79 30.9 11.1 51.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.110 min Scan# 990
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion:136 Resp: 687224
Ion Ratio Lower Upper
136 100
137 10.9 8.4 12.6
54 7.2 6.2 9.2
68 5.1 4.2 6.4


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#22
Acetophenone
Concen: 36.666 ng
RT: 7.239 min Scan# 842
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion:105 Resp: 614053
Ion Ratio Lower Upper
105 100
71 7.1 6.6 9.8
51 29.1 23.8 35.8
120 21.6 17.9 26.9






#23
Nitrobenzene-d5
Concen: 73.512 ng
RT: 7.398 min Scan# 869
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion: 82 Resp: 94372
Ion Ratio Lower Upper
82 100
128 44.9 36.1 54.1
54 50.1 40.0 60.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#24
Nitrobenzene
Concen: 36.200 ng
RT: 7.416 min Scan# 872
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion: 77 Resp: 468613
Ion Ratio Lower Upper
77 100
123 48.0 39.2 58.8
65 13.7 11.1 16.7






#25
Isophorone
Concen: 36.903 ng
RT: 7.651 min Scan# 912
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :




SSTDCCC040EC




Tgt Ion: 82 Resp: 81539
Ion Ratio Lower Upper
82 100
95 7.5 5.9 8.9
138 17.7 14.6 21.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#26
2-Nitrophenol
Concen: 38.082 ng
RT: 7.728 min Scan# 925
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion:139 Resp: 242483
Ion Ratio Lower Upper
139 100
109 30.7 24.4 36.6
65 52.1 40.9 61.3





#27

2,4-Dimethylphenol

Concen: 36.729 ng

RT: 7.769 min Scan# 91

Delta R.T. -0.000 min

Lab File: BF129780.D


Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion	Ratio	Lower	Upper
122	100		
107	119.2	92.2	138.4
121	59.9	46.1	69.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022

Supervised By :Sohil Jodhani 08/18/2022



#28

bis(2-Chloroethoxy)methane

Concen: 36.872 ng




RT: 7.857 min Scan# 947


Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34

Tgt Ion	Ratio	Lower	Upper
93	100		
95	33.0	25.6	38.4
123	11.7	9.0	13.4





#29

2,4-Dichlorophenol

Concen: 37.283 ng

RT: 7.981 min Scan# 90

Delta R.T. -0.000 min

Lab File: BF129780.D


Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC



Tgt Ion:162 Resp: 372269

Ion Ratio Lower Upper

162 100

164 64.7 43.7 83.7




98 36.3 16.5 56.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022

Supervised By :Sohil Jodhani 08/18/2022



#30

1,2,4-Trichlorobenzene

Concen: 36.811 ng

RT: 8.051 min Scan# 980

Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34




Tgt Ion:180 Resp: 387757


Ion Ratio Lower Upper

180 100

182 95.8 77.6 116.4


145 33.0 25.6 38.4





#31
Naphthalene
Concen: 36.642 ng
RT: 8.128 min Scan# 993
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC



Tgt Ion:128 Resp: 1316634
Ion Ratio Lower Upper
128 100
129 11.7 8.9 13.3
127 13.3 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#32
Benzoic acid
Concen: 40.388 ng
RT: 7.922 min Scan# 958
Delta R.T. 0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion:122 Resp: 156524
Ion Ratio Lower Upper
122 100
105 122.7 99.9 139.9
77 85.3 65.9 105.9





#33
4-Chloroaniline
Concen: 36.267 ng
RT: 8.186 min Scan# 1003
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040EC

Tgt Ion:127 Resp: 512460
Ion Ratio Lower Upper
127 100
129 32.1 26.7 40.1
65 30.0 24.2 36.2
92 19.6 15.2 22.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#34
Hexachlorobutadiene
Concen: 37.099 ng
RT: 8.239 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion:225 Resp: 237958
Ion Ratio Lower Upper
225 100
223 61.6 50.2 75.2
227 65.1 50.6 75.8





#35
 Caprolactam
 Concen: 37.428 ng m
 RT: 8.575 min Scan# 1069
 Delta R.T. 0.006 min
 Lab File: BF129780.D
 Acq: 15 Aug 2022 00:34




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC



Tgt Ion:113 Resp: 11621
 Ion Ratio Lower Upper
 113 100
 55 156.9 148.8 188.8
 56 119.0 109.1 149.1




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
 Supervised By :Sohil Jodhani 08/18/2022



#36
 4-Chloro-3-methylphenol
 Concen: 36.508 ng
 RT: 8.681 min Scan# 1087
 Delta R.T. -0.000 min
 Lab File: BF129780.D
 Acq: 15 Aug 2022 00:34

Tgt Ion:107 Resp: 393269
 Ion Ratio Lower Upper
 107 100
 144 23.6 19.9 29.9
 142 74.4 60.2 90.4





#37
2-Methylnaphthalene
Concen: 36.556 ng
RT: 8.822 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040EC

Tgt Ion:142 Resp: 86211
Ion Ratio Lower Upper
142 100
141 88.9 70.9 106.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#38
1-Methylnaphthalene
Concen: 36.434 ng
RT: 8.922 min Scan# 1128
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion:142 Resp: 835065
Ion Ratio Lower Upper
142 100
141 92.6 73.5 110.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.863 min Scan# 1127
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:164 Resp: 36870
Ion Ratio Lower Upper
164 100
162 102.2 82.0 123.0
160 45.2 35.4 53.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#40
1,2,4,5-Tetrachlorobenzene
Concen: 36.383 ng
RT: 8.986 min Scan# 1139
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:216 Resp: 373103
Ion Ratio Lower Upper
216 100
214 79.6 63.1 94.7
179 22.8 18.1 27.1
108 20.0 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 40.905 ng
RT: 8.963 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion: 237 Resp: 9154
Ion Ratio Lower Upper
237 100
235 63.8 41.5 81.5
272 12.2 0.0 33.3

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 08/15/2022
Supervised By : Sohil Jodhani 08/18/2022




#42
2,4,6-Tribromophenol
Concen: 73.494 ng
RT: 10.663 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion: 330 Resp: 271963
Ion Ratio Lower Upper
330 100
332 102.1 82.6 124.0
141 35.0 27.6 41.4





#43
2,4,6-Trichlorophenol
Concen: 36.517 ng
RT: 9.110 min Scan# 1160
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC



Tgt Ion:196 Resp: 24330
Ion Ratio Lower Upper
196 100
198 96.5 79.0 118.4
200 30.6 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#44
2,4,5-Trichlorophenol
Concen: 37.460 ng
RT: 9.157 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion:196 Resp: 268518
Ion Ratio Lower Upper
196 100
198 95.9 76.4 114.6
97 57.9 46.5 69.7
132 31.0 25.2 37.8





#45

2-Fluorobiphenyl

Concen: 72.175 ng

RT: 9.186 min Scan# 1173

Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:172 Resp: 1752098

Ion Ratio Lower Upper

172 100

171 35.4 28.4 42.6





170 24.0 19.1 28.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022

Supervised By :Sohil Jodhani 08/18/2022



#46

1,1'-Biphenyl

Concen: 36.369 ng

RT: 9.286 min Scan# 1190

Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34




Tgt Ion:154 Resp: 1018698


Ion Ratio Lower Upper

154 100

153 40.5 20.2 60.2

76 11.9 0.0 32.1






#47
2-Chloronaphthalene
Concen: 36.441 ng
RT: 9.310 min Scan# 1195
Delta R.T. -0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion:162 Resp: 807624
Ion Ratio Lower Upper
162 100
127 43.6 33.1 49.7
164 33.0 26.7 40.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#48
2-Nitroaniline
Concen: 37.730 ng
RT: 9.416 min Scan# 1212
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion: 65 Resp: 264252
Ion Ratio Lower Upper
65 100
92 69.1 55.0 82.4
138 99.7 81.3 121.9





#49

Acenaphthylene

Concen: 36.254 ng

RT: 9.727 min Scan# 1265

Delta R.T. -0.000 min

Lab File: BF129780.D


Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC



Tgt Ion:152 Resp: 121838

Ion Ratio Lower Upper

152 100

151 20.2 16.1 24.1




153 13.4 10.6 16.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022

Supervised By :Sohil Jodhani 08/18/2022



#50

Dimethylphthalate

Concen: 36.497 ng

RT: 9.586 min Scan# 1241

Delta R.T. 0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34




Tgt Ion:163 Resp: 964644


Ion Ratio Lower Upper

163 100

194 3.8 3.3 4.9


164 10.3 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 37.336 ng
RT: 9.657 min Scan# 1253
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:165 Resp: 214065
Ion Ratio Lower Upper
165 100
63 61.6 47.8 71.8
89 59.1 47.2 70.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#52
Acenaphthene
Concen: 36.675 ng
RT: 9.898 min Scan# 1294
Delta R.T. 0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion:154 Resp: 747780
Ion Ratio Lower Upper
154 100
153 111.1 89.8 134.6
152 52.9 42.1 63.1






#53
3-Nitroaniline
Concen: 35.561 ng
RT: 9.833 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion:138 Resp: 225480
Ion Ratio Lower Upper
138 100
108 10.9 8.4 12.6
92 117.6 89.8 134.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#54
2,4-Dinitrophenol
Concen: 37.975 ng
RT: 9.945 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:184 Resp: 90441
Ion Ratio Lower Upper
184 100
63 75.4 59.6 89.4
154 70.8 55.3 82.9





#55
Dibenzenofuran
Concen: 36.073 ng
RT: 10.069 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC



Tgt Ion:168 Resp: 110500
Ion Ratio Lower Upper
168 100
139 42.5 31.3 46.9
169 14.1 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#56
4-Nitrophenol
Concen: 34.382 ng
RT: 10.022 min Scan# 1315
Delta R.T. 0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion:139 Resp: 96986
Ion Ratio Lower Upper
139 100
109 74.1 52.9 92.9
65 95.5 79.5 119.5





#57
2,4-Dinitrotoluene
Concen: 36.555 ng
RT: 10.069 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:165 Resp: 272075
Ion Ratio Lower Upper
165 100
63 58.6 48.4 72.6
89 77.1 62.3 93.5
182 2.4 2.1 3.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#58
Fluorene
Concen: 35.982 ng
RT: 10.416 min Scan# 1382
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:166 Resp: 920305
Ion Ratio Lower Upper
166 100
165 97.7 77.9 116.9
167 13.6 11.0 16.4





#59
2,3,4,6-Tetrachlorophenol
Concen: 39.773 ng
RT: 10.198 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:232 Resp: 209229
Ion Ratio Lower Upper
232 100
131 48.4 29.9 44.9
130 2.7 1.5 2.3
166 31.7 17.8 26.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#60
Diethylphthalate
Concen: 36.342 ng
RT: 10.286 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:149 Resp: 957771
Ion Ratio Lower Upper
149 100
177 22.3 17.7 26.5
150 12.5 10.1 15.1





#61
4-Chlorophenyl-phenylether
Concen: 36.867 ng
RT: 10.404 min Scan# 1380
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:204 Resp: 42315
Ion Ratio Lower Upper
204 100
206 32.6 26.5 39.7
141 65.0 52.6 79.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#62
4-Nitroaniline
Concen: 35.978 ng m
RT: 10.451 min Scan# 1388
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:138 Resp: 230139
Ion Ratio Lower Upper
138 100
92 52.6 30.3 70.3
108 78.9 55.7 95.7





#63
Azobenzene
Concen: 36.796 ng
RT: 10.563 min Scan# 1407
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion: 77 Resp: 926923
Ion Ratio Lower Upper
77 100
182 23.0 2.2 42.2
105 21.2 0.8 40.8
51 29.8 10.3 50.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.357 min Scan# 1542
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion:188 Resp: 626819
Ion Ratio Lower Upper
188 100
94 8.6 7.4 11.2
80 9.4 7.9 11.9









#65
4,6-Dinitro-2-methylphenol
Concen: 41.038 ng
RT: 10.480 min Scan# 1393
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:198 Resp: 15101
Ion Ratio Lower Upper
198 100
51 51.4 29.4 69.4
105 49.5 29.2 69.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#66
n-Nitrosodiphenylamine
Concen: 36.871 ng
RT: 10.527 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion:169 Resp: 780055
Ion Ratio Lower Upper
169 100
168 65.9 53.4 80.2
167 35.9 29.1 43.7





#67
4-Bromophenyl-phenylether
Concen: 37.195 ng
RT: 10.892 min Scan# 1463
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC



Tgt Ion:248 Resp: 273070
Ion Ratio Lower Upper
248 100
250 97.8 77.8 116.6
141 80.8 66.2 99.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#68
Hexachlorobenzene
Concen: 37.838 ng
RT: 10.963 min Scan# 1475
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34


Tgt Ion:284 Resp: 300913
Ion Ratio Lower Upper
284 100
142 37.0 31.4 47.0
249 31.2 25.7 38.5





#69
 Atrazine
 Concen: 37.317 ng
 RT: 11.057 min Scan# 1491
 Delta R.T. -0.000 min
 Lab File: BF129780.D
 Acq: 15 Aug 2022 00:34




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC




Tgt Ion:200 Resp: 242314
 Ion Ratio Lower Upper
 200 100
 173 26.8 6.4 46.4
 215 45.5 27.1 67.1

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
 Supervised By :Sohil Jodhani 08/18/2022




#70
 Pentachlorophenol
 Concen: 32.939 ng m
 RT: 11.169 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: BF129780.D
 Acq: 15 Aug 2022 00:34



Tgt Ion:266 Resp: 70751
 Ion Ratio Lower Upper
 266 100
 268 60.2 49.0 73.4
 264 61.0 47.9 71.9






#71
Phenanthrene
Concen: 36.520 ng
RT: 11.380 min Scan# 1546
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion:178 Resp: 127846
Ion Ratio Lower Upper
178 100
176 19.5 15.5 23.3
179 15.5 12.5 18.7


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#72
Anthracene
Concen: 36.244 ng
RT: 11.433 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:178 Resp: 1261787
Ion Ratio Lower Upper
178 100
176 18.8 15.1 22.7
179 15.6 12.6 18.8





#73
 Carbazole
 Concen: 36.035 ng
 RT: 11.592 min Scan# 11582
 Delta R.T. -0.000 min
 Lab File: BF129780.D
 Acq: 15 Aug 2022 00:34




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC




Tgt Ion:167 Resp: 1138792
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 14.0 11.0 16.6

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
 Supervised By :Sohil Jodhani 08/18/2022




#74
 Di-n-butylphthalate
 Concen: 37.095 ng
 RT: 11.904 min Scan# 1635
 Delta R.T. -0.000 min
 Lab File: BF129780.D
 Acq: 15 Aug 2022 00:34




Tgt Ion:149 Resp: 1508730
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.8 11.6
 104 5.3 4.2 6.4





#75
Fluoranthene
Concen: 36.284 ng
RT: 12.569 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:202 Resp: 1299790
Ion Ratio Lower Upper
202 100
101 12.7 0.0 30.7
203 17.9 0.0 37.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.004 min Scan# 1992
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:240 Resp: 394151
Ion Ratio Lower Upper
240 100
120 11.6 9.2 13.8
236 25.5 20.7 31.1





#77
Benzidine
Concen: 39.755 ng
RT: 12.692 min Scan# 1769
Delta R.T. 0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:184 Resp: 381999
Ion Ratio Lower Upper
184 100
185 14.4 11.4 17.2
183 11.5 9.4 14.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#78
Pyrene
Concen: 38.182 ng
RT: 12.804 min Scan# 1788
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion:202 Resp: 1307803
Ion Ratio Lower Upper
202 100
200 21.8 17.0 25.6
203 17.5 14.1 21.1





#79

Terphenyl-d14

Concen: 76.095 ng

RT: 12.939 min Scan# 1811

Delta R.T. -0.000 min

Lab File: BF129780.D


Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC



Tgt Ion:244 Resp: 180252

Ion Ratio Lower Upper

244 100

212 7.1 5.9 8.9




122 12.0 9.9 14.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022

Supervised By :Sohil Jodhani 08/18/2022



#80

Butylbenzylphthalate

Concen: 38.736 ng

RT: 13.410 min Scan# 1891

Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34




Tgt Ion:149 Resp: 561408


Ion Ratio Lower Upper

149 100

91 69.6 55.4 83.0

206 20.0 15.8 23.6






#81
Benzo(a)anthracene
Concen: 36.483 ng
RT: 13.992 min Scan# 1990
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion:228 Resp: 98919
Ion Ratio Lower Upper
228 100
226 27.4 21.4 32.0
229 20.0 16.1 24.1


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#82
3,3'-Dichlorobenzidine
Concen: 35.701 ng
RT: 13.951 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Tgt Ion:252 Resp: 299796
Ion Ratio Lower Upper
252 100
254 65.6 51.4 77.2
126 12.6 10.2 15.4





#83
Chrysene
Concen: 36.373 ng
RT: 14.033 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:228 Resp: 92316
Ion Ratio Lower Upper
228 100
226 29.6 24.4 36.6
229 19.9 15.9 23.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022




#84
Bis(2-ethylhexyl)phthalate
Concen: 38.284 ng
RT: 13.963 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion:149 Resp: 658284
Ion Ratio Lower Upper
149 100
167 26.8 22.0 33.0
279 4.5 3.6 5.4





#85

Di-n-octyl phthalate

Concen: 37.083 ng

RT: 14.574 min Scan# 2089

Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040EC

Tgt Ion:149 Resp: 880220
 Ion Ratio Lower Upper
 149 100
 167 1.7 1.3 1.9
 43 9.6 7.8 11.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022
 Supervised By :Sohil Jodhani 08/18/2022



#86

Perylene-d12

Concen: 20.000 ng




RT: 15.468 min Scan# 2241


Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34

Tgt Ion:264 Resp: 307519
 Ion Ratio Lower Upper
 264 100
 260 24.0 19.1 28.7
 265 22.1 16.7 25.1






#87
Indeno(1,2,3-cd)pyrene
Concen: 39.813 ng
RT: 16.956 min Scan# 2494
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion: 276 Resp: 840515
Ion Ratio Lower Upper
276 100
138 24.5 19.8 29.8
277 25.3 20.4 30.6

Manual Integrations




APPROVED


Reviewed By : Jagrut Upadhyay 08/15/2022
Supervised By : Sohil Jodhani 08/18/2022



#88
Benzo(b)fluoranthene
Concen: 36.636 ng
RT: 15.039 min Scan# 2168
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Tgt Ion: 252 Resp: 771455
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 10.9 8.3 12.5





#89

Benzo(k)fluoranthene

Concen: 34.912 ng

RT: 15.068 min Scan# 2173

Delta R.T. -0.000 min

Lab File: BF129780.D

Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:252 Resp: 705028

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.1	25.7
125	10.9	8.3	12.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#90

Benzo(a)pyrene

Concen: 36.360 ng

RT: 15.409 min Scan# 2231




Delta R.T. -0.000 min


Lab File: BF129780.D

Acq: 15 Aug 2022 00:34

Tgt Ion:252 Resp: 602946

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.2	25.8
125	10.9	8.8	13.2






#91
Dibenzo(a,h)anthracene
Concen: 40.077 ng
RT: 16.962 min Scan# 2495
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC






Tgt Ion:278 Resp: 698110
Ion Ratio Lower Upper
278 100
139 15.8 12.2 18.4
279 24.0 18.7 28.1


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/15/2022
Supervised By :Sohil Jodhani 08/18/2022



#92
Benzo(g,h,i)perylene
Concen: 39.725 ng
RT: 17.403 min Scan# 2570
Delta R.T. -0.000 min
Lab File: BF129780.D
Acq: 15 Aug 2022 00:34



Tgt Ion:276 Resp: 690367
Ion Ratio Lower Upper
276 100
277 23.2 18.2 27.4
138 21.0 15.8 23.8

