

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081517\
 Data File : BF097678.D
 Acq On : 15 Aug 2017 9:08
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

Sohil
 8/16/2017 6:45:00 PM

Quant Time: Aug 15 13:12:59 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 15 12:05:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.78	152	128373	20.00	ng	0.00
21) Naphthalene-d8	8.06	136	507916	20.00	ng	0.00
38) Acenaphthene-d10	9.82	164	225893	20.00	ng	0.00
63) Phenanthrene-d10	11.30	188	437130	20.00	ng	0.00
75) Chrysene-d12	13.93	240	363540	20.00	ng	0.00
86) Perylene-d12	15.36	264	264956	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.37	112	43928	5.16	ng	-0.01
7) Phenol-d6	6.39	99	61698	6.35	ng	-0.02
23) Nitrobenzene-d5	0.00	82	0d	0.00	ng	
41) 2,4,6-Tribromophenol	10.60	330	7933	4.44	ng	-0.01
44) 2-Fluorobiphenyl	9.14	172	94102	7.07	ng	0.00
78) Terphenyl-d14	12.88	244	90435	5.07	ng	-0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.49	88	11993	3.375	ng	89
3) Pyridine	3.23	79	31078	2.856	ng	87
6) Aniline	6.44	93	41910	3.426	ng	94
8) 2-Chlorophenol	6.56	128	23112	2.538	ng	85
10) Phenol	6.40	94	34140	2.961	ng	94
11) bis(2-Chloroethyl)ether	6.51	93	26747	2.933	ng	98
12) 1,3-Dichlorobenzene	6.72	146	25721	2.621	ng	# 92
13) 1,4-Dichlorobenzene	6.80	146	26848	2.761	ng	# 90
14) 1,2-Dichlorobenzene	6.95	146	24497	2.885	ng	98
15) Benzyl Alcohol	6.92	79	22593	3.671	ng	93
17) 2-Methylphenol	7.02	107	19984	2.844	ng	94
18) Hexachloroethane	7.30	117	9594	2.697	ng	# 79
19) n-Nitroso-di-n-propylamine	7.19	70	20203	3.157	ng	# 90
20) 3+4-Methylphenols	7.17	107	25808	2.812	ng	# 73
22) Acetophenone	7.19	105	39266	3.219	ng	# 83
24) Nitrobenzene	7.36	77	19718	2.187	ng	91
25) Isophorone	7.60	82	49055	2.893	ng	95
27) 2,4-Dimethylphenol	7.71	122	21218	2.637	ng	91
28) bis(2-Chloroethoxy)methane	7.82	93	33006	2.983	ng	99
29) 2,4-Dichlorophenol	7.92	162	16138	2.328	ng	93
30) 1,2,4-Trichlorobenzene	8.00	180	20294	3.071	ng	99
31) Naphthalene	8.09	128	73731	3.079	ng	99
33) 4-Chloroaniline	8.13	127	30608	3.142	ng	95
34) Hexachlorobutadiene	8.21	225	11646	3.324	ng	95
35) Caprolactam	8.45	113	4844	2.179	ng	98
36) 4-Chloro-3-methylphenol	8.60	107	21194	2.802	ng	# 80
37) 2-Methylnaphthalene	8.77	142	44553	2.939	ng	97
39) 1,2,4,5-Tetrachlorobenzene	8.94	216	18802	3.119	ng	# 96
40) Hexachlorocyclopentadiene	8.93	237	6721	3.820	ng	94
42) 2,4,6-Trichlorophenol	9.05	196	10717	2.454	ng	99
43) 2,4,5-Trichlorophenol	9.08	196	10194	2.332	ng	93
45) 1,1'-Biphenyl	9.24	154	58263	3.020	ng	96
46) 2-Chloronaphthalene	9.26	162	42796	3.014	ng	94
48) Acenaphthylene	9.67	152	64182	2.945	ng	98

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081517\
 Data File : BF097678.D
 Acq On : 15 Aug 2017 9:08
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5


Manual Integrations
 APPROVED

Sohil
 8/16/2017 6:45:00 PM

Quant Time: Aug 15 13:12:59 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 15 12:05:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
49) Dimethylphthalate	9.53	163	43188	2.518	nq	98
51) Acenaphthene	9.85	154	41275	3.215	nq	98
54) Dibenzofuran	10.02	168	56483	3.303	nq	99
57) Fluorene	10.36	166	46455	3.615	nq	95
58) 2,3,4,6-Tetrachlorophenol	10.13	232	7938	2.630	nq #	94
59) Diethylphthalate	10.24	149	45482	2.890	nq	96
60) 4-Chlorophenyl-phenylether	10.36	204	21347	4.023	nq	92
62) Azobenzene	10.52	77	55042	3.770	nq	92
65) n-Nitrosodiphenylamine	10.47	169	40141	2.639	nq	97
66) 4-Bromophenyl-phenylether	10.84	248	11579	2.654	nq	91
67) Hexachlorobenzene	10.90	284	11988	2.451	nq #	73
68) Atrazine	11.00	200	10280	2.357	nq	97
69) Pentachlorophenol	11.10	266	4649	2.297	nq	97
70) Phenanthrene	11.32	178	66969	2.859	nq	98
71) Anthracene	11.37	178	64463	2.687	nq	97
72) Carbazole	11.53	167	61758	2.618	nq	97
73) Di-n-butylphthalate	11.87	149	62407	2.140	nq	98
74) Fluoranthene	12.50	202	66825	2.912	nq	98
76) Benzidine	12.63	184	28605	2.121	nq #	93
77) Pyrene	12.73	202	70805	2.379	nq	98
80) Benzo(a)anthracene	13.92	228	55164	2.345	nq	98
82) Chrysene	13.96	228	55611	2.421	nq	99
87) Benzo(b)fluoranthene	14.94	252	37839	2.376	nq	100
88) Benzo(k)fluoranthene	14.97	252	41691	2.747	nq #	92
89) Benzo(a)pyrene	15.30	252	34282	2.352	nq	100
90) Dibenzo(a,h)anthracene	16.77	278	24709	2.067	nq	98
91) Benzo(g,h,i)perylene	17.16	276	25608	2.043	nq #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.78 min Scan# 866
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
152	100		
150	156.7	126.2	189.2
115	69.1	52.2	78.2

Manual Integrations
APPROVED




Sohil
8/16/2017 6:45:00 PM




#2

1,4-Dioxane
Concen: 3.375 ng
RT: 2.49 min Scan# 137
Delta R.T. 0.00 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
88	100		
58	64.3	61.4	92.0
43	27.4	23.4	35.0









#3
 Pvridine
 Concen: 2.856 ng
 RT: 3.23 min Scan# 263
 Delta R.T. 0.02 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
79	100		
52	56.2	54.8	82.2
51	28.4	27.0	40.4




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#5
 2-Fluorophenol
 Concen: 5.158 ng
 RT: 5.37 min Scan# 627
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
112	100		
64	64.4	46.0	69.0
63	33.1	23.4	35.0





#6

Aniline

Concen: 3.426 ng

RT: 6.44 min Scan# 808

Delta R.T. -0.01 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Instrument :

BNA_F





ClientSampleId :

SSTDICC2.5

Tot Ion:	93	Resp:	41910
Ion Ratio	Lower	Upper	
93	100		
66	36.0	32.9	49.3
65	20.8	16.0	24.0

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM



#7

Phenol-d6

Concen: 6.346 ng




RT: 6.39 min Scan# 800


Delta R.T. -0.02 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Tgt Ion:	99	Resp:	61698
Ion Ratio	Lower	Upper	
99	100		
42	15.7	13.5	20.3
71	32.4	24.5	36.7





#8

2-Chlorophenol

Concen: 2.538 ng

RT: 6.56 min Scan# 828

Delta R.T. -0.01 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Instrument :

BNA_F





ClientSampleId :

SSTDICC2.5

Tot Ion:	128	Resp:	23112
Ion Ratio	Lower	Upper	
128	100		
130	32.1	12.9	52.9
64	64.6	28.4	68.4

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM



#10

Phenol

Concen: 2.961 ng




RT: 6.40 min Scan# 802


Delta R.T. -0.02 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Tgt Ion:	94	Resp:	34140
Ion Ratio	Lower	Upper	
94	100		
65	28.7	9.9	49.9
66	38.0	23.1	63.1









#11
 bis(2-Chloroethyl)ether
 Concen: 2.933 ng
 RT: 6.51 min Scan# 820
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion:	93	Resp:	26747
Ion	Ratio	Lower	Upper
93	100		
63	77.3	59.2	99.2
95	32.0	12.8	52.8




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#12
 1,3-Dichlorobenzene
 Concen: 2.621 ng
 RT: 6.72 min Scan# 856
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion:	146	Resp:	25721
Ion	Ratio	Lower	Upper
146	100		
148	66.2	51.8	77.6
75	40.4	23.1	34.7#









#13
 1,4-Dichlorobenzene
 Concen: 2.761 ng
 RT: 6.80 min Scan# 869
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampled :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
146	100		
148	60.6	52.2	78.2
111	47.6	30.3	45.5#




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#14
 1,2-Dichlorobenzene
 Concen: 2.885 ng
 RT: 6.95 min Scan# 895
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
146	100		
148	65.2	50.4	75.6
111	46.4	38.2	57.4









#15
Benzyl Alcohol
Concen: 3.671 ng
RT: 6.92 min Scan# 889
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
79	100			
108	73.2	63.4	95.0	
77	67.5	49.2	73.8	




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#17
2-Methylphenol
Concen: 2.844 ng
RT: 7.02 min Scan# 907
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Resp	Lower	Upper
107	100			
108	112.9	93.4	140.0	
77	52.1	35.8	53.6	
79	47.3	34.8	52.2	









#18
Hexachloroethane
Concen: 2.697 ng
RT: 7.30 min Scan# 954
Delta R.T. 0.00 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion:117 Resp: 9594
Ion Ratio Lower Upper
117 100
119 99.0 77.4 116.0
201 78.9 94.6 141.8#




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#19
n-Nitroso-di-n-propylamine
Concen: 3.157 ng
RT: 7.19 min Scan# 935
Delta R.T. -0.02 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion: 70 Resp: 20203
Ion Ratio Lower Upper
70 100
42 52.1 47.2 70.8
101 6.9 7.2 10.8#
130 13.7 15.9 23.9#









#20
3+4-Methylphenols
Concen: 2.812 ng
RT: 7.17 min Scan# 933
Delta R.T. -0.02 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	107	Resp	25808
Ion Ratio	Lower	Upper	
107	100		
108	87.6	71.0	111.0
77	56.1	90.5	130.5#
79	31.4	8.4	48.4




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.06 min Scan# 1084
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	136	Resp	507916
Ion Ratio	Lower	Upper	
136	100		
137	10.9	8.5	12.7
54	8.5	4.9	7.3#
68	7.5	4.1	6.1#





#22

Acetophenone

Concen: 3.219 ng

RT: 7.19 min Scan# 935

Delta R.T. -0.01 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Instrument :

BNA_F





ClientSampleId :

SSTDICC2.5

Tot Ion	Ratio	Resp Lower	Resp Upper
105	100		
71	9.6	2.8	4.2#
51	24.5	30.7	46.1#
120	18.6	17.4	26.0

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM



#23

Nitrobenzene-d5


Concen: 0.000 ng


Expected RT: 7.35 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Tgt Ion	Exp Ratio
82	100
128	42.5
54	52.7





#24

Nitrobenzene

Concen: 2.187 ng

RT: 7.36 min Scan# 964

Delta R.T. -0.01 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Instrument :

BNA_F





ClientSampleId :

SSTDICC2.5

Tot Ion:	77	Resp:	19718
Ion Ratio	Lower	Upper	
77	100		
123	37.5	35.8	53.8
65	13.3	10.6	15.8

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM



#25

Isophorone

Concen: 2.893 ng




RT: 7.60 min Scan# 1005


Delta R.T. -0.01 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Tgt Ion:	82	Resp:	49055
Ion Ratio	Lower	Upper	
82	100		
95	6.5	5.3	7.9
138	13.2	13.1	19.7









#27
 2,4-Dimethylphenol
 Concen: 2.637 ng
 RT: 7.71 min Scan# 1024
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion:	122	Resp:	21218
Ion Ratio	Lower	Upper	
122	100		
107	125.8	89.2	133.8
121	56.9	45.2	67.8




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#28
 bis(2-Chloroethoxy)methane
 Concen: 2.983 ng
 RT: 7.82 min Scan# 1042
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion:	93	Resp:	33006
Ion Ratio	Lower	Upper	
93	100		
95	33.1	26.5	39.7
123	9.7	9.0	13.4









#29
 2,4-Dichlorophenol
 Concen: 2.328 ng
 RT: 7.92 min Scan# 1059
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
162	100	16138		
164	59.9	44.6	84.6	
98	40.2	14.8	54.8	




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#30
 1,2,4-Trichlorobenzene
 Concen: 3.071 ng
 RT: 8.00 min Scan# 1074
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	20294		
182	96.0	75.8	113.6	
145	32.0	25.3	37.9	






#31
Naphthalene
Concen: 3.079 ng
RT: 8.09 min Scan# 1088
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5




Tot Ion	Ratio	Resp	Lower	Upper
128	100			
129	10.8	9.0	13.6	
127	13.4	10.8	16.2	

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM






Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E




#33
4-Chloroaniline
Concen: 3.142 ng
RT: 8.13 min Scan# 1096
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Resp	Lower	Upper
127	100			
129	29.9	26.2	39.2	
65	31.1	27.8	41.8	
92	19.9	16.8	25.2	

Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0





#34

Hexachlorobutadiene

Concen: 3.324 ng

RT: 8.21 min Scan# 1109

Delta R.T. 0.00 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Instrument :

BNA_F


ClientSampled :

SSTDICC2.5




Tot Ion:	225	Resp:	11646
Ion Ratio	Lower	Upper	
225	100		
223	67.2	48.8	73.2
227	65.8	51.1	76.7

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E



#35

Caprolactam

Concen: 2.179 ng

RT: 8.45 min Scan# 1150




Delta R.T. -0.06 min


Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Tgt Ion:	113	Resp:	4844
Ion Ratio	Lower	Upper	
113	100		
55	172.7	150.2	190.2
56	129.2	107.8	147.8

Abundance Ion 113.00 (112.70 to 113.70): E
Ion 55.00 (54.70 to 55.70): BF0
Ion 56.00 (55.70 to 56.70): BF0









#36
 4-Chloro-3-methylphenol
 Concen: 2.802 ng
 RT: 8.60 min Scan# 1176
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
107	100		
144	24.2	24.2	36.4#
142	70.0	73.4	110.0#




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#37
 2-Methylnaphthalene
 Concen: 2.939 ng
 RT: 8.77 min Scan# 1205
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.5	71.3	106.9
115	37.2	27.1	40.7





#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.82 min Scan# 1382

Delta R.T. -0.01 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Instrument :

BNA_F


ClientSampleId :

SSTDICC2.5




Tot Ion:	164	Resp:	225893
Ion Ratio	Lower	Upper	
164	100		
162	102.5	82.6	123.8
160	47.8	36.2	54.2

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E



#39

1,2,4,5-Tetrachlorobenzene

Concen: 3.119 ng

RT: 8.94 min Scan# 1233




Delta R.T. -0.01 min


Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Tgt Ion:	216	Resp:	18802
Ion Ratio	Lower	Upper	
216	100		
214	81.2	63.4	95.2
179	23.0	17.9	26.9
108	26.6	16.5	24.7#

Abundance Ion 216.00 (215.70 to 216.70): E
Ion 214.00 (213.70 to 214.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 108.00 (107.70 to 108.70): E









#40
Hexachlorocyclopentadiene
Concen: 3.820 ng
RT: 8.93 min Scan# 1232
Delta R.T. 0.00 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion:	237	Resp:	6721
Ion Ratio	Lower	Upper	
237	100		
235	57.2	42.6	82.6
272	12.1	0.0	31.9




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#41
2,4,6-Tribromophenol
Concen: 4.445 ng
RT: 10.60 min Scan# 1515
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion:	330	Resp:	7933
Ion Ratio	Lower	Upper	
330	100		
332	97.2	76.6	115.0
141	53.9	31.4	47.2#









#42
 2,4,6-Trichlorophenol
 Concen: 2.454 ng
 RT: 9.05 min Scan# 1252
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion:	196	Resp:	10717
Ion Ratio	Lower	Upper	
196	100		
198	91.5	74.2	111.4
200	31.0	24.0	36.0




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#43
 2,4,5-Trichlorophenol
 Concen: 2.332 ng
 RT: 9.08 min Scan# 1257
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion:	196	Resp:	10194
Ion Ratio	Lower	Upper	
196	100		
198	101.8	75.7	113.5
97	67.8	47.7	71.5
132	35.8	28.0	42.0









#44
2-Fluorobiphenyl
Concen: 7.070 ng
RT: 9.14 min Scan# 1267
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
172	100			
171	34.3	29.6	44.4	
170	23.4	19.8	29.8	




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#45
1,1'-Biphenyl
Concen: 3.020 ng
RT: 9.24 min Scan# 1284
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Resp	Lower	Upper
154	100			
153	40.6	21.2	61.2	
76	16.3	0.0	29.9	






#46
 2-Chloronaphthalene
 Concen: 3.014 ng
 RT: 9.26 min Scan# 1288
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5




Tot Ion:	162	Resp:	42796
Ion Ratio	Lower	Upper	
162	100		
127	42.1	28.3	42.5
164	33.5	27.0	40.4

Manual Integrations
 APPROVED

Sohil
 8/16/2017 6:45:00 PM




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E






#48
 Acenaphthylene
 Concen: 2.945 ng
 RT: 9.67 min Scan# 1358
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion:	152	Resp:	64182
Ion Ratio	Lower	Upper	
152	100		
151	19.4	16.8	25.2
153	13.5	10.8	16.2



Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 151.00 (150.70 to 151.70): E
 Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 2.518 ng
RT: 9.53 min Scan# 1334
Delta R.T. -0.02 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion:	163	Resp:	43188
Ion Ratio	Lower	Upper	
163	100		
194	2.9	2.6	4.0
164	9.7	8.4	12.6




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#51
Acenaphthene
Concen: 3.215 ng
RT: 9.85 min Scan# 1388
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion:	154	Resp:	41275
Ion Ratio	Lower	Upper	
154	100		
153	109.7	90.5	135.7
152	54.6	43.6	65.4









#54
Dibenzenofuran
Concen: 3.303 ng
RT: 10.02 min Scan# 1417
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
168	100	56483		
139	37.0		30.6	45.8
169	13.4		10.8	16.2




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#57
Fluorene
Concen: 3.615 ng
RT: 10.36 min Scan# 1475
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Resp	Lower	Upper
166	100	46455		
165	93.0		78.8	118.2
167	12.6		10.6	16.0









#58
 2,3,4,6-Tetrachlorophenol
 Concen: 2.630 ng
 RT: 10.13 min Scan# 1436
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
232	100		
131	52.2	44.8	67.2
130	1.7	2.3	3.5
166	31.7	29.0	43.4




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#59
 Diethylphthalate
 Concen: 2.890 ng
 RT: 10.24 min Scan# 1454
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
149	100		
177	19.0	16.7	25.1
150	11.6	10.2	15.2









#60
 4-Chlorophenyl-phenylether
 Concen: 4.023 ng
 RT: 10.36 min Scan# 1474
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion:	204	Resp:	21347
Ion Ratio	Lower	Upper	
204	100		
206	34.3	26.2	39.2
141	67.0	60.8	91.2




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#62
 Azobenzene
 Concen: 3.770 ng
 RT: 10.52 min Scan# 1501
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion:	77	Resp:	55042
Ion Ratio	Lower	Upper	
77	100		
182	15.8	0.1	40.1
105	20.5	3.6	43.6
51	24.7	9.4	49.4









#63
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.30 min Scan# 1634
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion:	188	Resp:	437130
Ion Ratio	Lower	Upper	
188	100		
94	12.0	9.4	14.2
80	13.6	10.2	15.2




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#65
 n-Nitrosodiphenylamine
 Concen: 2.639 ng
 RT: 10.47 min Scan# 1494
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion:	169	Resp:	40141
Ion Ratio	Lower	Upper	
169	100		
168	69.0	53.2	79.8
167	35.4	29.5	44.3









#66
 4-Bromophenyl-phenylether
 Concen: 2.654 ng
 RT: 10.84 min Scan# 1557
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
248	100		
250	99.9	76.8	115.2
141	98.5	68.6	103.0




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#67
 Hexachlorobenzene
 Concen: 2.451 ng
 RT: 10.90 min Scan# 1567
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
284	100		
142	50.7	22.8	34.2#
249	37.5	24.0	36.0#






#68
Atrazine
Concen: 2.357 ng
RT: 11.00 min Scan# 1583
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp Lower	Upper
200	100		
173	30.2	6.3	46.3
215	46.8	27.2	67.2

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM






Abundance

Ion 200.00 (199.70 to 200.70): E


Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#69
Pentachlorophenol
Concen: 2.297 ng
RT: 11.10 min Scan# 1600
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Resp Lower	Upper
266	100		
268	60.4	51.1	76.7
264	62.7	51.7	77.5






Abundance

Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#70
Phenanthrene
Concen: 2.859 ng
RT: 11.32 min Scan# 1638
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.0	16.2	24.4
179	15.5	12.6	19.0




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#71
Anthracene
Concen: 2.687 ng
RT: 11.37 min Scan# 1647
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.8	23.8
179	14.6	12.7	19.1









#72
 Carbazole
 Concen: 2.618 ng
 RT: 11.53 min Scan# 1673
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
167	100		
166	22.0	18.1	27.1
139	12.4	11.7	17.5




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#73
 Di-n-butylphthalate
 Concen: 2.140 ng
 RT: 11.87 min Scan# 1731
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
149	100		
150	10.1	7.7	11.5
104	5.8	4.0	6.0









#74
Fluoranthene
 Concen: 2.912 ng
 RT: 12.50 min Scan# 1839
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Resp Lower	Upper
202	100		
101	14.0	0.0	33.2
203	17.2	0.0	38.1




Manual Integrations
APPROVED


Sohil
 8/16/2017 6:45:00 PM



#75
Chrysene-d12
 Concen: 20.000 ng
 RT: 13.93 min Scan# 2082
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Resp Lower	Upper
240	100		
120	12.4	5.8	8.8#
236	26.3	20.5	30.7





#76

Benzidine

Concen: 2.121 ng

RT: 12.63 min Scan# 1861

Delta R.T. 0.00 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Instrument :

BNA_F





ClientSampled :

SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
184	100		
185	14.8	15.5	23.3#
183	10.9	9.8	14.6

Manual Integrations
APPROVED

Sohil
8/16/2017 6:45:00 PM



#77

Pyrene

Concen: 2.379 ng




RT: 12.73 min Scan# 1878


Delta R.T. -0.01 min

Lab File: BF097678.D

Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.6	17.6	26.4
203	18.3	14.4	21.6









#78
 Terphenyl-d14
 Concen: 5.072 ng
 RT: 12.88 min Scan# 1903
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampled :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
244	100		
212	7.9	6.0	9.0
122	14.1	10.3	15.5




Manual Integrations
 APPROVED


Sohil
 8/16/2017 6:45:00 PM



#80
 Benzo(a)anthracene
 Concen: 2.345 ng
 RT: 13.92 min Scan# 2080
 Delta R.T. -0.01 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.5	22.6	33.8
229	19.1	16.3	24.5









#82
Chrysene
Concen: 2.421 ng
RT: 13.96 min Scan# 2086
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion:	228	Resp:	55611
Ion Ratio	Lower	Upper	
228	100		
226	31.3	24.9	37.3
229	19.0	15.8	23.6




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.36 min Scan# 2325
Delta R.T. 0.00 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion:	264	Resp:	264956
Ion Ratio	Lower	Upper	
264	100		
260	24.6	19.5	29.3
265	21.5	17.2	25.8









#87
Benzo(b)fluoranthene
Concen: 2.376 ng
RT: 14.94 min Scan# 2254
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.6	18.1	27.1
125	12.0	9.8	14.8




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#88
Benzo(k)fluoranthene
Concen: 2.747 ng
RT: 14.97 min Scan# 2259
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.1	18.4	27.6
125	10.9	13.1	19.7#









#89
Benzo(a)pyrene
Concen: 2.352 ng
RT: 15.30 min Scan# 2314
Delta R.T. -0.01 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tot Ion:	252	Resp:	34282
Ion Ratio	Lower	Upper	
252	100		
253	21.8	17.5	26.3
125	13.5	11.0	16.4




Manual Integrations
APPROVED


Sohil
8/16/2017 6:45:00 PM



#90
Dibenzo(a,h)anthracene
Concen: 2.067 ng
RT: 16.77 min Scan# 2564
Delta R.T. -0.02 min
Lab File: BF097678.D
Acq: 15 Aug 2017 9:08

Tgt Ion:	278	Resp:	24709
Ion Ratio	Lower	Upper	
278	100		
139	22.7	16.6	24.8
279	23.9	19.3	28.9





#91
 Benzo(a,h,i)perylene
 Concen: 2.043 ng
 RT: 17.16 min Scan# 2631
 Delta R.T. -0.02 min
 Lab File: BF097678.D
 Acq: 15 Aug 2017 9:08

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.2	18.6	28.0
138	25.0	25.4	38.0

Manual Integrations
 APPROVED

Sohil
 8/16/2017 6:45:00 PM

