


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081518\
 Data File : BF108158.D
 Acq On : 15 Aug 2018 11:16
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 08 12:24:24 2018



AutoFind: Scans 1607, 1608, 1609; Background Corrected with Scan 1602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.2	51124	PASS
68	69	0.00	2	1.6	701	PASS
69	198	0.00	100	34.4	44875	PASS
70	69	0.00	2	0.8	367	PASS
127	198	10	80	43.4	56490	PASS
197	198	0.00	2	0.7	893	PASS
198	198	100	100	100.0	130304	PASS
199	198	5	9	6.9	8944	PASS
275	198	10	60	26.7	34754	PASS
365	198	1	100	3.6	4708	PASS
441	198	0.01	100	13.6	17740	PASS
442	198	50	100	84.5	110114	PASS
443	442	15	24	18.9	20793	PASS