

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081519\
 Data File : BF116304.D
 Acq On : 16 Aug 2019 3:00
 Operator : HP/JU
 Sample : K4319-07
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEST-PIT-2

Quant Time: Aug 16 05:18:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 11:13:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	98348	20.00	ng	-0.03
21) Naphthalene-d8	8.10	136	383664	20.00	ng	-0.03
39) Acenaphthene-d10	9.85	164	208295	20.00	ng	-0.04
64) Phenanthrene-d10	11.33	188	384180	20.00	ng	-0.04
76) Chrysene-d12	13.97	240	342774	20.00	ng	-0.04
87) Perylene-d12	15.43	264	326276	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	604699	102.93	ng	-0.03
7) Phenol-d6	6.46	99	793332	107.03	ng	-0.03
23) Nitrobenzene-d5	7.38	82	498998	75.08	ng	-0.04
42) 2,4,6-Tribromophenol	10.64	330	219683	108.78	ng	-0.04
45) 2-Fluorobiphenyl	9.17	172	894385	80.43	ng	-0.04
79) Terphenyl-d14	12.92	244	1075193	66.10	ng	-0.04
Target Compounds						
50) Dimethylphthalate	9.57	163	99118	6.989	ng	99
80) Butylbenzylphthalate	13.38	149	31939	2.922	ng	95
84) Bis(2-ethylhexyl)phthalate	13.94	149	45479	3.202	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081519\
Data File : BF116304.D
Acq On : 16 Aug 2019 3:00
Operator : HP/JU
Sample : K4319-07
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEST-PIT-2

Quant Time: Aug 16 05:18:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 07 11:13:18 2019
Response via : Initial Calibration

