

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081519\
 Data File : BF116306.D
 Acq On : 16 Aug 2019 3:54
 Operator : HP/JU
 Sample : K4319-05 5X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEST-PIT-1

Quant Time: Aug 16 05:25:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 11:13:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	60044	20.00	ng	-0.02
21) Naphthalene-d8	8.10	136	250198	20.00	ng	-0.03
39) Acenaphthene-d10	9.85	164	135511	20.00	ng	-0.04
64) Phenanthrene-d10	11.34	188	251212	20.00	ng	-0.03
76) Chrysene-d12	13.98	240	213315	20.00	ng	-0.03
87) Perylene-d12	15.44	264	196056	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	63199	17.62	ng	0.00
7) Phenol-d6	6.49	99	108140	23.90	ng	0.00
23) Nitrobenzene-d5	7.40	82	63979	14.76	ng	-0.02
42) 2,4,6-Tribromophenol	10.65	330	28369	21.59	ng	-0.02
45) 2-Fluorobiphenyl	9.17	172	127439	17.62	ng	-0.04
79) Terphenyl-d14	12.91	244	159532	15.76	ng	-0.04
Target Compounds						
80) Butylbenzylphthalate	13.39	149	357302	52.530	ng	98
84) Bis(2-ethylhexyl)phthalate	13.94	149	203991	23.078	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081519\
Data File : BF116306.D
Acq On : 16 Aug 2019 3:54
Operator : HP/JU
Sample : K4319-05 5X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEST-PIT-1

Quant Time: Aug 16 05:25:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 07 11:13:18 2019
Response via : Initial Calibration

