

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081524\
 Data File : BF139045.D
 Acq On : 16 Aug 2024 02:49
 Operator : RC/JU
 Sample : P3548-06RE
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 343RE

Quant Time: Aug 16 04:34:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.834	152	38887	20.000	ng	-0.01
21) Naphthalene-d8	8.116	136	152683	20.000	ng	-0.01
39) Acenaphthene-d10	9.869	164	72754	20.000	ng	-0.01
64) Phenanthrene-d10	11.351	188	100045	20.000	ng	-0.02
76) Chrysene-d12	13.998	240	78585	20.000	ng	# 0.00
86) Perylene-d12	15.463	264	61905	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	176504	70.065	ng	0.00
7) Phenol-d6	6.481	99	248601	73.502	ng	0.00
23) Nitrobenzene-d5	7.398	82	148265	47.477	ng	-0.01
42) 2,4,6-Tribromophenol	10.657	330	35070	58.847	ng	-0.01
45) 2-Fluorobiphenyl	9.187	172	177786	36.716	ng	-0.02
79) Terphenyl-d14	12.933	244	135464	28.861	ng	-0.01
Target Compounds						
						Qvalue
75) Fluoranthene	12.569	202	14490	3.013	ng	96
78) Pyrene	12.798	202	16674	2.254	ng	97
84) Bis(2-ethylhexyl)phtha...	13.957	149	8050	2.320	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081524\
Data File : BF139045.D
Acq On : 16 Aug 2024 02:49
Operator : RC/JU
Sample : P3548-06RE
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
343RE

Quant Time: Aug 16 04:34:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration

