

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081622\
 Data File : BF129813.D
 Acq On : 16 Aug 2022 16:22
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 17 02:33:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	37.437	6.4	112	0.00
3	Pyridine	40.000	38.425	3.9	115	0.00
4	n-Nitrosodimethylamine	40.000	38.128	4.7	110	0.00
5 S	2-Fluorophenol	80.000	72.858	8.9	111	0.00
6	Aniline	40.000	36.586	8.5	112	0.00
7 S	Phenol-d6	80.000	72.271	9.7	111	0.00
8	2-Chlorophenol	40.000	36.803	8.0	111	0.00
9	Benzaldehyde	40.000	38.901	2.7	110	0.00
10 C	Phenol	40.000	36.277	9.3	111	0.00
11	bis(2-Chloroethyl)ether	40.000	36.794	8.0	110	0.00
12	1,3-Dichlorobenzene	40.000	36.573	8.6	111	0.00
13 C	1,4-Dichlorobenzene	40.000	36.452	8.9	111	0.00
14	1,2-Dichlorobenzene	40.000	36.143	9.6	110	0.00
15	Benzyl Alcohol	40.000	35.585	11.0	108	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.262	9.3	109	0.00
17	2-Methylphenol	40.000	36.749	8.1	112	0.00
18	Hexachloroethane	40.000	37.199	7.0	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.574	8.6	109	0.00
20	3+4-Methylphenols	40.000	36.423	8.9	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	36.930	7.7	111	0.00
23 S	Nitrobenzene-d5	80.000	74.556	6.8	111	0.00
24	Nitrobenzene	40.000	37.029	7.4	110	0.00
25	Isophorone	40.000	36.892	7.8	109	0.00
26 C	2-Nitrophenol	40.000	38.012	5.0	111	0.00
27	2,4-Dimethylphenol	40.000	36.789	8.0	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.940	7.7	108	0.00
29 C	2,4-Dichlorophenol	40.000	37.378	6.6	109	0.00
30	1,2,4-Trichlorobenzene	40.000	37.008	7.5	110	0.00
31	Naphthalene	40.000	36.540	8.7	110	0.00
32	Benzoic acid	40.000	38.217	4.5	107	0.00
33	4-Chloroaniline	40.000	36.455	8.9	108	0.00
34 C	Hexachlorobutadiene	40.000	36.764	8.1	108	0.00
35	Caprolactam	40.000	37.831	5.4	112	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.346	6.6	111	0.00
37	2-Methylnaphthalene	40.000	36.672	8.3	109	0.00
38	1-Methylnaphthalene	40.000	36.954	7.6	111	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.518	8.7	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.828	-2.1	114	0.00
42 S	2,4,6-Tribromophenol	80.000	74.965	6.3	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.597	6.0	110	0.00
44	2,4,5-Trichlorophenol	40.000	37.344	6.6	108	0.00
45 S	2-Fluorobiphenyl	80.000	72.307	9.6	109	0.00
46	1,1'-Biphenyl	40.000	36.362	9.1	109	0.00
47	2-Chloronaphthalene	40.000	36.290	9.3	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081622\
 Data File : BF129813.D
 Acq On : 16 Aug 2022 16:22
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 17 02:33:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.683	5.8	110	0.00
49	Acenaphthylene	40.000	36.189	9.5	108	0.00
50	Dimethylphthalate	40.000	36.321	9.2	108	0.00
51	2,6-Dinitrotoluene	40.000	37.136	7.2	108	0.00
52 C	Acenaphthene	40.000	36.197	9.5	108	0.00
53	3-Nitroaniline	40.000	37.761	5.6	111	0.00
54 P	2,4-Dinitrophenol	40.000	37.561	6.1	109	0.00
55	Dibenzofuran	40.000	36.037	9.9	110	0.00
56 P	4-Nitrophenol	40.000	40.465	-1.2	112	0.00
57	2,4-Dinitrotoluene	40.000	36.526	8.7	111	0.00
58	Fluorene	40.000	35.800	10.5	109	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.825	7.9	105	0.00
60	Diethylphthalate	40.000	35.787	10.5	108	0.00
61	4-Chlorophenyl-phenylether	40.000	36.486	8.8	110	0.00
62	4-Nitroaniline	40.000	37.829	5.4	113	0.00
63	Azobenzene	40.000	36.523	8.7	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.067	-2.7	112	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.880	7.8	110	0.00
67	4-Bromophenyl-phenylether	40.000	36.546	8.6	108	0.00
68	Hexachlorobenzene	40.000	37.276	6.8	111	0.00
69	Atrazine	40.000	36.601	8.5	110	0.00
70 C	Pentachlorophenol	40.000	36.365	9.1	105	0.00
71	Phenanthrene	40.000	36.466	8.8	109	0.00
72	Anthracene	40.000	36.580	8.6	109	0.00
73	Carbazole	40.000	36.436	8.9	111	0.00
74	Di-n-butylphthalate	40.000	35.782	10.5	108	0.00
75 C	Fluoranthene	40.000	36.877	7.8	113	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	126	0.00
77	Benzydine	40.000	46.803	-17.0	146	0.00
78	Pyrene	40.000	35.667	10.8	115	0.00
79 S	Terphenyl-d14	80.000	69.108	13.6	113	0.00
80	Butylbenzylphthalate	40.000	37.552	6.1	125	0.00
81	Benzo(a)anthracene	40.000	36.330	9.2	125	0.00
82	3,3'-Dichlorobenzidine	40.000	36.720	8.2	129	0.00
83	Chrysene	40.000	36.201	9.5	125	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.186	-0.5	139	0.00
85 c	Di-n-octyl phthalate	40.000	40.090	-0.2	140	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.237	4.4	110	0.00
88	Benzo(b)fluoranthene	40.000	36.674	8.3	113	0.00
89	Benzo(k)fluoranthene	40.000	38.321	4.2	128	0.00
90 C	Benzo(a)pyrene	40.000	36.728	8.2	116	0.00
91	Dibenzo(a,h)anthracene	40.000	38.215	4.5	108	0.00
92	Benzo(g,h,i)perylene	40.000	39.093	2.3	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081622\
Data File : BF129813.D
Acq On : 16 Aug 2022 16:22
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 17 02:33:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 15 02:28:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0