

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081715\
 Data File : BF081035.D
 Acq On : 18 Aug 2015 7:56
 Operator : UM/IZ
 Sample : G3310-01
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P001-DS01-0012-01

Quant Time: Aug 18 08:31:41 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 17 17:57:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.63	152	61215	20.00	ng	0.00
21) Naphthalene-d8	7.92	136	215147	20.00	ng	0.01
38) Acenaphthene-d10	9.66	164	85008	20.00	ng	0.00
63) Phenanthrene-d10	11.13	188	228891	20.00	ng	0.00
75) Chrysene-d12	13.77	240	113352	20.00	ng	0.02
86) Perylene-d12	15.12	264	81787	20.00	ng	0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	392687	96.49	ng	0.02
7) Phenol-d6	6.27	99	542895	102.24	ng	0.01
23) Nitrobenzene-d5	7.20	82	380203	80.73	ng	0.01
41) 2,4,6-Tribromophenol	10.44	330	69448	94.25	ng	0.00
44) 2-Fluorobiphenyl	8.99	172	489480	75.45	ng	0.00
78) Terphenyl-d14	12.78	244	202664	38.33	ng	0.07


Target Compounds						Qvalue
9) Benzaldehyde	6.40	77	11634	3.40	ng	85
19) n-Nitroso-di-n-propylamine	7.20	70	52188	14.19	ng	# 69
30) 1,2,4-Trichlorobenzene	7.86	180	17520	5.17	ng	97
48) Acenaphthylene	9.52	152	85779	7.97	ng	99
49) Dimethylphthalate	9.39	163	102855	14.69	ng	99
66) 4-Bromophenyl-phenylether	10.44	248	5780	2.57	ng	# 1
71) Anthracene	11.21	178	47610	3.61	ng	99
73) Di-n-butylphthalate	11.68	149	103022	6.70	ng	# 1
74) Fluoranthene	12.39	202	67260	4.59	ng	93
76) Benzidine	12.50	184	603250	140.88	ng	# 81
77) Pyrene	12.62	202	89757	9.74	ng	98
80) Benzo(a)anthracene	13.77	228	52709	6.93	ng	# 58
81) 3,3'-Dichlorobenzidine	13.71	252	25747	10.46	ng	# 2
82) Chrysene	13.77	228	52709	7.69	ng	# 62
83) Bis(2-ethylhexyl)phthalate	13.77	149	22225	4.38	ng	# 97
85) Indeno(1,2,3-cd)pyrene	16.34	276	31622	6.57	ng	96
87) Benzo(b)fluoranthene	14.75	252	83978	14.52	ng	# 92
88) Benzo(k)fluoranthene	14.75	252	83626	15.51	ng	# 95
89) Benzo(a)pyrene	15.06	252	46218	9.17	ng	# 91
90) Dibenzo(a,h)anthracene	16.35	278	10287	2.62	ng	# 74
91) Benzo(g,h,i)perylene	16.72	276	48348	12.13	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF081715\
Data File : BF081035.D
Acq On : 18 Aug 2015 7:56
Operator : UM/IZ
Sample : G3310-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01

Quant Time: Aug 18 08:31:41 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 17 17:57:25 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.63 min Scan# 406

Delta R.T. -0.00 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01


Tot Ion:152 Resp: 61215

Ion Ratio Lower Upper

152 100

150 158.0 123.8 185.6




115 64.8 52.1 78.1



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 96.49 ng

RT: 5.21 min Scan# 282

Delta R.T. 0.02 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56


Tgt Ion:112 Resp: 392687

Ion Ratio Lower Upper

112 100

64 77.1 66.9 100.3



63 41.1 38.1 57.1




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 102.24 ng

RT: 6.27 min Scan# 375

Delta R.T. 0.01 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01


Tot Ion: 99 Resp: 542895

Ion Ratio Lower Upper

99 100

42 24.3 19.6 29.4




71 37.0 31.2 46.8



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#9

Benzaldehyde

Concen: 3.40 ng

RT: 6.40 min Scan# 386

Delta R.T. 0.24 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56


Tgt Ion: 77 Resp: 11634

Ion Ratio Lower Upper

77 100

105 80.1 69.6 109.6



106 66.6 64.0 104.0






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0





#19

n-Nitroso-di-n-propylamine

Concen: 14.19 ng

RT: 7.20 min Scan# 456

Delta R.T. 0.15 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01

Tot Ion: 70 Resp: 52188





Ion Ratio Lower Upper

70 100

42 53.5 65.0 97.4#

101 0.0 7.4 11.2#

130 1.7 13.0 19.4#



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.92 min Scan# 519

Delta R.T. 0.01 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Tgt Ion: 136 Resp: 215147


Ion Ratio Lower Upper

136 100

137 10.5 8.2 12.2

54 7.5 7.7 11.5#

68 3.9 3.4 5.2









#23
 Nitrobenzene-d5
 Concen: 80.73 ng
 RT: 7.20 min Scan# 456
 Delta R.T. 0.01 min
 Lab File: BF081035.D
 Acq: 18 Aug 2015 7:56




Instrument :
 BNA_F
 ClientSampleId :
 P001-DS01-0012-01


Tot Ion	82	Resp	380203
Ion Ratio	Lower	Upper	
82	100		
128	39.5	28.6	42.8
54	62.2	55.1	82.7



#30
 1,2,4-Trichlorobenzene
 Concen: 5.17 ng
 RT: 7.86 min Scan# 514
 Delta R.T. -0.00 min
 Lab File: BF081035.D
 Acq: 18 Aug 2015 7:56

Tgt Ion	180	Resp	17520
Ion Ratio	Lower	Upper	
180	100		
182	97.5	76.4	114.6
145	28.3	25.4	38.2









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.66 min Scan# 671
Delta R.T. -0.00 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01

Tot Ion:164 Resp: 85008
Ion Ratio Lower Upper
164 100
162 101.6 83.8 125.8
160 45.6 39.8 59.8




Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#41
2,4,6-Tribromophenol
Concen: 94.25 ng
RT: 10.44 min Scan# 740
Delta R.T. -0.00 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Tgt Ion:330 Resp: 69448
Ion Ratio Lower Upper
330 100
332 95.5 77.3 115.9
141 57.6 44.9 67.3



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#44
 2-Fluorobiphenyl
 Concen: 75.45 ng
 RT: 8.99 min Scan# 613
 Delta R.T. -0.00 min
 Lab File: BF081035.D
 Acq: 18 Aug 2015 7:56

Instrument :
 BNA_F
 ClientSampleId :
 P001-DS01-0012-01

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.9	43.3
170	24.2	19.4	29.2




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E






#48
 Acenaphthylene
 Concen: 7.97 ng
 RT: 9.52 min Scan# 659
 Delta R.T. -0.00 min
 Lab File: BF081035.D
 Acq: 18 Aug 2015 7:56

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.1	24.1
153	13.5	9.8	14.6



Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 151.00 (150.70 to 151.70): E
 Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 14.69 ng
RT: 9.39 min Scan# 648
Delta R.T. -0.00 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56




Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01


Tot Ion	Ratio	Lower	Upper
163	100		
194	3.3	2.5	3.7
164	9.9	8.1	12.1



#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.13 min Scan# 800
Delta R.T. -0.00 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Tgt Ion	Ratio	Lower	Upper
188	100		
94	13.3	8.2	12.2#
80	14.8	9.5	14.3#









#66
4-Bromophenyl-phenylether
Concen: 2.57 ng
RT: 10.44 min Scan# 740
Delta R.T. -0.24 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56




Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01


Tot Ion:248 Resp: 5780
Ion Ratio Lower Upper
248 100
250 216.1 77.8 116.6#
141 591.6 39.8 59.6#



#71
Anthracene
Concen: 3.61 ng
RT: 11.21 min Scan# 807
Delta R.T. 0.01 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Tgt Ion:178 Resp: 47610
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 15.6 12.7 19.1





#73

Di-n-butylphthalate

Concen: 6.70 ng

RT: 11.68 min Scan# 848

Delta R.T. -0.03 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01


Tot Ion:149 Resp: 103022

Ion Ratio Lower Upper

149 100

150 268.2 7.1 10.7#




104 0.0 3.9 5.9#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 4.59 ng

RT: 12.39 min Scan# 910

Delta R.T. 0.06 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56


Tgt Ion:202 Resp: 67260

Ion Ratio Lower Upper

202 100

101 12.2 0.0 24.8



203 17.9 0.0 36.9




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.77 min Scan# 1031

Delta R.T. 0.02 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01


Tot Ion:240 Resp: 113352

Ion Ratio Lower Upper

240 100

120 16.7 5.6 8.4#




236 25.5 20.6 30.8



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 140.88 ng

RT: 12.50 min Scan# 920

Delta R.T. 0.03 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56


Tgt Ion:184 Resp: 603250

Ion Ratio Lower Upper

184 100

185 19.6 12.0 18.0#



183 22.3 8.9 13.3#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pvrene

Concen: 9.74 ng

RT: 12.62 min Scan# 930

Delta R.T. 0.06 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01


Tot Ion:202 Resp: 89757

Ion Ratio Lower Upper

202 100

200 19.6 16.6 24.8




203 17.3 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 38.33 ng

RT: 12.78 min Scan# 944

Delta R.T. 0.07 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56


Tgt Ion:244 Resp: 202664

Ion Ratio Lower Upper

244 100

212 6.7 5.8 8.8



122 29.7 8.2 12.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#80
Benzo(a)anthracene
Concen: 6.93 ng
RT: 13.77 min Scan# 1031
Delta R.T. 0.03 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56




Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01


Tot Ion:228 Resp: 52709
Ion Ratio Lower Upper
228 100
226 60.3 22.3 33.5#
229 26.0 15.9 23.9#



#81
3,3'-Dichlorobenzidine
Concen: 10.46 ng
RT: 13.71 min Scan# 1026
Delta R.T. -0.00 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Tgt Ion:252 Resp: 25747
Ion Ratio Lower Upper
252 100
254 111.5 51.5 77.3#
126 122.3 12.5 18.7#









#82
Chrysene
Concen: 7.69 ng
RT: 13.77 min Scan# 1031
Delta R.T. -0.00 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56




Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01


Tot Ion:228 Resp: 52709
Ion Ratio Lower Upper
228 100
226 60.3 24.4 36.6#
229 26.0 16.0 24.0#



#83
Bis(2-ethylhexyl)phthalate
Concen: 4.38 ng
RT: 13.77 min Scan# 1031
Delta R.T. 0.01 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Tgt Ion:149 Resp: 22225
Ion Ratio Lower Upper
149 100
167 24.2 20.7 31.1
279 2.1 2.2 3.4#









#85
Indeno(1,2,3-cd)pyrene
Concen: 6.57 ng
RT: 16.34 min Scan# 1256
Delta R.T. 0.02 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56




Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01


Tot Ion	Ratio	Lower	Upper
276	100		
138	26.5	18.6	28.0
277	25.6	19.8	29.8



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.12 min Scan# 1149
Delta R.T. 0.02 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.6	29.4
265	23.3	17.5	26.3









#87
Benzo(b)fluoranthene
Concen: 14.52 ng
RT: 14.75 min Scan# 1117
Delta R.T. 0.02 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56




Instrument :
BNA_F
ClientSampleId :
P001-DS01-0012-01


Tot Ion:252 Resp: 83978
Ion Ratio Lower Upper
252 100
253 23.5 17.1 25.7
125 16.1 8.0 12.0#



#88
Benzo(k)fluoranthene
Concen: 15.51 ng
RT: 14.75 min Scan# 1117
Delta R.T. -0.01 min
Lab File: BF081035.D
Acq: 18 Aug 2015 7:56

Tgt Ion:252 Resp: 83626
Ion Ratio Lower Upper
252 100
253 23.5 17.6 26.4
125 16.1 10.3 15.5#





#89

Benzo(a)pyrene

Concen: 9.17 ng

RT: 15.06 min Scan# 1144

Delta R.T. 0.02 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01


Tot Ion:252 Resp: 46218

Ion Ratio Lower Upper

252 100

253 22.4 16.6 25.0




125 17.4 7.8 11.8#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 2.62 ng

RT: 16.35 min Scan# 1257

Delta R.T. 0.01 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56


Tgt Ion:278 Resp: 10287

Ion Ratio Lower Upper

278 100

139 33.7 10.8 16.2#



279 29.9 19.0 28.6#




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(a,h,i)perylene

Concen: 12.13 ng

RT: 16.72 min Scan# 1289

Delta R.T. 0.03 min

Lab File: BF081035.D

Acq: 18 Aug 2015 7:56

Instrument :

BNA_F

ClientSampleId :

P001-DS01-0012-01


Tot Ion:276 Resp: 48348

Ion Ratio Lower Upper

276 100

277 25.3 18.2 27.2

138 23.3 13.7 20.5#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

