

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF081715\
 Data File : BF081016.D
 Acq On : 17 Aug 2015 21:56
 Operator : UM/IZ
 Sample : G3239-04
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEC-COMP

Quant Time: Aug 18 01:42:37 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 17 17:57:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.63	152	68706	20.00	ng	0.00
21) Naphthalene-d8	7.91	136	248994	20.00	ng	0.00
38) Acenaphthene-d10	9.66	164	123213	20.00	ng	0.00
63) Phenanthrene-d10	11.13	188	257921	20.00	ng	0.00
75) Chrysene-d12	13.75	240	199382	20.00	ng	0.00
86) Perylene-d12	15.10	264	142956	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.22	112	626184	137.08	ng	0.03
7) Phenol-d6	6.27	99	675806	113.39	ng	0.01
23) Nitrobenzene-d5	7.20	82	577376	105.93	ng	0.01
41) 2,4,6-Tribromophenol	10.44	330	146673	137.33	ng	0.00
44) 2-Fluorobiphenyl	8.99	172	819881	87.19	ng	0.00
78) Terphenyl-d14	12.72	244	782703	84.16	ng	0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF081715\
Data File : BF081016.D
Acq On : 17 Aug 2015 21:56
Operator : UM/IZ
Sample : G3239-04
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEC-COMP

Quant Time: Aug 18 01:42:37 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 17 17:57:25 2015
Response via : Initial Calibration

