

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF081715\
 Data File : BF081034.D
 Acq On : 18 Aug 2015 7:25
 Operator : UM/IZ
 Sample : G3290-11
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 2A(0-2)-5

Quant Time: Aug 18 08:40:21 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 17 17:57:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.63	152	58159	20.00	ng	0.00
21) Naphthalene-d8	7.92	136	207916	20.00	ng	0.01
38) Acenaphthene-d10	9.66	164	89455	20.00	ng	0.00
63) Phenanthrene-d10	11.13	188	153846	20.00	ng	0.00
75) Chrysene-d12	13.75	240	125084	20.00	ng	0.00
86) Perylene-d12	15.11	264	89180	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	405001	104.74	ng	0.02
7) Phenol-d6	6.27	99	482476	95.63	ng	0.01
23) Nitrobenzene-d5	7.20	82	367372	80.72	ng	0.01
41) 2,4,6-Tribromophenol	10.44	330	69228	89.28	ng	0.00
44) 2-Fluorobiphenyl	8.99	172	394734	57.82	ng	0.00
78) Terphenyl-d14	12.72	244	297467	50.98	ng	0.01
Target Compounds						
49) Dimethylphthalate	9.39	163	94038	12.76	ng	Qvalue # 99
74) Fluoranthene	12.33	202	25503	2.59	ng	93
77) Pyrene	12.56	202	24251	2.38	ng	97
79) Butylbenzylphthalate	13.20	149	29604	6.77	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF081715\
Data File : BF081034.D
Acq On : 18 Aug 2015 7:25
Operator : UM/IZ
Sample : G3290-11
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
2A(0-2)-5

Quant Time: Aug 18 08:40:21 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 17 17:57:25 2015
Response via : Initial Calibration

