


Data Path : Z:\HPCHEM1\BNA F\Data\BF081717\
 Data File : BF097811.D
 Acq On : 18 Aug 2017 3:31
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 15 13:25:08 2017



AutoFind: Scans 1609, 1610, 1611; Background Corrected with Scan 1605

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.5	21421	PASS
68	69	0.00	2	1.6	375	PASS
69	198	0.00	100	46.3	23919	PASS
70	69	0.00	2	0.6	137	PASS
127	198	10	80	52.9	27341	PASS
197	198	0.00	2	1.1	594	PASS
198	198	100	100	100.0	51664	PASS
199	198	5	9	6.8	3521	PASS
275	198	10	60	24.1	12473	PASS
365	198	1	100	2.9	1475	PASS
441	442	0.01	24	16.7	7172	PASS
442	198	50	100	83.3	43045	PASS
443	442	15	24	19.6	8435	PASS