

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081718\
 Data File : BF108262.D
 Acq On : 18 Aug 2018 00:30
 Operator : JU/SJ
 Sample : J4501-06
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RA-081318-SSW-15-021

Manual Integrations
 APPROVED

Sohil
 8/20/2018 12:48:39 PM

Quant Time: Aug 18 01:22:50 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 12:24:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.01	152	152702	20.00	ng	0.00
21) Naphthalene-d8	8.30	136	568975	20.00	ng	0.00
38) Acenaphthene-d10	10.06	164	257101	20.00	ng	0.00
63) Phenanthrene-d10	11.55	188	437875	20.00	ng	0.00
75) Chrysene-d12	14.21	240	400151	20.00	ng	-0.01
86) Perylene-d12	15.77	264	310343	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.64	112	729637	86.51	ng	0.01
7) Phenol-d6	6.64	99	950867	84.84	ng	0.00
23) Nitrobenzene-d5	7.58	82	624225	61.22	ng	0.00
41) 2,4,6-Tribromophenol	10.85	330	199280	77.71	ng	0.00
44) 2-Fluorobiphenyl	9.37	172	971322	60.65	ng	0.00
78) Terphenyl-d14	13.14	244	761861	48.41	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.76	163	86942	4.855	ng	Qvalue # 99
70) Phenanthrene	11.58	178	277515	13.437	ng	98
71) Anthracene	11.63	178	67295	3.179	ng	98
73) Di-n-butylphthalate	12.10	149	60050	2.819	ng	97
74) Fluoranthene	12.78	202	360867	16.010	ng	94
77) Pyrene	13.01	202	304385	12.015	ng	99
80) Benzo(a)anthracene	14.20	228	179963	7.837	ng	98
82) Chrysene	14.24	228	167609	7.961	ng	97
85) Indeno(1,2,3-cd)pyrene	17.38	276	45083	2.471	ng	# 90
87) Benzo(b)fluoranthene	15.29	252	165909m	8.947	ng	
88) Benzo(k)fluoranthene	15.32	252	57331m	3.363	ng	
89) Benzo(a)pyrene	15.70	252	100917	6.077	ng	96
91) Benzo(a,h,i)perylene	17.87	276	37737	2.789	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081718\
Data File : BF108262.D
Acq On : 18 Aug 2018 00:30
Operator : JU/SJ
Sample : J4501-06
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RA-081318-SSW-15-021

Manual Integrations
APPROVED

Sohil
8/20/2018 12:48:39 PM

Quant Time: Aug 18 01:22:50 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 12:24:24 2018
Response via : Initial Calibration

