

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081722\
 Data File : BF129838.D
 Acq On : 17 Aug 2022 11:22
 Operator : CG\JU
 Sample : N4218-01DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OR-02-081522DL

Quant Time: Aug 17 13:33:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.822	152	156304	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	595297	20.000	ng	0.00
39) Acenaphthene-d10	9.863	164	275357	20.000	ng	0.00
64) Phenanthrene-d10	11.357	188	354704	20.000	ng	0.00
76) Chrysene-d12	14.010	240	294883	20.000	ng	0.00
86) Perylene-d12	15.486	264	261516	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	129881	13.758	ng	0.00
7) Phenol-d6	6.481	99	165075	13.963	ng	0.00
23) Nitrobenzene-d5	7.386	82	100254	9.015	ng	-0.01
42) 2,4,6-Tribromophenol	10.657	330	27475	9.942	ng	0.00
45) 2-Fluorobiphenyl	9.175	172	194240	10.714	ng	-0.01
79) Terphenyl-d14	12.939	244	137886	7.780	ng	0.00
Target Compounds					Qvalue	
9) Benzaldehyde	6.392	77	58	Below Cal	#	1
25) Isophorone	7.386	82	100252	5.238	ng	# 65
31) Naphthalene	8.128	128	320994	10.313	ng	99
32) Benzoic acid	7.916	122	2196	2.964	ng	# 30
33) 4-Chloroaniline	8.128	127	41489	3.390	ng	# 34
37) 2-Methylnaphthalene	8.816	142	163670	8.012	ng	99
38) 1-Methylnaphthalene	8.916	142	122459	6.168	ng	100
49) Acenaphthylene	9.722	152	93459	3.724	ng	98
51) 2,6-Dinitrotoluene	9.863	165	40551	9.470	ng	# 23
52) Acenaphthene	9.892	154	107155	7.037	ng	97
55) Dibenzofuran	10.069	168	139039	6.078	ng	99
56) 4-Nitrophenol	10.069	139	56097	26.628	ng	# 7
58) Fluorene	10.410	166	223340	11.692	ng	99
71) Phenanthrene	11.380	178	1039487	52.473	ng	100
72) Anthracene	11.433	178	296884	15.070	ng	100
73) Carbazole	11.592	167	92162	5.154	ng	99
75) Fluoranthene	12.574	202	891835	43.995	ng	98
77) Benzidine	12.721	184	359	Below Cal	#	1
78) Pyrene	12.804	202	813895	31.762	ng	99
81) Benzo(a)anthracene	13.998	228	439259	21.654	ng	97
83) Chrysene	14.033	228	342786	18.052	ng	98
87) Indeno(1,2,3-cd)pyrene	16.968	276	132105	7.358	ng	96
88) Benzo(b)fluoranthene	15.057	252	573676	32.036	ng	# 96
89) Benzo(k)fluoranthene	15.057	252	573676	33.405	ng	# 95
90) Benzo(a)pyrene	15.421	252	302325	21.438	ng	96
91) Dibenz(a,h)anthracene	16.968	278	37578	2.537	ng	# 88
92) Benzo(g,h,i)perylene	17.427	276	122029	8.257	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081722\
Data File : BF129838.D
Acq On : 17 Aug 2022 11:22
Operator : CG\JU
Sample : N4218-01DL 5X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL

Quant Time: Aug 17 13:33:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 15 02:28:31 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.822 min Scan# 771
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:152 Resp: 156304
Ion Ratio Lower Upper
152 100
150 157.7 126.9 190.3
115 66.5 53.4 80.0



#5
2-Fluorophenol
Concen: 13.758 ng
RT: 5.463 min Scan# 540
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22


Tgt Ion:112 Resp: 129881
Ion Ratio Lower Upper
112 100
64 62.3 47.6 71.4
63 32.4 24.6 37.0








#7
Phenol-d6
Concen: 13.963 ng
RT: 6.481 min Scan# 713
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL







Tgt Ion: 99 Resp: 165075
Ion Ratio Lower Upper
99 100
42 17.7 15.0 22.4
71 35.7 28.8 43.2



#9
Benzaldehyde
Concen: Below Cal
RT: 6.392 min Scan# 698
Delta R.T. 0.011 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion: 77 Resp: 58
Ion Ratio Lower Upper
77 100
105 0.0 80.3 120.3#
106 0.0 74.3 114.3#





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.104 min Scan# 91

Delta R.T. -0.006 min

Lab File: BF129838.D


Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL



Tgt Ion:136 Resp: 595297




Ion Ratio Lower Upper

136 100

137 10.9 8.4 12.6

54 8.2 6.2 9.2

68 5.8 4.2 6.4



#23

Nitrobenzene-d5

Concen: 9.015 ng

RT: 7.386 min Scan# 867

Delta R.T. -0.012 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion: 82 Resp: 100254


Ion Ratio Lower Upper

82 100

128 43.9 36.1 54.1

54 50.9 40.0 60.0









#25
Isophorone
Concen: 5.238 ng
RT: 7.386 min Scan# 80
Delta R.T. -0.265 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion: 82 Resp: 100252
Ion Ratio Lower Upper
82 100
95 0.0 5.9 8.9#
138 0.0 14.6 21.8#



#31
Naphthalene
Concen: 10.313 ng
RT: 8.128 min Scan# 993
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion: 128 Resp: 320994
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 12.9 10.7 16.1









#32
Benzoic acid
Concen: 2.964 ng
RT: 7.916 min Scan# 91
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:122 Resp: 2196
Ion Ratio Lower Upper
122 100
105 24.2 99.9 139.9#
77 43.2 65.9 105.9#



#33
4-Chloroaniline
Concen: 3.390 ng
RT: 8.128 min Scan# 993
Delta R.T. -0.059 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion:127 Resp: 41489
Ion Ratio Lower Upper
127 100
129 83.6 26.7 40.1#
65 0.0 24.2 36.2#
92 0.0 15.2 22.8#









#37
2-Methylnaphthalene
Concen: 8.012 ng
RT: 8.816 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:142 Resp: 163670
Ion Ratio Lower Upper
142 100
141 89.4 70.9 106.3



#38
1-Methylnaphthalene
Concen: 6.168 ng
RT: 8.916 min Scan# 1127
Delta R.T. -0.006 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion:142 Resp: 122459
Ion Ratio Lower Upper
142 100
141 92.0 73.5 110.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.863 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:164 Resp: 275357
Ion Ratio Lower Upper
164 100
162 105.0 82.0 123.0
160 44.8 35.4 53.0



#42
2,4,6-Tribromophenol
Concen: 9.942 ng
RT: 10.657 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion:330 Resp: 27475
Ion Ratio Lower Upper
330 100
332 97.4 82.6 124.0
141 39.8 27.6 41.4









#45
2-Fluorobiphenyl
Concen: 10.714 ng
RT: 9.175 min Scan# 1173
Delta R.T. -0.012 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:172 Resp: 194240
Ion Ratio Lower Upper
172 100
171 35.9 28.4 42.6
170 23.5 19.1 28.7



#49
Acenaphthylene
Concen: 3.724 ng
RT: 9.722 min Scan# 1264
Delta R.T. -0.006 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion:152 Resp: 93459
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 14.6 10.6 16.0









#51
2,6-Dinitrotoluene
Concen: 9.470 ng
RT: 9.863 min Scan# 11
Delta R.T. 0.206 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:165 Resp: 40551
Ion Ratio Lower Upper
165 100
63 1.3 47.8 71.8#
89 1.3 47.2 70.8#



#52
Acenaphthene
Concen: 7.037 ng
RT: 9.892 min Scan# 1293
Delta R.T. -0.006 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion:154 Resp: 107155
Ion Ratio Lower Upper
154 100
153 115.9 89.8 134.6
152 55.0 42.1 63.1





#55

Dibenzenofuran

Concen: 6.078 ng

RT: 10.069 min Scan# 1323

Delta R.T. -0.006 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL





Tgt Ion:168 Resp: 139039

Ion Ratio Lower Upper

168 100

139 38.6 31.3 46.9

169 13.8 10.7 16.1



#56

4-Nitrophenol

Concen: 26.628 ng

RT: 10.069 min Scan# 1323

Delta R.T. 0.053 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion:139 Resp: 56097


Ion Ratio Lower Upper

139 100

109 1.4 52.9 92.9#

65 1.8 79.5 119.5#









#58
 Fluorene
 Concen: 11.692 ng
 RT: 10.410 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BF129838.D
 Acq: 17 Aug 2022 11:22




Instrument :
 BNA_F
 ClientSampleId :
 OR-02-081522DL


Tgt Ion:166 Resp: 223340
 Ion Ratio Lower Upper
 166 100
 165 98.4 77.9 116.9
 167 14.4 11.0 16.4



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.357 min Scan# 1542
 Delta R.T. -0.000 min
 Lab File: BF129838.D
 Acq: 17 Aug 2022 11:22

Tgt Ion:188 Resp: 354704
 Ion Ratio Lower Upper
 188 100
 94 8.6 7.4 11.2
 80 10.2 7.9 11.9









#71
Phenanthrene
Concen: 52.473 ng
RT: 11.380 min Scan# 1546
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:178 Resp: 1039487
Ion Ratio Lower Upper
178 100
176 19.1 15.5 23.3
179 15.5 12.5 18.7



#72
Anthracene
Concen: 15.070 ng
RT: 11.433 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion:178 Resp: 296884
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.4 12.6 18.8





#73

Carbazole

Concen: 5.154 ng

RT: 11.592 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL





Tgt Ion:167 Resp: 92162

Ion Ratio Lower Upper

167 100

166 21.7 17.1 25.7

139 13.3 11.0 16.6



#75

Fluoranthene

Concen: 43.995 ng

RT: 12.574 min Scan# 1749

Delta R.T. -0.000 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion:202 Resp: 891835


Ion Ratio Lower Upper

202 100

101 12.1 0.0 30.7

203 18.1 0.0 37.5





#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.010 min Scan# 1992

Delta R.T. 0.006 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL





Tgt Ion:240 Resp: 294883

Ion Ratio Lower Upper

240 100

120 11.7 9.2 13.8

236 26.4 20.7 31.1



#77

Benzidine

Concen: Below Cal

RT: 12.721 min Scan# 1774

Delta R.T. 0.029 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion:184 Resp: 359


Ion Ratio Lower Upper

184 100

185 113.6 11.4 17.2#

183 72.9 9.4 14.2#





#78

Pyrene

Concen: 31.762 ng

RT: 12.804 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL





Tgt Ion:202 Resp: 813895

Ion Ratio Lower Upper

202 100

200 20.9 17.0 25.6

203 17.1 14.1 21.1



#79

Terphenyl-d14

Concen: 7.780 ng

RT: 12.939 min Scan# 1811

Delta R.T. -0.000 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion:244 Resp: 137886


Ion Ratio Lower Upper

244 100

212 6.3 5.9 8.9

122 10.1 9.9 14.9





#81

Benzo(a)anthracene

Concen: 21.654 ng

RT: 13.998 min Scan# 1990

Delta R.T. 0.006 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL





Tgt Ion:228 Resp: 439259

Ion Ratio Lower Upper

228 100

226 29.0 21.4 32.0

229 21.0 16.1 24.1



#83

Chrysene

Concen: 18.052 ng

RT: 14.033 min Scan# 1997

Delta R.T. -0.000 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion:228 Resp: 342786


Ion Ratio Lower Upper

228 100

226 30.6 24.4 36.6

229 22.6 15.9 23.9





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.486 min Scan# 21

Delta R.T. 0.018 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL





Tgt Ion:264 Resp: 261516

Ion Ratio Lower Upper

264 100

260 24.9 19.1 28.7

265 21.8 16.7 25.1



#87

Indeno(1,2,3-cd)pyrene

Concen: 7.358 ng

RT: 16.968 min Scan# 2496

Delta R.T. 0.012 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion:276 Resp: 132105


Ion Ratio Lower Upper

276 100

138 21.5 19.8 29.8

277 24.3 20.4 30.6









#88
Benzo(b)fluoranthene
Concen: 32.036 ng
RT: 15.057 min Scan# 2168
Delta R.T. 0.018 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22




Instrument :
BNA_F
ClientSampleId :
OR-02-081522DL


Tgt Ion:252 Resp: 573676
Ion Ratio Lower Upper
252 100
253 23.0 17.5 26.3
125 13.2 8.3 12.5#



#89
Benzo(k)fluoranthene
Concen: 33.405 ng
RT: 15.057 min Scan# 2171
Delta R.T. -0.012 min
Lab File: BF129838.D
Acq: 17 Aug 2022 11:22

Tgt Ion:252 Resp: 573676
Ion Ratio Lower Upper
252 100
253 23.0 17.1 25.7
125 13.2 8.3 12.5#





#90

Benzo(a)pyrene

Concen: 21.438 ng

RT: 15.421 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL





Tgt Ion:252 Resp: 302325

Ion Ratio Lower Upper

252 100

253 23.3 17.2 25.8

125 12.7 8.8 13.2



#91

Dibenzo(a,h)anthracene

Concen: 2.537 ng

RT: 16.968 min Scan# 2496

Delta R.T. 0.006 min

Lab File: BF129838.D

Acq: 17 Aug 2022 11:22




Tgt Ion:278 Resp: 37578


Ion Ratio Lower Upper

278 100

139 22.9 12.2 18.4#

279 27.6 18.7 28.1





#92

Benzo(g,h,i)perylene

Concen: 8.257 ng

RT: 17.427 min Scan# 21

Delta R.T. 0.023 min

Lab File: BF129838.D


Acq: 17 Aug 2022 11:22

Instrument :

BNA_F

ClientSampleId :

OR-02-081522DL



Tgt Ion:276 Resp: 122029

Ion Ratio Lower Upper

276 100

277 25.3 18.2 27.4

138 19.3 15.8 23.8

