

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081722\
 Data File : BF129836.D
 Acq On : 17 Aug 2022 10:13
 Operator : CG\JU
 Sample : PB146911TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB146911TB

Quant Time: Aug 17 13:32:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 02:28:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.822	152	154129	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	622521	20.000	ng	0.00
39) Acenaphthene-d10	9.863	164	348679	20.000	ng	0.00
64) Phenanthrene-d10	11.351	188	662417	20.000	ng	0.00
76) Chrysene-d12	14.004	240	574128	20.000	ng	0.00
86) Perylene-d12	15.480	264	476196	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	1158771	124.479	ng	0.00
7) Phenol-d6	6.481	99	1423904	122.145	ng	0.00
23) Nitrobenzene-d5	7.392	82	926373	79.661	ng	0.00
42) 2,4,6-Tribromophenol	10.663	330	453818	129.680	ng	0.00
45) 2-Fluorobiphenyl	9.181	172	1802634	78.521	ng	0.00
79) Terphenyl-d14	12.945	244	2342143	67.880	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081722\
Data File : BF129836.D
Acq On : 17 Aug 2022 10:13
Operator : CG\JU
Sample : PB146911TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB146911TB

Quant Time: Aug 17 13:32:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 15 02:28:31 2022
Response via : Initial Calibration

