

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081822\
 Data File : BF129873.D
 Acq On : 18 Aug 2022 13:39
 Operator : CG\JU
 Sample : N4189-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB19

Quant Time: Aug 18 14:13:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 18 12:03:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.792	152	129269	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	504103	20.000	ng	0.00
39) Acenaphthene-d10	9.827	164	271828	20.000	ng	0.00
64) Phenanthrene-d10	11.321	188	451900	20.000	ng	0.00
76) Chrysene-d12	13.968	240	236351	20.000	ng	0.00
86) Perylene-d12	15.427	264	263948	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	753674	96.532	ng	0.00
7) Phenol-d6	6.445	99	928098	94.925	ng	0.00
23) Nitrobenzene-d5	7.357	82	588493	62.493	ng	-0.01
42) 2,4,6-Tribromophenol	10.627	330	240543	88.169	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	992032	55.429	ng	0.00
79) Terphenyl-d14	12.904	244	870237	61.266	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081822\
Data File : BF129873.D
Acq On : 18 Aug 2022 13:39
Operator : CG\JU
Sample : N4189-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB19

Quant Time: Aug 18 14:13:51 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 18 12:03:10 2022
Response via : Initial Calibration

