

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081822\
 Data File : BF129868.D
 Acq On : 18 Aug 2022 10:50
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 18 13:35:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 18 12:03:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	86	0.00
2	1,4-Dioxane	40.000	40.069	-0.2	91	0.00
3	Pyridine	40.000	40.835	-2.1	93	0.00
4	n-Nitrosodimethylamine	40.000	40.649	-1.6	90	0.00
5 S	2-Fluorophenol	80.000	80.220	-0.3	93	0.00
6	Aniline	40.000	39.226	1.9	92	0.00
7 S	Phenol-d6	80.000	78.608	1.7	92	0.00
8	2-Chlorophenol	40.000	39.904	0.2	92	0.00
9	Benzaldehyde	40.000	41.614	-4.0	89	0.00
10 C	Phenol	40.000	39.682	0.8	93	0.00
11	bis(2-Chloroethyl)ether	40.000	39.048	2.4	90	0.00
12	1,3-Dichlorobenzene	40.000	39.866	0.3	92	0.00
13 C	1,4-Dichlorobenzene	40.000	39.186	2.0	92	0.00
14	1,2-Dichlorobenzene	40.000	39.513	1.2	92	0.00
15	Benzyl Alcohol	40.000	38.677	3.3	90	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.671	3.3	89	0.00
17	2-Methylphenol	40.000	39.202	2.0	92	0.00
18	Hexachloroethane	40.000	40.453	-1.1	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.560	3.6	88	0.00
20	3+4-Methylphenols	40.000	39.142	2.1	91	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	83	0.00
22	Acetophenone	40.000	39.668	0.8	90	0.00
23 S	Nitrobenzene-d5	80.000	79.693	0.4	90	0.00
24	Nitrobenzene	40.000	39.141	2.1	88	0.00
25	Isophorone	40.000	38.490	3.8	86	0.00
26 C	2-Nitrophenol	40.000	40.112	-0.3	89	0.00
27	2,4-Dimethylphenol	40.000	38.963	2.6	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.485	3.8	85	0.00
29 C	2,4-Dichlorophenol	40.000	39.845	0.4	88	0.00
30	1,2,4-Trichlorobenzene	40.000	39.264	1.8	89	0.00
31	Naphthalene	40.000	39.105	2.2	89	0.00
32	Benzoic acid	40.000	40.601	-1.5	87	0.00
33	4-Chloroaniline	40.000	37.282	6.8	84	0.00
34 C	Hexachlorobutadiene	40.000	39.308	1.7	88	0.00
35	Caprolactam	40.000	37.977	5.1	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.062	2.3	88	0.00
37	2-Methylnaphthalene	40.000	38.468	3.8	87	0.00
38	1-Methylnaphthalene	40.000	38.393	4.0	87	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	80	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.631	0.9	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	48.042	-20.1	107	0.00
42 S	2,4,6-Tribromophenol	80.000	78.171	2.3	83	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.961	0.1	85	0.00
44	2,4,5-Trichlorophenol	40.000	40.558	-1.4	85	0.00
45 S	2-Fluorobiphenyl	80.000	79.135	1.1	87	0.00
46	1,1'-Biphenyl	40.000	40.032	-0.1	87	0.00
47	2-Chloronaphthalene	40.000	39.837	0.4	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081822\
 Data File : BF129868.D
 Acq On : 18 Aug 2022 10:50
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 18 13:35:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 18 12:03:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.705	-1.8	86	0.00
49	Acenaphthylene	40.000	39.328	1.7	85	0.00
50	Dimethylphthalate	40.000	38.302	4.2	83	0.00
51	2,6-Dinitrotoluene	40.000	39.698	0.8	84	0.00
52 C	Acenaphthene	40.000	39.514	1.2	85	0.00
53	3-Nitroaniline	40.000	39.595	1.0	85	0.00
54 P	2,4-Dinitrophenol	40.000	42.388	-6.0	91	0.00
55	Dibenzofuran	40.000	39.095	2.3	86	0.00
56 P	4-Nitrophenol	40.000	40.504	-1.3	81	0.00
57	2,4-Dinitrotoluene	40.000	39.180	2.1	86	0.00
58	Fluorene	40.000	38.718	3.2	85	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.549	-3.9	86	0.00
60	Diethylphthalate	40.000	37.875	5.3	83	0.00
61	4-Chlorophenyl-phenylether	40.000	38.250	4.4	83	0.00
62	4-Nitroaniline	40.000	39.910	0.2	86	0.00
63	Azobenzene	40.000	39.264	1.8	85	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	81	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.429	-3.6	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.145	2.1	86	0.00
67	4-Bromophenyl-phenylether	40.000	38.621	3.4	84	0.00
68	Hexachlorobenzene	40.000	38.415	4.0	84	0.00
69	Atrazine	40.000	39.153	2.1	87	0.00
70 C	Pentachlorophenol	40.000	45.698	-14.2	100	0.00
71	Phenanthrene	40.000	39.529	1.2	87	0.00
72	Anthracene	40.000	38.738	3.2	85	0.00
73	Carbazole	40.000	38.962	2.6	87	0.00
74	Di-n-butylphthalate	40.000	38.541	3.6	85	0.00
75 C	Fluoranthene	40.000	39.363	1.6	88	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	90	0.00
77	Benzydine	40.000	41.143	-2.9	93	0.00
78	Pyrene	40.000	39.081	2.3	89	0.00
79 S	Terphenyl-d14	80.000	76.831	4.0	89	0.00
80	Butylbenzylphthalate	40.000	39.300	1.8	92	0.00
81	Benzo(a)anthracene	40.000	38.743	3.1	94	0.00
82	3,3'-Dichlorobenzidine	40.000	38.373	4.1	96	0.00
83	Chrysene	40.000	38.599	3.5	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.133	-2.8	101	0.00
85 c	Di-n-octyl phthalate	40.000	39.758	0.6	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	89	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.734	-6.8	94	0.00
88	Benzo(b)fluoranthene	40.000	40.499	-1.2	96	0.00
89	Benzo(k)fluoranthene	40.000	36.136	9.7	93	0.00
90 C	Benzo(a)pyrene	40.000	38.970	2.6	94	0.00
91	Dibenzo(a,h)anthracene	40.000	43.186	-8.0	94	0.00
92	Benzo(g,h,i)perylene	40.000	43.758	-9.4	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081822\
Data File : BF129868.D
Acq On : 18 Aug 2022 10:50
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 18 13:35:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 18 12:03:10 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0