

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081825\
 Data File : BF143433.D
 Acq On : 18 Aug 2025 11:02
 Operator : RC/JU
 Sample : PB169256BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB169256BL

Quant Time: Aug 18 12:04:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 18 04:40:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.928	152	126309	20.000	ng	0.00
21) Naphthalene-d8	8.204	136	491978	20.000	ng	0.00
39) Acenaphthene-d10	9.963	164	278073	20.000	ng	0.00
64) Phenanthrene-d10	11.445	188	476848	20.000	ng	0.00
76) Chrysene-d12	14.086	240	269560	20.000	ng	0.00
86) Perylene-d12	15.574	264	230909	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.569	112	691978	95.233	ng	0.01
7) Phenol-d6	6.557	99	881267	96.240	ng	0.00
23) Nitrobenzene-d5	7.481	82	825215	99.532	ng	0.00
42) 2,4,6-Tribromophenol	10.751	330	239290	96.681	ng	0.00
45) 2-Fluorobiphenyl	9.281	172	1517117	91.766	ng	0.00
79) Terphenyl-d14	13.027	244	1660445	103.476	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081825\
Data File : BF143433.D
Acq On : 18 Aug 2025 11:02
Operator : RC/JU
Sample : PB169256BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB169256BL


Quant Time: Aug 18 12:04:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 18 04:40:19 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.928 min Scan# 805
Delta R.T. -0.000 min
Lab File: BF143433.D
Acq: 18 Aug 2025 11:02


Instrument :
BNA_F
ClientSampleId :
PB169256BL





Tgt Ion:152 Resp: 126309
Ion Ratio Lower Upper
152 100
150 156.4 127.4 191.2
115 59.3 47.5 71.3




#5
2-Fluorophenol
Concen: 95.233 ng
RT: 5.569 min Scan# 574
Delta R.T. 0.012 min
Lab File: BF143433.D
Acq: 18 Aug 2025 11:02



Tgt Ion:112 Resp: 691978
Ion Ratio Lower Upper
112 100
64 64.8 52.2 78.2
63 33.7 27.0 40.6









#7
Phenol-d6
Concen: 96.240 ng
RT: 6.557 min Scan# 742
Delta R.T. -0.000 min
Lab File: BF143433.D
Acq: 18 Aug 2025 11:02




Instrument :
BNA_F
ClientSampleId :
PB169256BL


Tgt Ion: 99 Resp: 881267
Ion Ratio Lower Upper
99 100
42 20.1 17.0 25.6
71 34.3 27.4 41.2



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.204 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BF143433.D
Acq: 18 Aug 2025 11:02

Tgt Ion: 136 Resp: 491978
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 9.8 7.8 11.8
68 6.7 5.2 7.8





#23

Nitrobenzene-d5

Concen: 99.532 ng

RT: 7.481 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF143433.D

Acq: 18 Aug 2025 11:02

Instrument :

BNA_F

ClientSampleId :

PB169256BL





Tgt Ion: 82 Resp: 825215

Ion Ratio Lower Upper

82 100

128 42.9 34.6 52.0

54 57.6 46.0 69.0



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.963 min Scan# 1321

Delta R.T. -0.000 min

Lab File: BF143433.D

Acq: 18 Aug 2025 11:02




Tgt Ion:164 Resp: 278073


Ion Ratio Lower Upper

164 100

162 99.4 80.6 121.0

160 43.0 34.6 51.8





#42

2,4,6-Tribromophenol

Concen: 96.681 ng

RT: 10.751 min Scan# 1455

Delta R.T. -0.000 min

Lab File: BF143433.D


Acq: 18 Aug 2025 11:02

Instrument :

BNA_F

ClientSampleId :

PB169256BL






Tgt Ion:330 Resp: 239290

Ion Ratio Lower Upper

330 100

332 97.8 77.4 116.2

141 30.2 24.1 36.1



#45

2-Fluorobiphenyl

Concen: 91.766 ng

RT: 9.281 min Scan# 1205

Delta R.T. -0.000 min

Lab File: BF143433.D

Acq: 18 Aug 2025 11:02




Tgt Ion:172 Resp: 1517117


Ion Ratio Lower Upper

172 100

171 35.8 28.7 43.1

170 23.7 19.0 28.6









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.445 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF143433.D
Acq: 18 Aug 2025 11:02




Instrument :
BNA_F
ClientSampleId :
PB169256BL


Tgt Ion:188 Resp: 476848
Ion Ratio Lower Upper
188 100
94 10.3 8.0 12.0
80 11.3 8.6 12.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.086 min Scan# 2022
Delta R.T. -0.000 min
Lab File: BF143433.D
Acq: 18 Aug 2025 11:02

Tgt Ion:240 Resp: 269560
Ion Ratio Lower Upper
240 100
120 10.6 9.0 13.4
236 25.1 20.2 30.2









#79
 Terphenyl-d14
 Concen: 103.476 ng
 RT: 13.027 min Scan# 1842
 Delta R.T. -0.000 min
 Lab File: BF143433.D
 Acq: 18 Aug 2025 11:02

Instrument :
 BNA_F
 ClientSampleId :
 PB169256BL

Tgt Ion:244 Resp: 1660445
 Ion Ratio Lower Upper
 244 100
 212 6.7 5.2 7.8
 122 10.2 8.3 12.5



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.574 min Scan# 2275
 Delta R.T. -0.000 min
 Lab File: BF143433.D
 Acq: 18 Aug 2025 11:02

Tgt Ion:264 Resp: 230909
 Ion Ratio Lower Upper
 264 100
 260 23.2 18.8 28.2
 265 20.8 17.4 26.0

