

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081919\
 Data File : BF116393.D
 Acq On : 19 Aug 2019 17:51
 Operator : HP/JU
 Sample : K4350-04
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SS-18 081519

Quant Time: Aug 20 01:57:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 06:03:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	78353	20.00	ng	-0.01
21) Naphthalene-d8	8.08	136	304257	20.00	ng	-0.02
39) Acenaphthene-d10	9.83	164	164992	20.00	ng	-0.02
64) Phenanthrene-d10	11.32	188	314593	20.00	ng	-0.02
76) Chrysene-d12	13.96	240	296036	20.00	ng	-0.02
87) Perylene-d12	15.42	264	348743	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	396703	84.76	ng	-0.01
7) Phenol-d6	6.45	99	515937	87.37	ng	-0.02
23) Nitrobenzene-d5	7.36	82	378593	71.83	ng	-0.02
42) 2,4,6-Tribromophenol	10.63	330	155038	96.92	ng	-0.02
45) 2-Fluorobiphenyl	9.15	172	672886	76.39	ng	-0.02
79) Terphenyl-d14	12.90	244	867496	61.75	ng	-0.02
Target Compounds						
50) Dimethylphthalate	9.55	163	54690	4.868	ng	98
75) Fluoranthene	12.54	202	38364	2.279	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081919\
Data File : BF116393.D
Acq On : 19 Aug 2019 17:51
Operator : HP/JU
Sample : K4350-04
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SS-18_081519

Quant Time: Aug 20 01:57:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 06:03:31 2019
Response via : Initial Calibration

