

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081919\
 Data File : BF116405.D
 Acq On : 19 Aug 2019 23:17
 Operator : HP/JU
 Sample : K4290-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-2

Quant Time: Aug 20 03:29:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 06:03:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	97965	20.00	ng	-0.01
21) Naphthalene-d8	8.08	136	378329	20.00	ng	-0.02
39) Acenaphthene-d10	9.83	164	206062	20.00	ng	-0.02
64) Phenanthrene-d10	11.32	188	379606	20.00	ng	-0.02
76) Chrysene-d12	13.96	240	284336	20.00	ng	-0.02
87) Perylene-d12	15.42	264	272746	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	679053	116.04	ng	0.00
7) Phenol-d6	6.45	99	880961	119.32	ng	-0.01
23) Nitrobenzene-d5	7.37	82	561351	85.65	ng	-0.02
42) 2,4,6-Tribromophenol	10.63	330	247852	124.06	ng	-0.02
45) 2-Fluorobiphenyl	9.15	172	991122	90.09	ng	-0.02
79) Terphenyl-d14	12.90	244	1169618	86.68	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.55	163	233439	16.638	ng	99
75) Fluoranthene	12.54	202	60305	2.968	ng	98
78) Pyrene	12.77	202	48235	2.250	ng	97
84) Bis(2-ethylhexyl)phthalate	13.92	149	35827	3.041	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF081919\
Data File : BF116405.D
Acq On : 19 Aug 2019 23:17
Operator : HP/JU
Sample : K4290-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-2

Quant Time: Aug 20 03:29:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 06:03:31 2019
Response via : Initial Calibration

