

Data Path : U:\HPCHEM1\BNA F\DATA\BF082117\
 Data File : BF097902.D
 Acq On : 21 Aug 2017 13:13
 Operator : SJ/JU
 Sample : I4849-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CAPTAINSCOVE-WC-002

Manual Integrations
 APPROVED

Sohil
 8/22/2017 5:43:47 PM

Quant Time: Aug 22 01:49:34 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 15:59:04 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	150922	20.00	ng	0.00
21) Naphthalene-d8	8.03	136	589782	20.00	ng	0.00
38) Acenaphthene-d10	9.78	164	250632	20.00	ng	0.00
63) Phenanthrene-d10	11.26	188	404270	20.00	ng	0.00
75) Chrysene-d12	13.89	240	309117	20.00	ng	0.00
86) Perylene-d12	15.32	264	283291	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	1184359	119.37	ng	0.01
7) Phenol-d6	6.39	99	1610988	119.50	ng	0.00
23) Nitrobenzene-d5	7.32	82	1014019	93.40	ng	0.00
41) 2,4,6-Tribromophenol	10.57	330	259842	119.49	ng	0.00
44) 2-Fluorobiphenyl	9.11	172	1300015	76.21	ng	0.00
78) Terphenyl-d14	12.85	244	769411	51.91	ng	0.00
Target Compounds						
10) Phenol	6.40	94	39571	2.510	ng	99
32) Benzoic acid	7.73	122	2265	6.691	ng	88
49) Dimethylphthalate	9.50	163	292014	15.939	ng	99
74) Fluoranthene	12.47	202	69172	3.076	ng	99
77) Pyrene	12.69	202	56305	2.227	ng	97
87) Benzo(b)fluoranthene	14.90	252	44581m	2.497	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA F\DATA\BF082117\
Data File : BF097902.D
Acq On : 21 Aug 2017 13:13
Operator : SJ/JU
Sample : I4849-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CAPTAINSCOVE-WC-002

Manual Integrations
APPROVED

Sohil
8/22/2017 5:43:47 PM

Quant Time: Aug 22 01:49:34 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 15:59:04 2017
Response via : Initial Calibration

