


Data Path : U:\HPCHEM1\BNA F\DATA\BF082117\
 Data File : BF097895.D
 Acq On : 21 Aug 2017 9:59
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 21 15:59:04 2017



AutoFind: Scans 1603, 1604, 1605; Background Corrected with Scan 1599

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.4	33901	PASS
68	69	0.00	2	2.0	745	PASS
69	198	0.00	100	44.8	36736	PASS
70	69	0.00	2	0.4	160	PASS
127	198	10	80	54.2	44400	PASS
197	198	0.00	2	0.7	554	PASS
198	198	100	100	100.0	81981	PASS
199	198	5	9	6.8	5588	PASS
275	198	10	60	22.4	18382	PASS
365	198	1	100	2.5	2068	PASS
441	442	0.01	24	15.5	9494	PASS
442	198	50	100	74.7	61261	PASS
443	442	15	24	19.1	11726	PASS