

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082125\
 Data File : BF143560.D
 Acq On : 22 Aug 2025 06:40
 Operator : RC/JU
 Sample : Q2892-04
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 1-FLOOR-DRILL-B-BACK

Quant Time: Aug 22 07:00:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 21 06:12:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.928	152	96411	20.000	ng	0.00
21) Naphthalene-d8	8.204	136	363727	20.000	ng	0.00
39) Acenaphthene-d10	9.963	164	189282	20.000	ng	0.00
64) Phenanthrene-d10	11.451	188	257639	20.000	ng	0.00
76) Chrysene-d12	14.092	240	206512	20.000	ng	-0.01
86) Perylene-d12	15.592	264	218750	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.563	112	415376	75.231	ng	0.00
7) Phenol-d6	6.557	99	533312	78.247	ng	-0.02
23) Nitrobenzene-d5	7.481	82	353533	56.719	ng	-0.02
42) 2,4,6-Tribromophenol	10.751	330	120206	68.224	ng	-0.01
45) 2-Fluorobiphenyl	9.281	172	680528	59.419	ng	0.00
79) Terphenyl-d14	13.033	244	537826	45.447	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082125\
Data File : BF143560.D
Acq On : 22 Aug 2025 06:40
Operator : RC/JU
Sample : Q2892-04
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
1-FLOOR-DRILL-B-BACK

Quant Time: Aug 22 07:00:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 21 06:12:21 2025
Response via : Initial Calibration

